

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
 Português : Planilha de valores do Sistema de Medição BS da Mindray
 Español : Hoja de valores del sistema de medición Mindray BS
 Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
 Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский : Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

7. **BS-380**: BS-380, BS-390;

2. **BS-240**;

8. **BS-400**: BS-400, BS-420;

3. **BS-240E**: BS240E, BS240Pro;

9. **BS-430**: BS-450, BS-460;

4. **BS-300**: BS-300, BS-320;

10. **BS-480**: BS-480, BS-490;

5. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-");

11. **BS-600**: BS-600, BS-620;

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00)(Serial Number starts with "W8-"
 and software version starts with "35.00"),
BS-350E(V35.00) (Serial Number starts with "W9-"
 and software version starts with "35.00");

12. **BS-600M**: BS-600M, BS-620M;

13. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,
 BS-1800, BS-1800plus;

14. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15. **BS-2800M**: BS-2600M, BS-2800M.

16. **S0: 0.9% NaCl, Conc. Of S0=0;**

Русский : **S0: 0, 9% NaCl, конц. S0=0;**

Português : **S0: 0,9% NaCl, Conc. de S0=0;**

Español : **S0: 0,9% NaCl, Conc. de S0=0;**

Italiano : **S0: 0,9% NaCl, conc. di S0=0;**

Türkçe : **S0: %0,9 NaCl, S0 Kons.=0.**

LOT : 043423007

EXP : 2024-11-19

| Abbreviated name | HS-CRP | Calibration Rule | spline |
|-----------------------------|--------|--------------------------------|--------|
| Model | Level | Calibrator Value ¹⁶ | |
| | | mg/L | nmol/L |
| BS-200E ¹ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-240 ² | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-240E ³ | a | 5.30 | 50.5 |
| | b | 17.8 | 169 |
| | c | 37.5 | 357 |
| | d | 146 | 1390 |
| | e | 305 | 2904 |
| BS-300 ⁴ | a | 4.92 | 46.8 |
| | b | 17.8 | 169 |
| | c | 39.0 | 371 |
| | d | 152 | 1447 |
| | e | 321 | 3056 |
| BS-330E ⁵ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |

HS-CRP Calibrator



| Abbreviated name | HS-CRP | Calibration Rule | spline |
|-----------------------|--------|--------------------------------|--------|
| Model | Level | Calibrator Value ¹⁶ | |
| | | mg/L | nmol/L |
| BS-360E ⁶ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-380 ⁷ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-400 ⁸ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-430 ⁹ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-480 ¹⁰ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-600 ¹¹ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |
| BS-600M ¹² | a | 4.76 | 45.3 |
| | b | 18.1 | 172 |
| | c | 38.3 | 365 |
| | d | 154 | 1466 |
| | e | 320 | 3046 |
| BS-800 ¹³ | a | 4.80 | 45.7 |
| | b | 18.0 | 171 |
| | c | 38.1 | 363 |
| | d | 151 | 1438 |
| | e | 319 | 3037 |

HS-CRP Calibrator



| Abbreviated name | HS-CRP | Calibration Rule | spline |
|------------------------|--------|--------------------------------|--------|
| Model | Level | Calibrator Value ¹⁶ | |
| | | mg/L | nmol/L |
| BS-2000 ¹⁴ | a | 4.96 | 47.2 |
| | b | 18.6 | 177 |
| | c | 38.8 | 369 |
| | d | 154 | 1466 |
| | e | 321 | 3056 |
| BS-2800M ¹⁵ | a | 4.76 | 45.3 |
| | b | 18.1 | 172 |
| | c | 38.3 | 365 |
| | d | 154 | 1466 |
| | e | 320 | 3046 |

| English | Abbreviated name | Calibration Rule | Calibration Value | Model | Level |
|------------------|--------------------------|------------------------|------------------------|---------|---------|
| Русский | сокращенное наименование | Принцип калибровки | Эталонное значение | модель | Уровень |
| Português | Nome abreviado | Regra de calibração | Valor de calibração | Modelo | Nível |
| Español | nombre abreviado | Regla de calibración | Valor de calibración | modelo | Nivel |
| Italiano | abbreviazione | Regola di calibrazione | Valore di calibrazione | modelli | Livello |
| Türkçe | kısaltılmış ad | Kalibrasyon Kuralı | Kalibrasyon Değeri | model | Düzey |

| HS-CRP | |
|------------------|--|
| English | High Sensitivity C-reaction Protein |
| Русский | C-реактивный белок |
| Português | Proteína C Reativa de Alta Sensibilidade |
| Español | proteína C reactiva de alta sensibilidad |
| Italiano | proteina C reattiva ad alta sensibilità |
| Türkçe | Yüksek Duyarlıklılı C Reaksiyon Proteindir |