

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200**: BS-200, BS-220;

8. **BS-380**: BS-380, BS-390;

2. **BS-200E**: BS-200E, BS-220E;

9. **BS-400**: BS-400, BS-420;

3. **BS-230**: BS-230, BS-240;

10. **BS-430**: BS-430, BS-450, BS-460;

4. **BS-240E**: BS240E, BS240Pro;

11. **BS-480**: BS-480, BS-490;

5. **BS-300**: BS-300, BS-320;

12. **BS-600**: BS-600, BS-620;

6. **BS-330E**: BS-330E (Serial Number starts with "XQ-"),  
BS-350E (Serial Number starts with "XS-");

13. **BS-600M**: BS-600M, BS-620M;

7. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

14. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

**BS-330E(V35.00)** (Serial Number starts with "W8-"  
and software version starts with "35.00"),

BS-1800, BS-1800plus;

**BS-350E(V35.00)** (Serial Number starts with "W9-"  
and software version starts with "35.00");

15. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

16. **BS-2800M**: BS-2600M, BS-2800M.

**LOT**: 045623004

**EXP**: 2024-08-02

Abbreviated name		Hb	HbA1c
Model	Level	Calibrator Value (μmol/L)	Calibrator Value (μmol/L)
BS-200 <sup>1</sup>	S1	93.6	3.30
	S2	146	15.8
BS-200E <sup>2</sup>	S1	94.5	3.40
	S2	147	16.3
BS-230 <sup>3</sup>	S1	94.5	3.50
	S2	147	16.5
BS-240E <sup>4</sup>	S1	94.5	3.50
	S2	147	16.5
BS-300 <sup>5</sup>	S1	94.5	3.40
	S2	147	16.3
BS-330E <sup>6</sup>	S1	94.5	3.40
	S2	147	16.3
BS-360E <sup>7</sup>	S1	94.5	3.50
	S2	147	16.5
BS-380 <sup>8</sup>	S1	94.5	3.40
	S2	147	16.3
BS-400 <sup>9</sup>	S1	94.5	3.40
	S2	147	16.3
BS-430 <sup>10</sup>	S1	94.5	3.50
	S2	147	16.5
BS-480 <sup>11</sup>	S1	94.5	3.50
	S2	147	16.5
BS-600 <sup>12</sup>	S1	94.5	3.50
	S2	147	16.5
BS-600M <sup>13</sup>	S1	93.6	3.50
	S2	146	16.4
BS-800 <sup>14</sup>	S1	94.5	3.50
	S2	147	16.5

# HbA1C Calibrator

**mindray**

<b>BS-2000</b> <sup>15</sup>	<b>S1</b>	95.3	3.50	
	<b>S2</b>	149	16.6	
<b>BS-2800M</b> <sup>16</sup>	<b>S1</b>	95.3	3.50	
	<b>S2</b>	148	16.6	
English	Abbreviated name	Calibration Value	Model	Level
<b>Русский</b>	сокращенное наименование	Эталонное значение	модель	Уровень
<b>Português</b>	Nome abreviado	Valor de calibração	Modelo	Nível
<b>Español</b>	nombre abreviado	Valor de calibración	modelo	Nivel
<b>Italiano</b>	abbreviazione	Valore di calibrazione	modelli	Livello
<b>Türkçe</b>	kısaltılmış ad	Kalibrasyon Değeri	model	Düzey

<b>HbA1c/Hb</b>	
<b>English</b>	Hemoglobin A1c/ Hemoglobin
<b>Русский</b>	гемоглобин A1c / гемоглобин
<b>Português</b>	Hemoglobina A1c /Hemoglobina
<b>Español</b>	hemoglobina A1c /hemoglobina
<b>Italiano</b>	Emoglobina A1c/Emoglobina
<b>Türkçe</b>	Hemoglobin A1c / Hemoglobin