

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-240**;
2. **BS-240E**: BS-240E, BS-240Pro;
3. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)**(Serial Number starts with "W9-" and software version starts with "35.00");
4. **BS-380**: BS-380, BS-390;
5. **BS-400**: BS-400, BS-420;
6. **BS-430**: BS-450, BS-460;
7. **BS-480**: BS-480, BS-490;
8. **BS-600**: BS-600, BS-620;
9. **BS-600M**: BS-600M, BS-620M;
10. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
11. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;
12. **BS-2800M**: BS-2600M, BS-2800M.

13. a:0.9% NaCl is not needed, Conc. Of a=0;

Русский: a: 0,9% NaCl не необходим, конц. a=0;

Português: a: 0,9% O NaCl não é needed, Conc. de a=0;

Español: a: 0,9% El NaCl no es necesario, Conc. de a=0;

Italiano: a: 0.9% Il NaCl non è necessario, conc. di a=0;

Türkçe: a: %0,9 NaCl gerekli değildir, a Kons.=0.

LOT: 045723002

EXP: 2023-12-30

| Abbreviated name | | | D-Dimer | | | Calibration Rule | | | spline | | |
|---------------------|-------|--|----------------------|-------|--|----------------------|-------|--|----------------------|-------|--|
| Model | Level | Calibration Value ¹³ (µg/mL) | Model | Level | Calibration Value ¹³ (µg/mL) | Model | Level | Calibration Value ¹³ (µg/mL) | Model | Level | Calibration Value ¹³ (µg/mL) |
| BS-240 ¹ | a | 0 | BS-240E ² | a | 0 | BS-360E ³ | a | 0 | BS-360E ³ | a | 0 |
| | b | 2.41 | | b | 2.41 | | b | 2.41 | | | |
| | c | 8.84 | | c | 8.84 | | c | 8.84 | | | |
| | d | 16.8 | | d | 16.8 | | d | 16.8 | | | |
| | e | 34.5 | | e | 34.5 | | e | 34.5 | | | |
| | f | 50.7 | | f | 50.7 | | f | 50.7 | | | |
| BS-380 ⁴ | a | 0 | BS-400 ⁵ | a | 0 | BS-430 ⁶ | a | 0 | BS-430 ⁶ | a | 0 |
| | b | 2.41 | | b | 2.41 | | b | 2.41 | | | |
| | c | 8.84 | | c | 8.84 | | c | 8.84 | | | |
| | d | 16.8 | | d | 16.8 | | d | 16.8 | | | |
| | e | 34.5 | | e | 34.5 | | e | 34.5 | | | |
| | f | 50.7 | | f | 50.7 | | f | 50.7 | | | |
| BS-480 ⁷ | a | 0 | BS-600 ⁸ | a | 0 | BS-600M ⁹ | a | 0 | BS-600M ⁹ | a | 0 |
| | b | 2.41 | | b | 2.41 | | b | 2.44 | | | |
| | c | 8.84 | | c | 8.84 | | c | 8.97 | | | |
| | d | 16.8 | | d | 16.8 | | d | 17.0 | | | |
| | e | 34.5 | | e | 34.5 | | e | 35.1 | | | |
| | f | 50.7 | | f | 50.7 | | f | 51.1 | | | |

D-Dimer Calibrator

| Abbreviated name | | | D-Dimer | | | Calibration Rule | | | spline | | |
|----------------------|-------|--|-----------------------|-------|--|------------------------|-------|--|--------|--|--|
| Model | Level | Calibration Value ¹³ (µg/mL) | Model | Level | Calibration Value ¹³ (µg/mL) | Model | Level | Calibration Value ¹³ (µg/mL) | | | |
| BS-800 ¹⁰ | a | 0 | BS-2000 ¹¹ | a | 0 | BS-2800M ¹² | a | 0 | | | |
| | b | 2.41 | | b | 2.40 | | b | 2.44 | | | |
| | c | 8.84 | | c | 8.95 | | c | 8.97 | | | |
| | d | 16.8 | | d | 17.0 | | d | 17.0 | | | |
| | e | 34.5 | | e | 35.4 | | e | 35.1 | | | |
| | f | 50.7 | | f | 51.3 | | f | 51.1 | | | |

| English | Abbreviated name | Calibration Rule | Calibration Value | Model | Level |
|-----------|--------------------------|------------------------|------------------------|---------|---------|
| Русский | сокращенное наименование | Принцип калибровки | Эталонное значение | модель | Уровень |
| Português | Nome abreviado | Regra de calibração | Valor de calibração | Modelo | Nível |
| Español | nombre abreviado | Regla de calibración | Valor de calibración | modelo | Nivel |
| Italiano | abbreviazione | Regola di calibrazione | Valore di calibrazione | modelli | Livello |
| Türkçe | kısaltılmış ad | Kalibrasyon Kuralı | Kalibrasyon Değeri | model | Düzey |

| D-Dimer | |
|-----------|----------|
| English | D-Dimer |
| Русский | D-димера |
| Português | D-Dimer |
| Español | Dímero D |
| Italiano | D-Dimero |
| Türkçe | D-Dimer |