

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-240**;

2. **BS-240E**: BS-240E, BS-240Pro;

3. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

4. **BS-380**: BS-380, BS-390;

5. **BS-400**: BS-400, BS-420;

6. **BS-430**: BS-450, BS-460;

7. **BS-480**: BS-480, BS-490;

8. **BS-600**: BS-600, BS-620;

9. **BS-600M**: BS-600M, BS-620M;

10. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

11. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

12. **BS-2800M**: BS-2600M, BS-2800M.

13. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Русский: Cal0: 0,9% NaCl, конц. Cal0=0;

Português: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Español: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Italiano: Cal0: 0,9% NaCl, conc. di Cal0=0;

Türkçe: Cal0: %0,9 NaCl, Cal0 Kons.=0

LOT : 045822017

EXP : 2023-12-21

Abbreviated name		MYO	Calibration Rule			spline		
Model	Level	Calibration Value ¹³ (ng/mL)	Model	Level	Calibration Value ¹³ (ng/mL)	Model	Level	Calibration Value ¹³ (ng/mL)
BS-240	a	90.0	BS-240E	a	90.0	BS-360E	a	90.0
	b	185		b	185		b	185
	c	384		c	384		c	384
	d	742		d	742		d	742
BS-380	a	90.0	BS-400	a	90.0	BS-430	a	90.0
	b	185		b	185		b	185
	c	384		c	384		c	384
	d	742		d	742		d	742
BS-480	a	90.0	BS-600	a	90.0	BS-600M	a	94.0
	b	185		b	185		b	190
	c	384		c	384		c	391
	d	742		d	742		d	749
BS-800	a	90.0	BS-2000	a	93.0	BS-2800M	a	94.0
	b	185		b	189		b	190
	c	384		c	392		c	391
	d	742		d	747		d	749

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

MYO	
English	Myoglobin
Русский	Миоглобин
Português	Mioglobina
Español	Mioglobina
Italiano	Mioglobina
Türkçe	Myoglobin