

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

2.**BS-200**: BS-200, BS-220;

3.**BS-200E**: BS-200E, BS-220E;

4.**BS-230**: BS-230, BS-240;

5.**BS-240E**: BS-240E, BS-240Pro;

6.**BS-360E**: BS-360E,BS-370E,BS-350S,BS-360S,

BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

7.**BS-380**: BS-380, BS-390;

8.**BS-400**: BS-400, BS-420;

9.**BS-430**: BS-430, BS-450, BS-460;

10.**BS-480**: BS-480, BS-490;

11.**BS-600**: BS-600, BS-620;

12.**BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

13.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

14.**BS-2800M**: BS-2600M, BS-2800M.

15. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Português: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Italiano: Cal0: 0,9% NaCl, conc. di Cal0=0;

Русский: Cal0: 0,9% NaCl, конц. Cal0=0;

Español: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Türkçe: Cal0: %0,9 NaCl, Cal0 Kons.=0.

LOT: 046022003

EXP: 2023-09-19

| Abbreviated name | | MALB | | Calibration Rule | | spline | |
|-----------------------------|-------|---------------------------------|--------|-----------------------------|-------|---------------------------------|--------|
| Model | Level | Calibration Value ¹⁵ | | Model | Level | Calibration Value ¹⁵ | |
| | | mg/L | µmol/L | | | mg/L | µmol/L |
| BS-120 ¹ | a | 21.3 | 0.324 | BS-400 ⁸ | a | 21.3 | 0.324 |
| | b | 56.4 | 0.857 | | b | 56.4 | 0.857 |
| | c | 106 | 1.61 | | c | 106 | 1.61 |
| | d | 206 | 3.13 | | d | 206 | 3.13 |
| | e | 382 | 5.81 | | e | 382 | 5.81 |
| BS-200 ² | a | 21.3 | 0.324 | BS-430 ⁹ | a | 21.3 | 0.324 |
| | b | 56.4 | 0.857 | | b | 56.4 | 0.857 |
| | c | 106 | 1.61 | | c | 106 | 1.61 |
| | d | 206 | 3.13 | | d | 206 | 3.13 |
| | e | 382 | 5.81 | | e | 382 | 5.81 |
| BS-200E ³ | a | 21.3 | 0.324 | BS-480 ¹⁰ | a | 21.3 | 0.324 |
| | b | 56.4 | 0.857 | | b | 56.4 | 0.857 |
| | c | 106 | 1.61 | | c | 106 | 1.61 |
| | d | 206 | 3.13 | | d | 206 | 3.13 |
| | e | 382 | 5.81 | | e | 382 | 5.81 |

| Abbreviated name | | MALB | | Calibration Rule | | spline | |
|----------------------|-------|---------------------------------|--------|------------------------|-------|---------------------------------|--------|
| Model | Level | Calibration Value ¹⁵ | | Model | Level | Calibration Value ¹⁵ | |
| | | mg/L | μmol/L | | | mg/L | μmol/L |
| BS-230 ⁴ | a | 21.3 | 0.324 | BS-600 ¹¹ | a | 21.3 | 0.324 |
| | b | 56.4 | 0.857 | | b | 56.4 | 0.857 |
| | c | 106 | 1.61 | | c | 106 | 1.61 |
| | d | 206 | 3.13 | | d | 206 | 3.13 |
| | e | 382 | 5.81 | | e | 382 | 5.81 |
| BS-240E ⁵ | a | 21.3 | 0.324 | BS-800 ¹² | a | 21.3 | 0.324 |
| | b | 56.4 | 0.857 | | b | 56.4 | 0.857 |
| | c | 106 | 1.61 | | c | 106 | 1.61 |
| | d | 206 | 3.13 | | d | 206 | 3.13 |
| | e | 382 | 5.81 | | e | 382 | 5.81 |
| BS-360E ⁶ | a | 21.3 | 0.324 | BS-2000 ¹³ | a | 21.5 | 0.327 |
| | b | 56.4 | 0.857 | | b | 56.5 | 0.859 |
| | c | 106 | 1.61 | | c | 105 | 1.60 |
| | d | 206 | 3.13 | | d | 205 | 3.12 |
| | e | 382 | 5.81 | | e | 376 | 5.72 |
| BS-380 ⁷ | a | 21.3 | 0.324 | BS-2800M ¹⁴ | a | 21.0 | 0.319 |
| | b | 56.4 | 0.857 | | b | 55.7 | 0.847 |
| | c | 106 | 1.61 | | c | 105 | 1.60 |
| | d | 206 | 3.13 | | d | 205 | 3.12 |
| | e | 382 | 5.81 | | e | 372 | 5.65 |

| English | Abbreviated name | Calibration Rule | Calibration Value | Model | Level |
|-----------|--------------------------|------------------------|------------------------|---------|---------|
| Русский | сокращенное наименование | Принцип калибровки | Эталонное значение | модель | Уровень |
| Português | Nome abreviado | Regra de calibração | Valor de calibração | Modelo | Nível |
| Español | nombre abreviado | Regla de calibración | Valor de calibración | modelo | Nivel |
| Italiano | abbreviazione | Regola di calibrazione | Valore di calibrazione | modelli | Livello |
| Türkçe | kısaltılmış ad | Kalibrasyon Kuralı | Kalibrasyon Değeri | model | Düzey |

| MALB | |
|-----------|------------------|
| English | Microalbumin |
| Русский | микроальбумина |
| Português | Microalbuminúria |
| Español | Microalbúmina |
| Italiano | Microalbumina |
| Türkçe | Mikroalbumin |