

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

9.BS-430: BS-430, BS-450,BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

10.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

11.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

12.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

14.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

7.BS-380: BS-380, BS-390;

15.BS-2800M:BS-2600M, BS-2800M;

8.BS-400: BS-400, BS-420;

16.For applicable analyzers of the analyte, please subject to the parameter sheet and instrument.

LOT : 046024003

: 2025-09-12

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kisaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

MALB	
English	Microalbumin
Русский	микроальбумина
Português	Microalbuminúria
Español	Microalbúmina
Italiano	Microalbumina
Türkçe	Mikroalbumin

Abbreviated name		MALB		Calibration Rule		spline			
Model		BS-120 ¹		BS-200 ²		BS-200E ³		BS-240 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	mg/L	21.8	1.9	21.8	1.9	21.8	1.9	21.8	1.9
b		56.4	2.7	56.4	2.7	56.4	2.7	56.4	2.7
c		104	5	104	5	104	5	104	5

d		209	13	209	13	209	13	209	13
e		383	17	383	17	383	17	383	17
0.9%NaCl		0.000	/	0.000	/	0.000	/	0.000	/
a	μmol/L	0.331	0.029	0.331	0.029	0.331	0.029	0.331	0.029
b		0.857	0.04	0.857	0.04	0.857	0.04	0.857	0.04
c		1.58	0.08	1.58	0.08	1.58	0.08	1.58	0.08
d		3.18	0.20	3.18	0.20	3.18	0.20	3.18	0.20
e		5.82	0.25	5.82	0.25	5.82	0.25	5.82	0.25
Model		BS-240E⁵		BS-360E⁶		BS-380⁷		BS-400⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	mg/L	21.8	1.9	21.8	1.9	21.8	1.9	21.8	1.9
b		56.4	2.7	56.4	2.7	56.4	2.7	56.4	2.7
c		104	5	104	5	104	5	104	5
d		209	13	209	13	209	13	209	13
e		383	17	383	17	383	17	383	17
0.9%NaCl		0.000	/	0.000	/	0.000	/	0.000	/
a	μmol/L	0.331	0.029	0.331	0.029	0.331	0.029	0.331	0.029
b		0.857	0.04	0.857	0.04	0.857	0.04	0.857	0.04
c		1.58	0.08	1.58	0.08	1.58	0.08	1.58	0.08
d		3.18	0.20	3.18	0.20	3.18	0.20	3.18	0.20
e		5.82	0.25	5.82	0.25	5.82	0.25	5.82	0.25
Model		BS-430⁹		BS-480¹⁰		BS-600¹¹		BS-600M¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	mg/L	21.8	1.9	21.8	1.9	21.8	1.9	21.8	1.9
b		56.4	2.7	56.4	2.7	56.4	2.7	56.4	2.7
c		104	5	104	5	104	5	104	5
d		209	13	209	13	209	13	209	13
e		383	17	383	17	383	17	383	17
0.9%NaCl		0.000	/	0.000	/	0.000	/	0.000	/
a	μmol/L	0.331	0.029	0.331	0.029	0.331	0.029	0.331	0.029
b		0.857	0.04	0.857	0.04	0.857	0.04	0.857	0.04
c		1.58	0.08	1.58	0.08	1.58	0.08	1.58	0.08
d		3.18	0.20	3.18	0.20	3.18	0.20	3.18	0.20
e		5.82	0.25	5.82	0.25	5.82	0.25	5.82	0.25
Model		BS-800¹³		BS-2000¹⁴		BS-2800M¹⁵			
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty		
0.9%NaCl		0.0	/	0.0	/	0.0	/		
a	mg/L	21.8	1.9	22.2	1.9	22.1	1.9		
b		56.4	2.7	56.8	2.7	57.1	2.7		
c		104	5	105	5	106	5		
d		209	13	209	13	205	13		
e		383	17	376	16	381	17		
0.9%NaCl		0.000	/	0.000	/	0.000	/		
a	μmol/L	0.331	0.029	0.337	0.03	0.336	0.03		
b		0.857	0.04	0.863	0.04	0.868	0.04		

MALB Calibrator



c	1.58	0.08	1.60	0.08	1.61	0.08
d	3.18	0.20	3.18	0.20	3.12	0.20
e	5.82	0.25	5.72	0.25	5.79	0.25

mindray

MALB Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 046024003

🕒 2025-09-12

