

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-230**: BS-230,BS-240;

3. **BS-240E**: BS-240E, BS-240Pro;

4. **BS-360E**: BS-360E, BS-370E,BS-350S,BS-360S,

**BS-330E(V35.00)** (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

5. **BS-380**: BS-380, BS-390;

14. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Português: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Italiano: **Cal0: 0,9% NaCl, conc. di Cal0=0;**

6. **BS-400**: BS-400,BS-420;

7. **BS-430**: BS-430,BS-450,BS-460;

8. **BS-480**: BS-480,BS-490;

9. **BS-600**: BS-600, BS-620;

10. **BS-600M**: BS-600M, BS-620M;

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

Русский: **Cal0: 0,9% NaCl, конц. Cal0=0;**

Español: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Türkçe: **Cal0: %0,9 NaCl, Cal0 Kons.=0**

**LOT** : 046323001

**EXP** : 2023-12-25

Abbreviated name		RBP	Calibration Rule		Spline
Model	Level	Calibration Value <sup>14</sup> (mg/dL)	Sample Vol for Dilution / $\mu$ L	Diluent Vol/ $\mu$ L	Sample Vol For Analysis / $\mu$ L
<b>BS-200E</b> <sup>1</sup> R1: R2: S= 180: 60: 4	Cal1	2.79	45	135	4
	Cal2	5.50	45	135	8
	Cal3	13.3	/	/	4
	Cal4	17.4	/	/	6
<b>BS-380</b> <sup>5</sup> R1: R2: S= 180: 60: 4	Cal1	2.72	40	120	4
	Cal2	5.46	40	120	8
	Cal3	12.8	/	/	4
	Cal4	17.4	/	/	6
<b>BS-400</b> <sup>6</sup> R1: R2: S= 180: 60: 4	Cal1	2.61	40	120	4
	Cal2	5.17	40	120	8
	Cal3	12.6	/	/	4
	Cal4	17.2	/	/	6
Abbreviated name		RBP	Calibration Rule		Spline
Model	Level	Calibration Value <sup>14</sup> (mg/dL)	Sample Vol For Analysis / $\mu$ L	Sample Vol for Dilution / $\mu$ L	Diluent Vol/ $\mu$ L
<b>BS-230</b> <sup>2</sup> R1: R2: S= 180: 60: 4	Cal1	2.79	4	35	105
	Cal2	5.50	8	35	105
	Cal3	13.3	4	/	/
	Cal4	17.4	6	/	/

Abbreviated name		RBP	Calibration Rule		Spline
Model	Level	Calibration Value <sup>14</sup> (mg/dL)	Sample Vol For Analysis / $\mu$ L	Sample Vol for Dilution / $\mu$ L	Diluent Vol/ $\mu$ L
<b>BS-240E</b> <sup>3</sup> R1: R2: S= 180: 60: 4	Cal1	2.72	4	34	102
	Cal2	5.37	8	34	102
	Cal3	12.8	4	/	/
	Cal4	17.4	6	/	/
<b>BS-360E</b> <sup>4</sup> R1: R2: S= 180: 60: 4	Cal1	2.68	4	34	102
	Cal2	5.33	8	34	102
	Cal3	12.7	4	/	/
	Cal4	17.2	6	/	/
<b>BS-430</b> <sup>7</sup> R1: R2: S= 180: 60: 4	Cal1	2.61	4	30	90
	Cal2	5.17	8	30	90
	Cal3	12.6	4	/	/
	Cal4	17.2	6	/	/
<b>BS-480</b> <sup>8</sup> R1: R2: S= 180: 60: 4	Cal1	2.71	4	40	120
	Cal2	5.36	8	40	120
	Cal3	12.8	4	/	/
	Cal4	17.2	6	/	/
<b>BS-600</b> <sup>9</sup> R1: R2: S= 180: 60: 4	Cal1	2.68	4	30	90
	Cal2	5.32	8	30	90
	Cal3	12.8	4	/	/
	Cal4	17.2	6	/	/
<b>BS-600M</b> <sup>10</sup> R1: R2: S= 180: 60: 4	Cal1	2.82	4	25	75
	Cal2	5.80	8	25	75
	Cal3	12.5	4	/	/
	Cal4	17.4	6	/	/
<b>BS-800</b> <sup>11</sup> R1: R2: S= 180: 60: 4	Cal1	2.61	4	30	90
	Cal2	5.17	8	30	90
	Cal3	12.6	4	/	/
	Cal4	17.2	6	/	/
<b>BS-2000</b> <sup>12</sup> R1: R2: S= 180: 60: 4	Cal1	2.80	4	25	75
	Cal2	5.54	8	25	75
	Cal3	12.6	4	/	/
	Cal4	17.2	6	/	/
<b>BS-2800M</b> <sup>13</sup> R1: R2: S= 180: 60: 4	Cal1	2.82	4	25	75
	Cal2	5.80	8	25	75
	Cal3	12.5	4	/	/
	Cal4	17.4	6	/	/

English	Abbreviated name	Calibration Rule	Model	Level
<b>Русский</b>	сокращенное наименование	Принцип калибровки	модель	Уровень
<b>Português</b>	Nome abreviado	Regra de calibração	Modelo	Nível
<b>Español</b>	nombre abreviado	Regla de calibración	modelo	Nivel
<b>Italiano</b>	abbreviazione	Regola di calibrazione	modelli	Livello
<b>Türkçe</b>	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Sample Vol. for Dilution	Diluent Vol.	Sample Vol. For Analysis
<b>Русский</b>	Эталонное значение	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
<b>Português</b>	Valor de calibração	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
<b>Español</b>	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
<b>Italiano</b>	Valore di calibrazione	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
<b>Türkçe</b>	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

<b>RBP</b>	
<b>English</b>	Retinol Binding Protein
<b>Русский</b>	Ретинол-связывающий белок
<b>Português</b>	Proteína transportadora do retinol
<b>Español</b>	Proteína fijadora de retinol
<b>Italiano</b>	Proteina legante il retinolo
<b>Türkçe</b>	Retinol Bağlayıcı Protein