

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray

Português : Planilha de valores do Sistema de Medição BS da Mindray

Español : Hoja de valores del sistema de medición Mindray BS

Italiano : Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский : Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

2.**BS-180**: BS-180, BS-190;

3.**BS-200**: BS-200, BS-220;

4.**BS-200E**: BS-200E, BS-220E;

5.**BS-230**: BS-230, BS-240;

6.**BS-240E**: BS-240E, BS-240Pro;

7.**BS-300**: BS-300, BS-320;

8.**BS-330**: BS-330, BS-350;

9.**BS-330E**:BS-330E(Serial Number starts with "XQ-"),

BS-350E(Serial Number starts with "XS-");

10.**BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"),

BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

11.**BS-380**: BS-380, BS-390;

12.**BS-400**: BS-400, BS-420;

13.**BS-430**: BS-430, BS-450, BS-460;

14.**BS-480**: BS-480, BS-490;

15.**BS-600**: BS-600, BS-620;

16.**BS-600M**: BS-600M, BS-620M;

17.**BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

18.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

19.**BS-2800M**: BS-2600M, BS-2800M.

20.**S0: 0.9% NaCl, Conc. Of S0=0;**

Português: S0: 0,9% NaCl, Conc. de S0=0;

Italiano: S0: 0,9% NaCl, conc. di S0=0;

Русский: S0: 0, 9% NaCl, конц. S0=0;

Español: S0: 0,9% NaCl, Conc. de S0=0;

Türkçe: S0: %0,9 NaCl, S0 Kons.=0.

LOT : 143223001



: 2024-01-13

Abbreviated name	Unit	Model	Calibration Value ²⁰	Model	Calibration Value ²⁰
TBA	µmol/L	BS-120 ¹	33.9	BS-380 ¹¹	32.3
		BS-180 ²	33.9	BS-400 ¹²	32.3
		BS-200 ³	32.8	BS-430 ¹³	33.3
		BS-200E ⁴	33.3	BS-480 ¹⁴	32.3
		BS-230 ⁵	33.4	BS-600 ¹⁵	32.3
		BS-240E ⁶	32.9	BS-600M ¹⁶	32.3
		BS-300 ⁷	32.3	BS-800 ¹⁷	32.3
		BS-330 ⁸	32.8	BS-2000 ¹⁸	32.3
		BS-330E ⁹	33.3	BS-2800M ¹⁹	32.9
		BS-360E ¹⁰	32.9		

English	Abbreviated name	Model	Unit	Calibration Value
Русский	сокращенное наименование	модель	Прибор	Принцип калибровки
Português	Nome abreviado	Modelo	Unidade	Regra de calibração
Español	nombre abreviado	modelo	Unidad	Regla de calibración
Italiano	abbreviazione	modelli	Unità	Regola di calibrazione
Türkçe	kısaltılmış ad	model	Ünite	Kalibrasyon Kuralı

TBA

English	Total Bile Acids
Русский	желчные кислоты
Português	Ácido Biliar Total
Español	ácidos biliares totales
Italiano	acidi biliari totali
Türkçe	Total Safra Asitleridir

CO2 and TBA Multi Control

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- | | |
|--|--|
| 1. BS-120 : BS-120, BS-130; | 10. BS-360E : BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00) (Serial Number starts with "W8-"
and software version starts with "35.00"),
BS-350E(V35.00) (Serial Number starts with "W9-"
and software version starts with "35.00"); |
| 2. BS-180 : BS-180, BS-190; | 11. BS-380 : BS-380, BS-390; |
| 3. BS-200 : BS-200, BS-220; | 12. BS-400 : BS-400, BS-420; |
| 4. BS-200E : BS-200E, BS-220E; | 13. BS-430 : BS-410, BS-430, BS-450, BS-460, BS-470; |
| 5. BS-230 : BS-230, BS-240, BS-280; | 14. BS-480 : BS-480, BS-490; |
| 6. BS-240E : BS-240E, BS-240Pro; | 15. BS-600 : BS-600, BS-620; |
| 7. BS-300 : BS-300, BS-320; | 16. BS-600M : BS-600M, BS-620M; |
| 8. BS-330 : BS-330, BS-350; | 17. BS-800 : BS-800, BS-820, BS-800M, BS-820M,
BS-1800, BS-1800plus; |
| 9. BS-330E : BS-330E (Serial Number starts with "XQ-"),
BS-350E (Serial Number starts with "XS-"); | 18. BS-830 : BS-830, BS-830S,BS-840, BS-850, BS-860,
BS-870; |
| | 19. BS-2000 : BS-2000, BS-2200, BS-2000M, BS-2200M. |
| | 20. BS-2800M : BS-2600M, BS-2800M. |

Abbreviated name	TBA Control(L)					TBA Control(H)				
	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD
TBA (apply to lot 061821003, 061821004)	BS-120	¹ μmol/L	20.7	16.5 — 24.9	1.4	BS-120	¹ μmol/L	39.5	31.6 — 47.4	2.6
	BS-180	² μmol/L	20.7	16.5 — 24.9	1.4	BS-180	² μmol/L	39.5	31.6 — 47.4	2.6
	BS-200	³ μmol/L	20.1	16.1 — 24.1	1.3	BS-200	³ μmol/L	39.3	31.4 — 47.2	2.6
	BS-200E	⁴ μmol/L	20.1	16.1 — 24.1	1.3	BS-200E	⁴ μmol/L	40.4	32.3 — 48.5	2.7
	BS-230	⁵ μmol/L	20.1	16.1 — 24.1	1.3	BS-230	⁵ μmol/L	38.8	31.0 — 46.6	2.6
	BS-240E	⁶ μmol/L	19.3	15.4 — 23.2	1.3	BS-240E	⁶ μmol/L	37.3	29.8 — 44.8	2.5
	BS-300	⁷ μmol/L	20.5	16.4 — 24.6	1.4	BS-300	⁷ μmol/L	39.3	31.4 — 47.2	2.6
	BS-330	⁸ μmol/L	20.1	16.1 — 24.1	1.3	BS-330	⁸ μmol/L	39.3	31.4 — 47.2	2.6
	BS-330E	⁹ μmol/L	20.1	16.1 — 24.1	1.3	BS-330E	⁹ μmol/L	40.4	32.3 — 48.5	2.7
	BS-360E	¹⁰ μmol/L	20.2	16.1 — 24.3	1.4	BS-360E	¹⁰ μmol/L	38.2	30.5 — 45.9	2.6
	BS-380	¹¹ μmol/L	20.1	16.1 — 24.1	1.3	BS-380	¹¹ μmol/L	39.1	31.2 — 47.0	2.6
	BS-400	¹² μmol/L	20.5	16.4 — 24.6	1.4	BS-400	¹² μmol/L	38.7	30.9 — 46.5	2.6
	BS-430	¹³ μmol/L	20.6	16.5 — 24.7	1.4	BS-430	¹³ μmol/L	39.2	31.3 — 47.1	2.6
	BS-480	¹⁴ μmol/L	20.4	16.3 — 24.5	1.4	BS-480	¹⁴ μmol/L	39.3	31.4 — 47.2	2.6
	BS-600	¹⁵ μmol/L	20.9	16.7 — 25.1	1.4	BS-600	¹⁵ μmol/L	39.8	31.8 — 47.8	2.7
	BS-600M	¹⁶ μmol/L	20.7	16.5 — 24.9	1.4	BS-600M	¹⁶ μmol/L	39.2	31.3 — 47.1	2.6
	BS-800	¹⁷ μmol/L	20.5	16.4 — 24.6	1.4	BS-800	¹⁷ μmol/L	39.3	31.4 — 47.2	2.6
	BS-830	¹⁸ μmol/L	21.2	16.9 — 25.5	1.4	BS-830	¹⁸ μmol/L	40.2	32.1 — 48.3	2.7
	BS-2000	¹⁹ μmol/L	21.3	17.0 — 25.6	1.4	BS-2000	¹⁹ μmol/L	40.6	32.4 — 48.8	2.7
	BS-2800M	²⁰ μmol/L	21.4	17.1 — 25.7	1.4	BS-2800M	²⁰ μmol/L	40.7	32.5 — 48.9	2.7

CO2 and TBA Multi Control

Abbreviated name	TBA Control(L)						TBA Control(H)					
	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD
TBA (apply to lot 061822001)	BS-120	¹	μmol/L	20.7	16.5 — 24.9	1.4	BS-120	¹	μmol/L	37.9	30.3 — 45.5	2.5
	BS-180	²	μmol/L	20.7	16.5 — 24.9	1.4	BS-180	²	μmol/L	37.9	30.3 — 45.5	2.5
	BS-200	³	μmol/L	20.1	16.1 — 24.1	1.3	BS-200	³	μmol/L	37.2	29.7 — 44.7	2.5
	BS-200E	⁴	μmol/L	20.1	16.1 — 24.1	1.3	BS-200E	⁴	μmol/L	37.4	29.9 — 44.9	2.5
	BS-230	⁵	μmol/L	20.1	16.1 — 24.1	1.3	BS-230	⁵	μmol/L	37.0	29.6 — 44.4	2.5
	BS-240E	⁶	μmol/L	19.3	15.4 — 23.2	1.3	BS-240E	⁶	μmol/L	35.9	28.7 — 43.1	2.4
	BS-300	⁷	μmol/L	20.5	16.4 — 24.6	1.4	BS-300	⁷	μmol/L	37.2	29.7 — 44.7	2.5
	BS-330	⁸	μmol/L	20.1	16.1 — 24.1	1.3	BS-330	⁸	μmol/L	37.2	29.7 — 44.7	2.5
	BS-330E	⁹	μmol/L	20.1	16.1 — 24.1	1.3	BS-330E	⁹	μmol/L	37.4	29.9 — 44.9	2.5
	BS-360E	¹⁰	μmol/L	20.2	16.1 — 24.3	1.4	BS-360E	¹⁰	μmol/L	36.2	28.9 — 43.5	2.4
	BS-380	¹¹	μmol/L	20.1	16.1 — 24.1	1.3	BS-380	¹¹	μmol/L	37.1	29.6 — 44.6	2.5
	BS-400	¹²	μmol/L	20.5	16.4 — 24.6	1.4	BS-400	¹²	μmol/L	36.8	29.4 — 44.2	2.5
	BS-430	¹³	μmol/L	20.6	16.5 — 24.7	1.4	BS-430	¹³	μmol/L	37.5	30.0 — 45.0	2.5
	BS-480	¹⁴	μmol/L	20.4	16.3 — 24.5	1.4	BS-480	¹⁴	μmol/L	37.1	29.6 — 44.6	2.5
	BS-600	¹⁵	μmol/L	20.9	16.7 — 25.1	1.4	BS-600	¹⁵	μmol/L	37.8	30.2 — 45.4	2.5
	BS-600M	¹⁶	μmol/L	20.7	16.5 — 24.9	1.4	BS-600M	¹⁶	μmol/L	37.5	30.0 — 45.0	2.5
	BS-800	¹⁷	μmol/L	20.5	16.4 — 24.6	1.4	BS-800	¹⁷	μmol/L	37.3	29.8 — 44.8	2.5
	BS-830	¹⁸	μmol/L	21.2	16.9 — 25.5	1.4	BS-830	¹⁸	μmol/L	38.4	30.7 — 46.1	2.6
	BS-2000	¹⁹	μmol/L	21.3	17.0 — 25.6	1.4	BS-2000	¹⁹	μmol/L	38.7	30.9 — 46.5	2.6
	BS-2800M	²⁰	μmol/L	21.4	17.1 — 25.7	1.4	BS-2800M	²⁰	μmol/L	38.7	30.9 — 46.5	2.6

Abbreviated name	TBA Control(L)						TBA Control(H)					
	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD
TBA (apply to lot 061822002)	BS-120	¹	μmol/L	20.7	16.5 — 24.9	1.4	BS-120	¹	μmol/L	36.2	28.9 — 43.5	2.4
	BS-180	²	μmol/L	20.7	16.5 — 24.9	1.4	BS-180	²	μmol/L	36.2	28.9 — 43.5	2.4
	BS-200	³	μmol/L	20.1	16.1 — 24.1	1.3	BS-200	³	μmol/L	36.3	29.0 — 43.6	2.4
	BS-200E	⁴	μmol/L	20.1	16.1 — 24.1	1.3	BS-200E	⁴	μmol/L	37.7	30.1 — 45.3	2.5
	BS-230	⁵	μmol/L	20.1	16.1 — 24.1	1.3	BS-230	⁵	μmol/L	35.5	28.4 — 42.6	2.4
	BS-240E	⁶	μmol/L	19.3	15.4 — 23.2	1.3	BS-240E	⁶	μmol/L	34.3	27.4 — 41.2	2.3
	BS-300	⁷	μmol/L	20.5	16.4 — 24.6	1.4	BS-300	⁷	μmol/L	36.9	29.5 — 44.3	2.5
	BS-330	⁸	μmol/L	20.1	16.1 — 24.1	1.3	BS-330	⁸	μmol/L	36.3	29.0 — 43.6	2.4
	BS-330E	⁹	μmol/L	20.1	16.1 — 24.1	1.3	BS-330E	⁹	μmol/L	37.7	30.1 — 45.3	2.5
	BS-360E	¹⁰	μmol/L	20.2	16.1 — 24.3	1.4	BS-360E	¹⁰	μmol/L	35.0	28.0 — 42.0	2.3
	BS-380	¹¹	μmol/L	20.1	16.1 — 24.1	1.3	BS-380	¹¹	μmol/L	35.7	28.5 — 42.9	2.4
	BS-400	¹²	μmol/L	20.5	16.4 — 24.6	1.4	BS-400	¹²	μmol/L	35.5	28.4 — 42.6	2.4
	BS-430	¹³	μmol/L	20.6	16.5 — 24.7	1.4	BS-430	¹³	μmol/L	35.9	28.7 — 43.1	2.4
	BS-480	¹⁴	μmol/L	20.4	16.3 — 24.5	1.4	BS-480	¹⁴	μmol/L	35.4	28.3 — 42.5	2.4
	BS-600	¹⁵	μmol/L	20.9	16.7 — 25.1	1.4	BS-600	¹⁵	μmol/L	36.4	29.1 — 43.7	2.4
	BS-600M	¹⁶	μmol/L	20.7	16.5 — 24.9	1.4	BS-600M	¹⁶	μmol/L	36.0	28.8 — 43.2	2.4
	BS-800	¹⁷	μmol/L	20.5	16.4 — 24.6	1.4	BS-800	¹⁷	μmol/L	35.9	28.7 — 43.1	2.4
	BS-830	¹⁸	μmol/L	21.2	16.9 — 25.5	1.4	BS-830	¹⁸	μmol/L	36.8	29.4 — 44.2	2.5
	BS-2000	¹⁹	μmol/L	21.3	17.0 — 25.6	1.4	BS-2000	¹⁹	μmol/L	37.3	29.8 — 44.8	2.5
	BS-2800M	²⁰	μmol/L	21.4	17.1 — 25.7	1.4	BS-2800M	²⁰	μmol/L	37.3	29.8 — 44.8	2.5

CO2 and TBA Multi Control

Abbreviated name	TBA Control(L)						TBA Control(H)					
	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD
TBA (apply to lot 061822003、061822004、061822005)	BS-120	¹	μmol/L	21.5	17.2 — 25.8	1.4	BS-120	¹	μmol/L	36.2	28.9 — 43.5	2.4
	BS-180	²	μmol/L	21.5	17.2 — 25.8	1.4	BS-180	²	μmol/L	36.2	28.9 — 43.5	2.4
	BS-200	³	μmol/L	20.8	16.6 — 25.0	1.4	BS-200	³	μmol/L	36.3	29.0 — 43.6	2.4
	BS-200E	⁴	μmol/L	21.0	16.8 — 25.2	1.4	BS-200E	⁴	μmol/L	37.7	30.1 — 45.3	2.5
	BS-230	⁵	μmol/L	21.0	16.8 — 25.2	1.4	BS-230	⁵	μmol/L	35.5	28.4 — 42.6	2.4
	BS-240E	⁶	μmol/L	20.3	16.2 — 24.4	1.4	BS-240E	⁶	μmol/L	34.3	27.4 — 41.2	2.3
	BS-300	⁷	μmol/L	21.3	17.0 — 25.6	1.4	BS-300	⁷	μmol/L	36.9	29.5 — 44.3	2.5
	BS-330	⁸	μmol/L	20.8	16.6 — 25.0	1.4	BS-330	⁸	μmol/L	36.3	29.0 — 43.6	2.4
	BS-330E	⁹	μmol/L	21.0	16.8 — 25.2	1.4	BS-330E	⁹	μmol/L	37.7	30.1 — 45.3	2.5
	BS-360E	¹⁰	μmol/L	20.9	16.7 — 25.1	1.4	BS-360E	¹⁰	μmol/L	35.0	28.0 — 42.0	2.3
	BS-380	¹¹	μmol/L	21.3	17.0 — 25.6	1.4	BS-380	¹¹	μmol/L	35.7	28.5 — 42.9	2.4
	BS-400	¹²	μmol/L	21.6	17.3 — 25.9	1.4	BS-400	¹²	μmol/L	35.5	28.4 — 42.6	2.4
	BS-430	¹³	μmol/L	21.6	17.3 — 25.9	1.4	BS-430	¹³	μmol/L	35.9	28.7 — 43.1	2.4
	BS-480	¹⁴	μmol/L	21.6	17.3 — 25.9	1.4	BS-480	¹⁴	μmol/L	35.4	28.3 — 42.5	2.4
	BS-600	¹⁵	μmol/L	21.9	17.5 — 26.3	1.5	BS-600	¹⁵	μmol/L	36.4	29.1 — 43.7	2.4
	BS-600M	¹⁶	μmol/L	21.6	17.3 — 25.9	1.4	BS-600M	¹⁶	μmol/L	36.0	28.8 — 43.2	2.4
	BS-800	¹⁷	μmol/L	21.4	17.1 — 25.7	1.4	BS-800	¹⁷	μmol/L	35.9	28.7 — 43.1	2.4
	BS-830	¹⁸	μmol/L	22.1	17.7 — 26.5	1.5	BS-830	¹⁸	μmol/L	36.8	29.4 — 44.2	2.5
	BS-2000	¹⁹	μmol/L	22.3	17.8 — 26.8	1.5	BS-2000	¹⁹	μmol/L	37.3	29.8 — 44.8	2.5
	BS-2800M	²⁰	μmol/L	22.3	17.8 — 26.8	1.5	BS-2800M	²⁰	μmol/L	37.3	29.8 — 44.8	2.5

Abbreviated name	TBA Control(L)						TBA Control(H)					
	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD
TBA (apply to lot 061822006)	BS-120	¹	μmol/L	21.5	17.2 — 25.8	1.4	BS-120	¹	μmol/L	36.1	28.8 — 43.4	2.4
	BS-180	²	μmol/L	21.5	17.2 — 25.8	1.4	BS-180	²	μmol/L	36.1	28.8 — 43.4	2.4
	BS-200	³	μmol/L	20.8	16.6 — 25.0	1.4	BS-200	³	μmol/L	35.7	28.5 — 42.9	2.4
	BS-200E	⁴	μmol/L	21.0	16.8 — 25.2	1.4	BS-200E	⁴	μmol/L	38.2	30.5 — 45.9	2.6
	BS-230	⁵	μmol/L	21.0	16.8 — 25.2	1.4	BS-230	⁵	μmol/L	35.6	28.4 — 42.8	2.4
	BS-240E	⁶	μmol/L	20.3	16.2 — 24.4	1.4	BS-240E	⁶	μmol/L	34.3	27.4 — 41.2	2.3
	BS-300	⁷	μmol/L	21.3	17.0 — 25.6	1.4	BS-300	⁷	μmol/L	37.4	29.9 — 44.9	2.5
	BS-330	⁸	μmol/L	20.8	16.6 — 25.0	1.4	BS-330	⁸	μmol/L	35.7	28.5 — 42.9	2.4
	BS-330E	⁹	μmol/L	21.0	16.8 — 25.2	1.4	BS-330E	⁹	μmol/L	38.2	30.5 — 45.9	2.6
	BS-360E	¹⁰	μmol/L	20.9	16.7 — 25.1	1.4	BS-360E	¹⁰	μmol/L	35.3	28.2 — 42.4	2.4
	BS-380	¹¹	μmol/L	21.3	17.0 — 25.6	1.4	BS-380	¹¹	μmol/L	35.6	28.4 — 42.8	2.4
	BS-400	¹²	μmol/L	21.6	17.3 — 25.9	1.4	BS-400	¹²	μmol/L	35.6	28.4 — 42.8	2.4
	BS-430	¹³	μmol/L	21.6	17.3 — 25.9	1.4	BS-430	¹³	μmol/L	35.9	28.7 — 43.1	2.4
	BS-480	¹⁴	μmol/L	21.6	17.3 — 25.9	1.4	BS-480	¹⁴	μmol/L	35.7	28.5 — 42.9	2.4
	BS-600	¹⁵	μmol/L	21.9	17.5 — 26.3	1.5	BS-600	¹⁵	μmol/L	36.1	28.8 — 43.4	2.4
	BS-600M	¹⁶	μmol/L	21.6	17.3 — 25.9	1.4	BS-600M	¹⁶	μmol/L	35.8	28.6 — 43.0	2.4
	BS-800	¹⁷	μmol/L	21.4	17.1 — 25.7	1.4	BS-800	¹⁷	μmol/L	35.9	28.7 — 43.1	2.4
	BS-830	¹⁸	μmol/L	22.1	17.7 — 26.5	1.5	BS-830	¹⁸	μmol/L	36.6	29.2 — 44.0	2.5
	BS-2000	¹⁹	μmol/L	22.3	17.8 — 26.8	1.5	BS-2000	¹⁹	μmol/L	37.3	29.8 — 44.8	2.5
	BS-2800M	²⁰	μmol/L	22.3	17.8 — 26.8	1.5	BS-2800M	²⁰	μmol/L	37.3	29.8 — 44.8	2.5

CO2 and TBA Multi Control

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа \pm 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise \pm 3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo \pm 3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione \pm 3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri \pm 3SD)
TBA					
English	Total Bile Acids		Español	ácidos biliars totais	
Русский	общие желчные кислоты		Italiano	acidi biliari totali	
Português	ácidos biliars totais		Türkçe	total safra asitleridir	