

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120:** BS-120, BS-130;

10. **BS-360E:** BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

2. **BS-180:** BS-180, BS-190;

11. **BS-380:** BS-380, BS-390;

3. **BS-200:** BS-200, BS-220;

12. **BS-400:** BS-400, BS-420;

4. **BS-200E:** BS-200E, BS-220E;

13. **BS-430:** BS-430, BS-450, BS-460;

5. **BS-230:** BS-230, BS-240;

14. **BS-480:** BS-480, BS-490;

6. **BS-240E:** BS-240E, BS-240Pro;

15. **BS-600:** BS-600, BS-620;

7. **BS-300:** BS-300, BS-320;

16. **BS-600M:** BS-600M, BS-620M;

8. **BS-330:** BS-330, BS-350;

17. **BS-800:** BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

9. **BS-330E:** BS-330E(Serial Number starts with "XQ-"),

18. **BS-2000:** BS-2000, BS-2200, BS-2000M, BS-2200M;

BS-350E(Serial Number starts with "XS-");

19. **BS-2800M:** BS-2600M, BS-2800M.

20. **S0: 0.9% NaCl, Conc. Of S0=0;**

Русский: S0: 0, 9% NaCl, конц. S0=0;

Português: S0: 0,9% NaCl, Conc. de S0=0;

Español: S0: 0,9% NaCl, Conc. de S0=0;

Italiano: S0: 0,9% NaCl, conc. di S0=0;

Türkçe: S0: %0,9 NaCl, S0 Kons.=0.

LOT : 145323009

: 2024-12-14

Abbreviated name			Fe		
Model	Unit	Calibration Value ²⁰	Model	Unit	Calibration Value ²⁰
BS-120 ¹	µmol/L	41.0	BS-380 ¹¹	µmol/L	43.1
	mg/L	2.30		mg/L	2.41
BS-180 ²	µmol/L	41.0	BS-400 ¹²	µmol/L	42.4
	mg/L	2.30		mg/L	2.37
BS-200 ³	µmol/L	41.3	BS-430 ¹³	µmol/L	43.1
	mg/L	2.31		mg/L	2.41
BS-200E ⁴	µmol/L	43.1	BS-480 ¹⁴	µmol/L	43.1
	mg/L	2.41		mg/L	2.41
BS-230 ⁵	µmol/L	41.0	BS-600 ¹⁵	µmol/L	43.1
	mg/L	2.30		mg/L	2.41
BS-240E ⁶	µmol/L	43.1	BS-600M ¹⁶	µmol/L	43.1
	mg/L	2.41		mg/L	2.41
BS-300 ⁷	µmol/L	41.8	BS-800 ¹⁷	µmol/L	43.1
	mg/L	2.34		mg/L	2.41
BS-330 ⁸	µmol/L	41.3	BS-2000 ¹⁸	µmol/L	43.2
	mg/L	2.31		mg/L	2.42
BS-330E ⁹	µmol/L	43.1	BS-2800M ¹⁹	µmol/L	43.7
	mg/L	2.41		mg/L	2.45
BS-360E ¹⁰	µmol/L	43.1			
	mg/L	2.41			

English	Abbreviated name	Model	Unit	Calibration Value
Русский	сокращенное наименование	модель	Прибор	Принцип калибровки
Português	Nome abreviado	Modelo	Unidade	Regra de calibração
Español	nombre abreviado	modelo	Unidad	Regla de calibración
Italiano	abbreviazione	modelli	Unità	Regola di calibrazione
Türkçe	kısaltılmış ad	model	Ünite	Kalibrasyon Kuralı

Fe	
English	Iron
Русский	Железо
Português	Ferro
Español	Hierro
Italiano	Ferro
Türkçe	Demir

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

1. **BS-120**: BS-120, BS-130;

2. **BS-180**: BS-180, BS-190;

3. **BS-200**: BS-200, BS-220;

4. **BS-200E**: BS-200E, BS-220E;

5. **BS-230**: BS-230, BS-240;

6. **BS-240E**: BS-240E, BS-240Pro;

7. **BS-300**: BS-300, BS-320;

8. **BS-330**: BS-330, BS-350;

9. **BS-330E**: BS-330E (Serial Number starts with "XQ-"),

BS-350E (Serial Number starts with "XS-");

10. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

11. **BS-380**: BS-380, BS-390;

12. **BS-400**: BS-400, BS-420;

13. **BS-430**: BS-430, BS-450, BS-460;

14. **BS-480**: BS-480, BS-490;

15. **BS-600**: BS-600, BS-620;

16. **BS-600M**: BS-600M, BS-620M;

17. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

18. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

19. **BS-2800M**: BS-2600M, BS-2800M.

LOT : 145323009

EXP : 2024-12-14

Abbreviated name						Fe							
Model	Unit	Assay Value	Range(Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range(Assay Value \pm 3SD)		1 SD		
BS-120 ¹	μ mol/L	19.3	15.4	—	23.2	1.3	BS-380 ¹¹	μ mol/L	20.2	16.0	—	24.4	1.4
	mg/L	1.08	0.86	—	1.30	0.07		mg/L	1.13	0.89	—	1.36	0.08
BS-180 ¹	μ mol/L	19.3	15.4	—	23.2	1.3	BS-400 ¹²	μ mol/L	19.9	16.0	—	23.8	1.3
	mg/L	1.08	0.86	—	1.30	0.07		mg/L	1.11	0.89	—	1.33	0.07
BS-200 ³	μ mol/L	19.8	15.9	—	23.7	1.3	BS-430 ¹³	μ mol/L	19.6	15.7	—	23.5	1.3
	mg/L	1.11	0.89	—	1.32	0.07		mg/L	1.09	0.88	—	1.31	0.07
BS-200E ⁴	μ mol/L	19.8	15.9	—	23.7	1.3	BS-480 ¹⁴	μ mol/L	19.8	15.9	—	23.7	1.3
	mg/L	1.11	0.89	—	1.32	0.07		mg/L	1.11	0.89	—	1.32	0.07
BS-230 ⁵	μ mol/L	19.3	15.4	—	23.2	1.3	BS-600 ¹⁵	μ mol/L	19.6	15.7	—	23.5	1.3
	mg/L	1.08	0.86	—	1.30	0.07		mg/L	1.09	0.88	—	1.31	0.07
BS-240E ⁶	μ mol/L	19.7	15.8	—	23.6	1.3	BS-600M ¹⁶	μ mol/L	19.6	15.7	—	23.5	1.3
	mg/L	1.10	0.88	—	1.32	0.07		mg/L	1.09	0.88	—	1.31	0.07
BS-300 ⁷	μ mol/L	19.7	15.8	—	23.6	1.3	BS-800 ¹⁷	μ mol/L	19.6	15.7	—	23.5	1.3
	mg/L	1.10	0.88	—	1.32	0.07		mg/L	1.09	0.88	—	1.31	0.07
BS-330 ⁸	μ mol/L	19.8	15.9	—	23.7	1.3	BS-2000 ¹⁸	μ mol/L	19.5	15.6	—	23.4	1.3
	mg/L	1.11	0.89	—	1.32	0.07		mg/L	1.09	0.87	—	1.31	0.07
BS-330E ⁹	μ mol/L	19.8	15.9	—	23.7	1.3	BS-2800M ¹⁹	μ mol/L	19.6	15.7	—	23.5	1.3
	mg/L	1.11	0.89	—	1.32	0.07		mg/L	1.09	0.88	—	1.31	0.07
BS-360E ¹⁰	μ mol/L	19.9	16.0	—	23.8	1.3							
	mg/L	1.11	0.89	—	1.33	0.07							

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа \pm 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise \pm 3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo \pm 3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione \pm 3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri \pm 3SD)

Fe	
English	Iron
Русский	Железо
Português	Ferro
Español	Hierro
Italiano	Ferro
Türkçe	Demir