

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;

9.BS-600: BS-600, BS-620;

2.BS-240: BS-230, BS-240;

10.BS-600M: BS-600M, BS-620M;

3.BS-240E: BS240E, BS240Pro;

4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12.BS-1000M: BS-1000M, BS-1100M;

5.BS-380: BS-380, BS-390;

13.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

6.BS-400: BS-400, BS-420;

14.BS-2800M:BS-2600M, BS-2800M;

7.BS-430: BS-430, BS-450,BS-460;

15.For applicable models of the analyte, please refer to the

8.BS-480: BS-480, BS-490;

reagent parameter sheet and instruction.

LOT : 147224002

: 2025-09-10

| English | Abbreviated name | Calibration Value | Level | Model |
|-----------|--------------------------|------------------------|---------|---------|
| Русский | сокращенное наименование | Эталонное значение | Уровень | модель |
| Português | Nome abreviado | Valor de calibração | Nível | Modelo |
| Español | nombre abreviado | Valor de calibración | Nivel | modelo |
| Italiano | abbreviazione | Valore di calibrazione | Livello | modelli |
| Türkçe | kisaltılmış ad | Kalibrasyon Değeri | Düzey | model |

| English | Expanded Uncertainty |
|-----------|------------------------------|
| Русский | Расширенная неопределенность |
| Português | Incerteza Expandida |
| Español | Incertidumbre expandida |
| Italiano | Incertezza estesa |
| Türkçe | Daha Uzun Süreli Belirsizlik |

| English | CysC |
|-----------|-------------|
| Русский | цистатин C |
| Português | Cistatina C |
| Español | cistatina C |
| Italiano | cistatina C |
| Türkçe | Cystatin C |

| Abbreviated name | CysC II | Calibration Rule | | spline | | | | | |
|------------------|----------------------|---------------------|----------------------|----------------------|----------------------|----------------------|----------------------|------------------|----------------------|
| Model | BS-200E ¹ | BS-240 ² | | BS-240E ³ | | BS-360E ⁴ | | | |
| Level | Unit | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty |
| 0.9%NaCl | | 0.000 | / | 0.000 | / | 0.000 | / | 0.000 | / |
| a | mg/L | 0.530 | 0.022 | 0.490 | 0.020 | 0.600 | 0.025 | 0.530 | 0.022 |
| b | | 1.07 | 0.027 | 1.04 | 0.026 | 1.06 | 0.027 | 1.07 | 0.027 |
| c | | 2.13 | 0.05 | 2.13 | 0.05 | 2.11 | 0.04 | 2.13 | 0.05 |

Cystatin C Calibrator




| d | | 4.26 | 0.16 | 4.26 | 0.16 | 4.22 | 0.16 | 4.26 | 0.16 |
|-----------------|------|-----------------------|----------------------|------------------------|----------------------|----------------------|----------------------|------------------------|----------------------|
| e | | 8.54 | 0.19 | 8.54 | 0.19 | 8.54 | 0.19 | 8.54 | 0.19 |
| Model | | BS-380 ⁵ | | BS-400 ⁶ | | BS-430 ⁷ | | BS-480 ⁸ | |
| Level | Unit | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty |
| 0.9%NaCl | | 0.000 | / | 0.000 | / | 0.000 | / | 0.000 | / |
| a | mg/L | 0.530 | 0.022 | 0.530 | 0.022 | 0.530 | 0.022 | 0.530 | 0.022 |
| b | | 1.07 | 0.027 | 1.07 | 0.027 | 1.07 | 0.027 | 1.07 | 0.027 |
| c | | 2.13 | 0.05 | 2.13 | 0.05 | 2.13 | 0.05 | 2.13 | 0.05 |
| d | | 4.26 | 0.16 | 4.26 | 0.16 | 4.26 | 0.16 | 4.26 | 0.16 |
| e | | 8.54 | 0.19 | 8.54 | 0.19 | 8.54 | 0.19 | 8.54 | 0.19 |
| Model | | BS-600 ⁹ | | BS-600M ¹⁰ | | BS-800 ¹¹ | | BS-1000M ¹² | |
| Level | Unit | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty |
| 0.9%NaCl | | 0.000 | / | 0.000 | / | 0.000 | / | 0.000 | / |
| a | mg/L | 0.530 | 0.022 | 0.560 | 0.023 | 0.530 | 0.022 | 0.560 | 0.028 |
| b | | 1.07 | 0.027 | 1.07 | 0.027 | 1.07 | 0.027 | 1.07 | 0.027 |
| c | | 2.13 | 0.05 | 2.11 | 0.04 | 2.13 | 0.05 | 2.11 | 0.04 |
| d | | 4.26 | 0.16 | 4.29 | 0.17 | 4.26 | 0.16 | 4.29 | 0.17 |
| e | | 8.54 | 0.19 | 8.51 | 0.19 | 8.54 | 0.19 | 8.51 | 0.19 |
| Model | | BS-2000 ¹³ | | BS-2800M ¹⁴ | | | | | |
| Level | Unit | Calibrator Value | Expanded Uncertainty | Calibrator Value | Expanded Uncertainty | | | | |
| 0.9%NaCl | | 0.000 | / | 0.000 | / | | | | |
| a | mg/L | 0.540 | 0.021 | 0.560 | 0.028 | | | | |
| b | | 1.07 | 0.027 | 1.07 | 0.027 | | | | |
| c | | 2.12 | 0.04 | 2.11 | 0.04 | | | | |
| d | | 4.27 | 0.17 | 4.29 | 0.17 | | | | |
| e | | 8.54 | 0.19 | 8.51 | 0.19 | | | | |

mindray

Cystatin C Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 147224002

 2025-09-10

