

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120**: BS-120, BS-130;

11. **BS-380**: BS-380, BS-390;

2. **BS-180**: BS-180, BS-190;

12. **BS-400**: BS-400, BS-420;

3. **BS-200**: BS-200, BS-220;

13. **BS-430**: BS-430, BS-450, BS-460;

4. **BS-200E**: BS-200E, BS-220E;

14. **BS-480**: BS-480, BS-490;

5. **BS-230**: BS-230, BS-240;

15. **BS-600**: BS-600, BS-620;

6. **BS-240E**: BS-240E, BS-240Pro;

16. **BS-600M**: BS-600M, BS-620M;

7. **BS-300**: BS-300, BS-320;

17. **BS-800**: BS-800, BS-820, BS-800M,

8. **BS-330**: BS-330, BS-350;

BS-820M, BS-1800, BS-1800plus;

9. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),

BS-350E(Serial Number starts with "XS-");

18. **BS-2000**: BS-2000, BS-2200,

10. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**

(Serial Number starts with "W8-" and software version starts with

BS-2000M, BS-2200M;

"35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and

software version starts with "35.00");

19. **BS-2800M**: BS-2800M, BS-2600M.

20. **S1:0.9% NaCl, Conc. Of S1=0;**

Русский: S1: 0,9% NaCl, конц. S1=0;

Português: S1:0,9% NaCl, Conc. de S1=0;

Español: S1:0,9% NaCl, Conc. de S1=0;

Italiano: S1:0,9% NaCl, conc. di S1=0;

Türkçe: S1:%0,9 NaCl, S1 Kons.=0.

LOT: 150423011 : 2025-05-06

Abbreviated name	Unit	Calibration Value ²⁰					
mmol/L		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
		1.73	1.73	1.75	1.79	1.73	1.79
		BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		1.79	1.75	1.79	1.79	1.79	1.79
		BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		1.79	1.79	1.79	1.79	1.79	1.77
HDL-C		BS-2800M ¹⁹					
		1.83					
		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
		66.9	66.9	67.7	69.2	66.9	69.2
		BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		69.2	67.7	69.2	69.2	69.2	69.2
mg/dL		BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		69.2	69.2	69.2	69.2	69.2	68.4
		BS-2800M ¹⁹					
		70.7					
		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
		3.83	3.83	3.83	3.95	3.83	3.95
LDL-C		BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		3.95	3.83	3.95	3.95	3.95	3.95
		BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		3.95	3.95	3.95	3.96	3.95	3.88
		BS-2800M ¹⁹					
		3.96					

Lipids Calibrator



Abbreviated name	Unit	Calibration Value ²⁰					
		BS-120 ¹	BS-180 ²	BS-200 ³	BS-200E ⁴	BS-230 ⁵	BS-240E ⁶
LDL-C	mg/dL	148	148	148	153	148	153
		BS-300 ⁷	BS-330 ⁸	BS-330E ⁹	BS-360E ¹⁰	BS-380 ¹¹	BS-400 ¹²
		153	148	153	153	153	153
		BS-430 ¹³	BS-480 ¹⁴	BS-600 ¹⁵	BS-600M ¹⁶	BS-800 ¹⁷	BS-2000 ¹⁸
		153	153	153	153	153	150
		BS-2800M ¹⁹	153				

Abbreviated name		ApoA1	Calibration Rule		Logit-Log(5P)	
Model	Level	Calibrator Value ²⁰		Sample Vol for Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		g/L	μmol/L			
BS-120 ¹	S2	0.040	1.43	20	180	3
	S3	0.350	12.5	5	194	32
	S4	0.920	32.8	9	187	35
	S5	1.76	62.8	45	180	12
	S6	3.11	111	/	/	3.5
	BS-180 ²	S2	0.040	1.43	20	180
S3		0.350	12.5	5	194	32
S4		0.920	32.8	9	187	35
S5		1.76	62.8	45	180	12
S6		3.11	111	/	/	3.5
BS-200 ³		S2	0.050	1.79	20	180
	S3	0.340	12.1	5	194	32
	S4	0.960	34.3	9	187	35
	S5	1.78	63.5	45	180	12
	S6	3.11	111	/	/	3.5
	BS-200E ⁴	S2	0.050	1.79	15	135
S3		0.560	20.0	45	135	2
S4		1.15	41.1	45	135	4
S5		1.87	66.8	35	140	8
S6		3.11	111	/	/	2.3
BS-300 ⁷		S2	0.050	1.79	20	180
	S3	0.440	15.7	5	194	32
	S4	1.09	38.9	9	187	35
	S5	1.98	70.7	45	180	12
	S6	3.11	111	/	/	3.5
	BS-330 ⁸	S2	0.050	1.79	20	180
S3		0.340	12.1	5	194	32
S4		0.960	34.3	9	187	35
S5		1.78	63.5	45	180	12
S6		3.11	111	/	/	3.5
BS-330E ⁹		S2	0.050	1.79	15	135
	S3	0.560	20.0	45	135	2
	S4	1.15	41.1	45	135	4
	S5	1.87	66.8	35	140	8
	S6	3.11	111	/	/	2.3

Lipids Calibrator



BS-380 ¹¹	S2	0.050	1.79	14	126	2
	S3	0.560	20.0	40	120	2
	S4	1.15	41.1	40	120	4
	S5	1.87	66.8	30	120	8
	S6	3.11	111	/	/	2.3
	BS-400 ¹²	S2	0.060	2.14	14	126
S3		0.540	19.3	40	120	2
S4		1.24	44.3	40	120	4
S5		1.95	69.6	30	120	8
S6		3.11	111	/	/	2.3
Abbreviated name		ApoB	Calibration Rule		Logit-Log(5P)	
Model	Level	Calibrator Value²⁰		Sample Vol for Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		g/L	μmol/L			
BS-120 ¹	S2	0.210	0.410	45	180	3
	S3	0.470	0.917	45	180	6
	S4	0.960	1.87	45	180	12
	S5	1.40	2.73	/	/	3
	S6	2.70	5.27	/	/	6
	BS-180 ²	S2	0.210	0.410	45	180
S3		0.470	0.917	45	180	6
S4		0.960	1.87	45	180	12
S5		1.40	2.73	/	/	3
S6		2.70	5.27	/	/	6
BS-200 ³		S2	0.210	0.410	45	180
	S3	0.530	1.03	45	180	6
	S4	1.00	1.95	45	180	12
	S5	1.41	2.75	/	/	3
	S6	2.70	5.27	/	/	6
	BS-200E ⁴	S2	0.220	0.429	35	140
S3		0.500	0.975	35	140	6
S4		1.00	1.95	35	140	12
S5		1.39	2.71	/	/	3
S6		2.70	5.27	/	/	6
BS-300 ⁷		S2	0.200	0.390	45	180
	S3	0.500	0.975	45	180	6
	S4	1.04	2.03	45	180	12
	S5	1.40	2.73	/	/	3
	S6	2.70	5.27	/	/	6
	BS-330 ⁸	S2	0.210	0.410	45	180
S3		0.530	1.03	45	180	6
S4		1.00	1.95	45	180	12
S5		1.41	2.75	/	/	3
S6		2.70	5.27	/	/	6
BS-330E ⁹		S2	0.220	0.429	35	140
	S3	0.500	0.975	35	140	6
	S4	1.00	1.95	35	140	12
	S5	1.39	2.71	/	/	3
	S6	2.70	5.27	/	/	6

Lipids Calibrator



BS-380 ¹¹	S2	0.220	0.429	30	120	2
	S3	0.500	0.975	30	120	4
	S4	1.00	1.95	30	120	8
	S5	1.39	2.71	/	/	2
	S6	2.70	5.27	/	/	4
	BS-400 ¹²	S2	0.220	0.429	30	120
S3		0.500	0.975	30	120	4
S4		0.990	1.93	30	120	8
S5		1.38	2.69	/	/	2
S6		2.70	5.27	/	/	4
Abbreviated name		ApoA1	Calibration Rule		Logit-Log(5P)	
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)
		g/L	μmol/L			
BS-230 ⁵	S2	0.050	1.79	2	13	117
	S3	0.410	14.6	23.8	3	129
	S4	1.02	36.4	23.4	6	125
	S5	1.79	63.9	8	30	120
	S6	3.11	111	2.3	/	/
	BS-240E ⁶	S2	0.050	1.79	2	13
S3		0.560	20.0	2	34	102
S4		1.15	41.1	4	34	102
S5		1.87	66.8	8	25	100
S6		3.11	111	2.3	/	/
BS-360E ¹⁰		S2	0.050	1.79	2	13
	S3	0.560	20.0	2	34	102
	S4	1.15	41.1	4	34	102
	S5	1.87	66.8	8	25	100
	S6	3.11	111	2.3	/	/
	BS-430 ¹³	S2	0.050	1.79	2	11
S3		0.560	20.0	2	30	90
S4		1.15	41.1	4	30	90
S5		1.87	66.8	8	25	100
S6		3.11	111	2.3	/	/
BS-480 ¹⁴		S2	0.050	1.79	2	14
	S3	0.560	20.0	2	40	120
	S4	1.15	41.1	4	40	120
	S5	1.87	66.8	8	30	120
	S6	3.11	111	2.3	/	/
	BS-600 ¹⁵	S2	0.050	1.79	2	11
S3		0.560	20.0	2	30	90
S4		1.15	41.1	4	30	90
S5		1.87	66.8	8	25	100
S6		3.11	111	2.3	/	/
BS-600M ¹⁶		S2	0.090	3.21	2	10
	S3	0.590	21.1	2	25	75
	S4	1.25	44.6	4	25	75
	S5	2.05	73.2	8	25	100
	S6	3.15	112	2.3	/	/

Lipids Calibrator



BS-800 ¹⁷	S2	0.050	1.79	2	10	90
	S3	0.560	20.0	2	30	90
	S4	1.15	41.1	4	30	90
	S5	1.87	66.8	8	25	100
	S6	3.11	111	2.3	/	/
BS-2000 ¹⁸	S2	0.090	3.21	2	10	90
	S3	0.590	21.1	2	25	75
	S4	1.15	41.1	4	25	75
	S5	1.87	66.8	8	25	100
	S6	3.11	111	2.3	/	/
BS-2800M ¹⁹	S2	0.090	3.21	2	10	90
	S3	0.590	21.1	2	25	75
	S4	1.25	44.6	4	25	75
	S5	2.05	73.2	8	25	100
	S6	3.15	112	2.3	/	/
Abbreviated name	ApoB	Calibration Rule		Logit-Log(5P)		
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)
		g/L	μmol/L			
BS-230 ⁵	S2	0.220	0.429	2	30	120
	S3	0.510	0.995	4	30	120
	S4	0.980	1.91	8	30	120
	S5	1.38	2.69	2	/	/
	S6	2.70	5.27	4	/	/
BS-240E ⁶	S2	0.200	0.390	2	25	100
	S3	0.450	0.878	4	25	100
	S4	0.930	1.81	8	25	100
	S5	1.41	2.75	2	/	/
	S6	2.70	5.27	4	/	/
BS-360E ¹⁰	S2	0.230	0.449	2	25	100
	S3	0.480	0.936	4	25	100
	S4	0.960	1.87	8	25	100
	S5	1.43	2.79	2	/	/
	S6	2.70	5.27	4	/	/
BS-430 ¹³	S2	0.190	0.371	2	25	100
	S3	0.460	0.897	4	25	100
	S4	0.990	1.93	8	25	100
	S5	1.37	2.67	2	/	/
	S6	2.70	5.27	4	/	/
BS-480 ¹⁴	S2	0.220	0.429	2	30	120
	S3	0.500	0.975	4	30	120
	S4	1.03	2.01	8	30	120
	S5	1.42	2.77	2	/	/
	S6	2.70	5.27	4	/	/
BS-600 ¹⁵	S2	0.210	0.410	2	25	100
	S3	0.490	0.956	4	25	100
	S4	1.00	1.95	8	25	100
	S5	1.43	2.79	2	/	/
	S6	2.70	5.27	4	/	/

BS-600M ¹⁶	S2	0.230	0.449	2	25	100
	S3	0.560	1.09	4	25	100
	S4	1.12	2.18	8	25	100
	S5	1.40	2.73	2	/	/
	S6	2.70	5.27	4	/	/
BS-800 ¹⁷	S2	0.210	0.410	2	25	100
	S3	0.470	0.917	4	25	100
	S4	0.960	1.87	8	25	100
	S5	1.40	2.73	2	/	/
	S6	2.70	5.27	4	/	/
BS-2000 ¹⁸	S2	0.230	0.449	2	25	100
	S3	0.550	1.07	4	25	100
	S4	1.08	2.11	8	25	100
	S5	1.40	2.73	2	/	/
	S6	2.72	5.30	4	/	/
BS-2800M ¹⁹	S2	0.230	0.449	2	25	100
	S3	0.550	1.07	4	25	100
	S4	1.11	2.16	8	25	100
	S5	1.42	2.77	2	/	/
	S6	2.73	5.32	4	/	/

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis
Русский	Эталонное значение	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
Português	Valor de calibração	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
Español	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
Italiano	Valore di calibrazione	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
Türkçe	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

	HDL-C	LDL-C	ApoA1	ApoB
English	HDL-Cholesterol	LDL-Cholesterol	Apolipoprotein A1	Apolipoprotein B
Русский	Холестерин ЛПВП	Холестерин ЛПНП	Аполипопротеин А1	Аполипопротеин В
Português	Colesterol HDL	Colesterol LDL	Apolipoproteína A1	Apolipoproteína B
Español	Colesterol HDL	Colesterol LDL	Apolipoproteína A1	Apolipoproteína B
Italiano	Colesterolo HDL	Colesterolo LDL	Apolipoproteina A1	Apolipoproteina B
Türkçe	HDL-Kolesterol	LDL-Kolesterol	Apolipoprotein A1	Apolipoprotein B