

Prealbumin Calibrator



Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-120:** BS-120, BS-130;

11. **BS-380:** BS-380, BS-390;

2. **BS-180:** BS-180, BS-190;

12. **BS-400:** BS-400, BS-420;

3. **BS-200:** BS-200, BS-220;

13. **BS-430:** BS-430, BS-450, BS-460;

4. **BS-200E:** BS-200E, BS-220E;

14. **BS-480:** BS-480, BS-490;

5. **BS-230:** BS-230, BS-240;

15. **BS-600:** BS-600, BS-620;

6. **BS-240E:** BS-240E, BS-240Pro;

16. **BS-600M:** BS-600M, BS-620M;

7. **BS-300:** BS-300, BS-320;

17. **BS-800:** BS-800, BS-820, BS-800M,

8. **BS-330:** BS-330, BS-350;

BS-820M, BS-1800, BS-1800plus;

9. **BS-330E: BS-330E(Serial Number starts with "XQ-"),**

BS-350E(Serial Number starts with "XS-");

18. **BS-2000:** BS-2000, BS-2200,

10. **BS-360E:** BS-360E, BS-370E, BS-350S, BS-360S,

BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

BS-2000M, BS-2200M;

19. **BS-2800M:** BS-2800M, BS-2600M.

20. **S1:0.9% NaCl, Conc. Of S1=0;**

Русский: S1: 0,9% NaCl, конц. S1=0;

Português: S1:0,9% NaCl, Conc. de S1=0;

Türkçe: S1:%0,9 NaCl, S1 Kons.=0;

Italiano: S1:0,9% NaCl, conc. di S1=0;

Español: S1:0,9% NaCl, Conc. de S1=0.

LOT: 151023005

EXP: 2024-09-30

Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol for Dilution (µL)	Diluent Vol (µL)	Sample Vol For Analysis (µL)
		mg/L	µmol/L			
BS-120 ¹	S2	59.6	1.08	33	187	7
	S3	121	2.20	45	180	10.5
	S4	224	4.08	45	180	21
	S5	406	7.39	/	/	7
	S6	996	18.1	/	/	20
BS-180 ²	S2	59.6	1.08	33	187	7
	S3	121	2.20	45	180	10.5
	S4	224	4.08	45	180	21
	S5	406	7.39	/	/	7
	S6	996	18.1	/	/	20
BS-200 ³	S2	68.0	1.24	33	187	7
	S3	123	2.24	45	180	10.5
	S4	218	3.97	45	180	21
	S5	402	7.32	/	/	7
	S6	996	18.1	/	/	20
BS-200E ⁴	S2	63.5	1.16	30	170	8.4
	S3	135	2.46	35	140	12.8
	S4	236	4.30	35	140	25.9
	S5	412	7.50	/	/	8.4
	S6	996	18.1	/	/	24
BS-300 ⁷	S2	64.0	1.16	33	187	7
	S3	125	2.28	15	195	32
	S4	232	4.22	35	230	34
	S5	412	7.50	/	/	7
	S6	996	18.1	/	/	20

Prealbumin Calibrator



Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Dilution (μL)	Diluent Vol (μL)	Sample Vol For Analysis (μL)
		mg/L	μmol/L			
BS-330 ⁸	S2	68.0	1.24	33	187	7
	S3	123	2.24	45	180	10.5
	S4	218	3.97	45	180	21
	S5	402	7.32	/	/	7
	S6	996	18.1	/	/	20
BS-330E ⁹	S2	63.5	1.16	30	170	8.4
	S3	135	2.46	35	140	12.8
	S4	236	4.30	35	140	25.9
	S5	412	7.50	/	/	8.4
	S6	996	18.1	/	/	24
BS-380 ¹¹	S2	59.6	1.08	24	136	7
	S3	121	2.20	10	130	32
	S4	224	4.08	21	138	34
	S5	406	7.39	/	/	7
	S6	996	18.1	/	/	20
BS-400 ¹²	S2	66.0	1.20	24	136	7
	S3	123	2.24	10	130	32
	S4	228	4.15	21	138	34
	S5	415	7.55	/	/	7
	S6	996	18.1	/	/	20
Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (μL)	Sample Vol for Dilution (μL)	Diluent Vol (μL)
		mg/L	μmol/L			
BS-230 ⁵	S2	59.6	1.08	5.6	27	153
	S3	121	2.20	8.4	36	144
	S4	224	4.08	16.8	36	144
	S5	406	7.39	5.6	/	/
	S6	996	18.1	16	/	/
BS-240E ⁶	S2	64.0	1.16	5.6	21	119
	S3	124	2.26	25.6	10	130
	S4	227	4.13	27.2	21	138
	S5	410	7.46	5.6	/	/
	S6	996	18.1	16	/	/
BS-360E ¹⁰	S2	63.0	1.15	7	21	119
	S3	125	2.28	32	10	130
	S4	228	4.15	34	21	138
	S5	420	7.64	7	/	/
	S6	996	18.1	20	/	/
BS-430 ¹³	S2	59.6	1.08	7	18	102
	S3	121	2.20	32	10	130
	S4	224	4.08	34	21	138
	S5	406	7.39	7	/	/
	S6	996	18.1	20	/	/
BS-480 ¹⁴	S2	66.0	1.20	7	24	136
	S3	126	2.29	32	10	130
	S4	230	4.19	34	21	138
	S5	416	7.57	7	/	/
	S6	945	17.2	20	/	/

Prealbumin Calibrator



Abbreviated name		PA		Calibration Rule		spline
Model	Level	Calibrator Value ²⁰		Sample Vol For Analysis (µL)	Sample Vol for Dilution (µL)	Diluent Vol (µL)
		mg/L	µmol/L			
BS-600 ¹⁵	S2	61.4	1.12	7	18	102
	S3	121	2.20	32	10	130
	S4	228	4.15	34	21	138
	S5	416	7.57	7	/	/
	S6	945	17.2	20	/	/
	BS-600M ¹⁶	S2	64.0	1.16	5.6	18
S3		130	2.37	8.4	25	100
S4		234	4.26	16.8	25	100
S5		400	7.28	5.6	/	/
S6		996	18.1	16	/	/
BS-800 ¹⁷ R1: R2: S =120:45:4.2		S2	59.6	1.08	4.2	18
	S3	121	2.20	19.2	10	130
	S4	224	4.08	20.4	21	138
	S5	406	7.39	4.2	/	/
	S6	996	18.1	12	/	/
	BS-800 ¹⁷ R1: R2: S =200:75:7	S2	59.6	1.08	7	18
S3		121	2.20	32	10	130
S4		224	4.08	34	21	138
S5		406	7.39	7	/	/
S6		996	18.1	20	/	/
BS-2000 ¹⁸		S2	63.3	1.15	4.2	18
	S3	123	2.24	19.2	10	130
	S4	228	4.15	20.4	21	138
	S5	407	7.41	4.2	/	/
	S6	945	17.2	12	/	/
	BS-2800M ¹⁹	S2	66.3	1.21	4.2	18
S3		130	2.37	19.2	10	130
S4		241	4.39	20.4	21	138
S5		417	7.59	4.2	/	/
S6		999	18.2	12	/	/
English		Abbreviated name	Calibration Rule		Model	Level
Русский	сокращенное наименование	Принцип калибровки		модель	Уровень	
Português	Nome abreviado	Regra de calibração		Modelo	Nível	
Español	nombre abreviado	Regla de calibración		modelo	Nivel	
Italiano	abbreviazione	Regola di calibrazione		modelli	Livello	
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı		model	Düzey	
English	Calibration Value	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis		
Русский	Эталонное значение	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа		
Português	Valor de calibração	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise		
Español	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis		
Italiano	Valore di calibrazione	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi		
Türkçe	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi		
PA						
English	Prealbumin					
Русский	преальбумина					
Português	pré-albumina					
Español	Prealbúmina					
Italiano	prealbumina					
Türkçe	Prealbümin					