

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português : Planilha de valores do Sistema de Medição BS da Mindray

Español : Hoja de valores del sistema de medición Mindray BS

Italiano : Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450,BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-300: BS-300, BS-320;

16.BS-1000M: BS-1000M, BS-1100M;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

18.BS-2800M:BS-2600M, BS-2800M;

9.BS-380: BS-380, BS-390;

19.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

10.BS-400: BS-400, BS-420;

LOT : 151023008

EXP : 2025-03-23

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey
English	Calibration Value	Expanded Uncertainty		
Русский	Эталонное значение	Расширенная неопределенность		
Português	Valor de calibração	Incerteza Expandida		
Español	Valor de calibración	Incertidumbre expandida		
Italiano	Valore di calibrazione	Incertezza estesa		
Türkçe	Kalibrasyon Değeri	Daha Uzun Süreli Belirsizlik		
English	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis	
Русский	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа	
Português	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise	
Español	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis	
Italiano	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi	
Türkçe	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi	
	PA			
English	Prealbumin			
Русский	преальбумина			
Português	pré-albumina			
Español	Prealbúmina			
Italiano	prealbumina			
Türkçe	Prealbümin			

Abbreviated name		PA		Calibration Rule			spline	
Model	Level	Calibrator Value		Expanded Uncertainty		Sample Vol For Dilution(μL)	Dilution Vol (μL)	Dilution Vol For Analysis(μL)
		mg/L	μmol/L	mg/L	μmol/L			
BS-120 ¹	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	33	187	7
	S3	/	/	/	/	45	180	10.5
	S4	/	/	/	/	45	180	21
	S5	/	/	/	/	/	/	7
	S6	/	/	/	/	/	/	20
BS-200 ²	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	33	187	7
	S3	/	/	/	/	45	180	10.5
	S4	/	/	/	/	45	180	21
	S5	/	/	/	/	/	/	7
	S6	/	/	/	/	/	/	20
BS-200E ³	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	30	170	8.4
	S3	/	/	/	/	35	140	12.8
	S4	/	/	/	/	35	140	25.9
	S5	/	/	/	/	/	/	8.4
	S6	/	/	/	/	/	/	24
BS-300 ⁶	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	33	187	7
	S3	/	/	/	/	15	195	32
	S4	/	/	/	/	35	230	34
	S5	/	/	/	/	/	/	7
	S6	/	/	/	/	/	/	20
BS-330E ⁷	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	30	170	8.4
	S3	/	/	/	/	35	140	12.8
	S4	/	/	/	/	35	140	25.9
	S5	/	/	/	/	/	/	8.4
	S6	/	/	/	/	/	/	24
BS-360E ⁸	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	21	119	7
	S3	/	/	/	/	10	130	32
	S4	/	/	/	/	21	138	34
	S5	/	/	/	/	/	/	7
	S6	/	/	/	/	/	/	20
BS-380 ⁹	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	24	136	7
	S3	/	/	/	/	10	130	32
	S4	/	/	/	/	21	138	34
	S5	/	/	/	/	/	/	7
	S6	/	/	/	/	/	/	20
BS-400 ¹⁰	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	24	136	7
	S3	/	/	/	/	10	130	32
	S4	/	/	/	/	21	138	34
	S5	/	/	/	/	/	/	7

		S6	/	/	/	/	/	20
Abbreviated name		PA		Calibration Rule			spline	
Model	Level	Calibrator Value		Expanded Uncertainty		Dilution Vol For Analysis(μL)	Sample Vol For Dilution(μL)	Dilution Vol (μL)
		mg/L	μmol/L	mg/L	μmol/L			
BS-240 ⁴	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	5.6	27	153
	S3	/	/	/	/	8.4	36	144
	S4	/	/	/	/	16.8	36	144
	S5	/	/	/	/	5.6	/	/
	S6	/	/	/	/	16	/	/
BS-240E ⁵	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	5.6	21	119
	S3	/	/	/	/	25.6	10	130
	S4	/	/	/	/	27.2	21	138
	S5	/	/	/	/	5.6	/	/
	S6	/	/	/	/	16	/	/
BS-430 ¹¹	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	7	18	102
	S3	/	/	/	/	32	10	130
	S4	/	/	/	/	34	21	138
	S5	/	/	/	/	7	/	/
	S6	/	/	/	/	20	/	/
BS-480 ¹²	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	7	24	136
	S3	/	/	/	/	32	10	130
	S4	/	/	/	/	34	21	138
	S5	/	/	/	/	7	/	/
	S6	/	/	/	/	20	/	/
BS-600 ¹³	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	7	18	102
	S3	/	/	/	/	32	10	130
	S4	/	/	/	/	34	21	138
	S5	/	/	/	/	7	/	/
	S6	/	/	/	/	20	/	/
BS-600M ¹⁴	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	5.6	18	102
	S3	/	/	/	/	8.4	25	100
	S4	/	/	/	/	16.8	25	100
	S5	/	/	/	/	5.6	/	/
	S6	/	/	/	/	16	/	/
BS-800 ¹⁵ (R1:R2 :S=200:75:7)	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	7	18	102
	S3	/	/	/	/	32	10	130
	S4	/	/	/	/	34	21	138
	S5	/	/	/	/	7	/	/
	S6	/	/	/	/	20	/	/
BS-800 ¹⁵ (R1:R2 :S=120:45:4.2)	0.9%NaCl	/	/	/	/	/	/	/
	S2	/	/	/	/	4.2	18	102
	S3	/	/	/	/	19.2	10	130
	S4	/	/	/	/	20.4	21	138

	S5	/	/	/	/	4.2	/	/
	S6	/	/	/	/	12	/	/
	0.9%NaCl	0.0	0.00	/	/	/	/	/
BS-1000M¹⁶	S2	62.3	1.13	6	0.10	4.2	18	102
	S3	126	2.29	11	0.21	19.2	10	130
	S4	236	4.30	21	0.4	20.4	21	138
	S5	399	7.26	36	0.7	4.2	/	/
	S6	977	17.8	88	1.6	12	/	/
		0.9%NaCl	/	/	/	/	/	/
BS-2000¹⁷	S2	/	/	/	/	4.2	18	102
	S3	/	/	/	/	19.2	10	130
	S4	/	/	/	/	20.4	21	138
	S5	/	/	/	/	4.2	/	/
	S6	/	/	/	/	12	/	/
		0.9%NaCl	/	/	/	/	/	/
BS-2800M¹⁸	S2	/	/	/	/	4.2	18	102
	S3	/	/	/	/	19.2	10	130
	S4	/	/	/	/	20.4	21	138
	S5	/	/	/	/	4.2	/	/
	S6	/	/	/	/	12	/	/
		0.9%NaCl	/	/	/	/	/	/

mindray

Prealbumin Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 151023008

EXP 2025-03-23

