

**Value sheet of Mindray BS Measurement System**

**Русский:** Таблица результатов для системы BS компании Mindray

**Português:** Planilha de valores do Sistema de Medição BS da Mindray

**Español:** Hoja de valores del sistema de medición Mindray BS

**Italiano:** Scheda dei valori del sistema di misurazione BS di Mindray

**Türkçe:** Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

**Русский:** Данные совпадают во всех группах.

**Português:** A dados de cada grupo é a mesma.

**Español:** la datos de cada grupo es la misma.

**Italiano:** la dati di ogni gruppo è la stessa.

**Türkçe:** her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

10.BS-400: BS-400, BS-420;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

11.BS-430: BS-430, BS-450,BS-460;

3.BS-200E: BS-200E, BS-220E,;

12.BS-480: BS-480, BS-490;

4.BS-240: BS-230, BS-240;

13.BS-600: BS-600, BS-620;

5.BS-240E: BS240E, BS240Pro;

14.BS-600M: BS-600M, BS-620M;

6.BS-300: BS-300, BS-320;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800,

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

BS-1800plus;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

17.BS-2800M:BS-2600M, BS-2800M;

9.BS-380: BS-380, BS-390;

18.Applicable models of the chemistry shall be subject to the parameter list and instructions.

**LOT** : 151023009



: 2025-05-27

English	Abbreviated name	Calibration Rule	Model	Level
<b>Русский</b>	сокращенное наименование	Принцип калибровки	модель	Уровень
<b>Português</b>	Nome abreviado	Regra de calibração	Modelo	Nível
<b>Español</b>	nombre abreviado	Regla de calibración	modelo	Nivel
<b>Italiano</b>	abbreviazione	Regola di calibrazione	modelli	Livello
<b>Türkçe</b>	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Expanded Uncertainty
<b>Русский</b>	Эталонное значение	Расширенная неопределенность
<b>Português</b>	Valor de calibração	Incerteza Expandida
<b>Español</b>	Valor de calibración	Incertidumbre expandida
<b>Italiano</b>	Valore di calibrazione	Incertezza estesa
<b>Türkçe</b>	Kalibrasyon Değeri	Daha Uzun Süreli Belirsizlik

English	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis
<b>Русский</b>	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
<b>Português</b>	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
<b>Español</b>	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
<b>Italiano</b>	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
<b>Türkçe</b>	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

PA

<b>English</b>	Prealbumin
<b>Русский</b>	преальбумина
<b>Português</b>	pré-albumina
<b>Español</b>	Prealbúmina
<b>Italiano</b>	prealbumina
<b>Türkçe</b>	Prealbümin

Abbreviated name		PA		Calibration Rule			spline	
Model	Level	Calibrator Value		Expanded Uncertainty		Sample Vol For Dilution(μL)	Dilution Vol (μL)	Dilution Vol For Analysis(μL)
		mg/L	μmol/L	mg/L	μmol/L			
BS-120 <sup>1</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	59.0	1.07	5	0.10	33	187	7
	S3	113	2.06	10	0.18	45	180	10.5
	S4	224	4.08	20	0.4	45	180	21
	S5	414	7.53	37	0.7	/	/	7
	S6	977	17.8	87	1.6	/	/	20
BS-200 <sup>2</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	62.5	1.14	6	0.10	33	187	7
	S3	117	2.13	10	0.19	45	180	10.5
	S4	215	3.91	19	0.3	45	180	21
	S5	414	7.53	37	0.7	/	/	7
	S6	956	17.4	86	1.6	/	/	20
BS-200E <sup>3</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	61.0	1.11	5	0.10	30	170	8.4
	S3	124	2.26	11	0.20	35	140	12.8
	S4	242	4.40	22	0.4	35	140	25.9
	S5	440	8.01	39	0.7	/	/	8.4
	S6	999	18.2	89	1.6	/	/	24
BS-300 <sup>6</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	62.0	1.13	6	0.10	33	187	7
	S3	118	2.15	11	0.19	15	195	32
	S4	238	4.33	21	0.4	35	230	34
	S5	425	7.74	38	0.7	/	/	7
	S6	999	18.2	89	1.6	/	/	20
BS-330E <sup>7</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	61.0	1.11	5	0.10	30	170	8.4
	S3	124	2.26	11	0.20	35	140	12.8
	S4	242	4.40	22	0.4	35	140	25.9
	S5	440	8.01	39	0.7	/	/	8.4
	S6	999	18.2	89	1.6	/	/	24
BS-360E <sup>8</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	62.0	1.13	6	0.10	21	119	7
	S3	117	2.13	10	0.19	10	130	32
	S4	230	4.19	21	0.4	21	138	34
	S5	424	7.72	38	0.7	/	/	7
	S6	999	18.2	89	1.6	/	/	20
BS-380 <sup>9</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	59.0	1.07	5	0.10	24	136	7
	S3	113	2.06	10	0.18	10	130	32
	S4	224	4.08	20	0.4	21	138	34
	S5	414	7.53	37	0.7	/	/	7
	S6	977	17.8	87	1.6	/	/	20
BS-400 <sup>10</sup>	生理盐水	0.0	0.00	/	/	/	/	/
	S2	59.0	1.07	5	0.10	24	136	7
	S3	113	2.06	10	0.18	10	130	32
	S4	224	4.08	20	0.4	21	138	34
	S5	414	7.53	37	0.7	/	/	7

BS-400 <sup>10</sup>		S6	977	17.8	87	1.6	/	/	20
Abbreviated name		PA		Calibration Rule			spline		
Model	Level	Calibrator Value		Expanded Uncertainty		Dilution Vol For Analysis(μL)	Sample Vol For Dilution(μL)	Dilution Vol (μL)	
		mg/L	μmol/L	mg/L	μmol/L				
BS-240 <sup>4</sup>	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	5.6	27	153	
	S3	113	2.06	10	0.18	8.4	36	144	
	S4	224	4.08	20	0.4	16.8	36	144	
	S5	414	7.53	37	0.7	5.6	/	/	
	S6	977	17.8	87	1.6	16	/	/	
BS-240E <sup>5</sup>	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	5.6	21	119	
	S3	113	2.06	10	0.18	25.6	10	130	
	S4	224	4.08	20	0.4	27.2	21	138	
	S5	414	7.53	37	0.7	5.6	/	/	
	S6	977	17.8	87	1.6	16	/	/	
BS-430 <sup>11</sup>	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	18	102	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-480 <sup>12</sup>	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	24	136	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-600 <sup>13</sup>	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	18	102	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-600M <sup>14</sup>	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	61.0	1.11	5	0.10	5.6	18	102	
	S3	130	2.37	12	0.21	8.4	25	100	
	S4	233	4.24	21	0.4	16.8	25	100	
	S5	414	7.53	37	0.7	5.6	/	/	
	S6	977	17.8	87	1.6	16	/	/	
BS-800 <sup>15</sup> (R1:R2 :S=200:75:7 )	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	18	102	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-800 <sup>15</sup> (R1:R2 :S=120:45:4.2)	生理盐水	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	4.2	18	102	
	S3	113	2.06	10	0.18	19.2	10	130	
	S4	224	4.08	20	0.4	20.4	21	138	

# Prealbumin Calibrator



<b>BS-800<sup>15</sup></b> (R1:R2 :S=120:45:4.	<b>S5</b>	414	7.53	37	0.7	4.2	/	/
	<b>S6</b>	977	17.8	87	1.6	12	/	/
	生理盐水	0.0	0.00	/	/	/	/	/
<b>BS-2000<sup>16</sup></b>	<b>S2</b>	61.0	1.11	5	0.10	4.2	18	102
	<b>S3</b>	113	2.06	10	0.18	19.2	10	130
	<b>S4</b>	218	3.97	20	0.4	20.4	21	138
	<b>S5</b>	411	7.48	37	0.7	4.2	/	/
	<b>S6</b>	931	16.9	83	1.5	12	/	/
		生理盐水	0.0	0.00	/	/	/	/
<b>BS-2800M<sup>17</sup></b>	<b>S2</b>	62.2	1.13	6	0.10	4.2	18	102
	<b>S3</b>	120	2.18	11	0.20	19.2	10	130
	<b>S4</b>	231	4.20	21	0.4	20.4	21	138
	<b>S5</b>	410	7.46	37	0.7	4.2	/	/
	<b>S6</b>	983	17.9	88	1.6	12	/	/
		生理盐水	0.0	0.00	/	/	/	/

# **mindray**

## **Prealbumin Calibrator**

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

**LOT** 151023009

**EXP** 2025-05-27

