

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

10.BS-400: BS-400, BS-420;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

11.BS-430: BS-430, BS-450,BS-460;

3.BS-200E: BS-200E, BS-220E,;

12.BS-480: BS-480, BS-490;

4.BS-240: BS-230, BS-240;

13.BS-600: BS-600, BS-620;

5.BS-240E: BS240E, BS240Pro;

14.BS-600M: BS-600M, BS-620M;

6.BS-300: BS-300, BS-320;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800,

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

BS-1800plus;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

17.BS-2800M:BS-2600M, BS-2800M;

9.BS-380: BS-380, BS-390;

18.Applicable models of the chemistry shall be subject to the parameter list and instructions.

LOT : 151023010



: 2025-06-12

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Expanded Uncertainty
Русский	Эталонное значение	Расширенная неопределенность
Português	Valor de calibração	Incerteza Expandida
Español	Valor de calibración	Incertidumbre expandida
Italiano	Valore di calibrazione	Incertezza estesa
Türkçe	Kalibrasyon Değeri	Daha Uzun Süreli Belirsizlik

English	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis
Русский	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
Português	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
Español	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
Italiano	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
Türkçe	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

PA

English	Prealbumin
Русский	преальбумина
Português	pré-albumina
Español	Prealbúmina
Italiano	prealbumina
Türkçe	Prealbümin

Abbreviated name		PA		Calibration Rule			spline	
Model	Level	Calibrator Value		Expanded Uncertainty		Sample Vol For Dilution(μL)	Dilution Vol (μL)	Dilution Vol For Analysis(μL)
		mg/L	μmol/L	mg/L	μmol/L			
BS-120 ¹	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	59.0	1.07	5	0.10	33	187	7
	S3	113	2.06	10	0.18	45	180	10.5
	S4	224	4.08	20	0.4	45	180	21
	S5	414	7.53	37	0.7	/	/	7
	S6	977	17.8	87	1.6	/	/	20
BS-200 ²	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	62.5	1.14	6	0.10	33	187	7
	S3	117	2.13	10	0.19	45	180	10.5
	S4	215	3.91	19	0.3	45	180	21
	S5	414	7.53	37	0.7	/	/	7
	S6	956	17.4	86	1.6	/	/	20
BS-200E ³	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	61.0	1.11	5	0.10	30	170	8.4
	S3	124	2.26	11	0.20	35	140	12.8
	S4	242	4.40	22	0.4	35	140	25.9
	S5	440	8.01	39	0.7	/	/	8.4
	S6	999	18.2	89	1.6	/	/	24
BS-300 ⁶	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	62.0	1.13	6	0.10	33	187	7
	S3	118	2.15	11	0.19	15	195	32
	S4	238	4.33	21	0.4	35	230	34
	S5	425	7.74	38	0.7	/	/	7
	S6	999	18.2	89	1.6	/	/	20
BS-330E ⁷	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	61.0	1.11	5	0.10	30	170	8.4
	S3	124	2.26	11	0.20	35	140	12.8
	S4	242	4.40	22	0.4	35	140	25.9
	S5	440	8.01	39	0.7	/	/	8.4
	S6	999	18.2	89	1.6	/	/	24
BS-360E ⁸	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	62.0	1.13	6	0.10	21	119	7
	S3	117	2.13	10	0.19	10	130	32
	S4	230	4.19	21	0.4	21	138	34
	S5	424	7.72	38	0.7	/	/	7
	S6	999	18.2	89	1.6	/	/	20
BS-380 ⁹	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	59.0	1.07	5	0.10	24	136	7
	S3	113	2.06	10	0.18	10	130	32
	S4	224	4.08	20	0.4	21	138	34
	S5	414	7.53	37	0.7	/	/	7
	S6	977	17.8	87	1.6	/	/	20
BS-400 ¹⁰	0.9%NaCl	0.0	0.00	/	/	/	/	/
	S2	59.0	1.07	5	0.10	24	136	7
	S3	113	2.06	10	0.18	10	130	32
	S4	224	4.08	20	0.4	21	138	34
	S5	414	7.53	37	0.7	/	/	7

		S6	977	17.8	87	1.6	/	/	20
Abbreviated name		PA		Calibration Rule			spline		
Model	Level	Calibrator Value		Expanded Uncertainty		Dilution Vol For Analysis(μL)	Sample Vol For Dilution(μL)	Dilution Vol (μL)	
		mg/L	μmol/L	mg/L	μmol/L				
BS-240 ⁴	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	5.6	27	153	
	S3	113	2.06	10	0.18	8.4	36	144	
	S4	224	4.08	20	0.4	16.8	36	144	
	S5	414	7.53	37	0.7	5.6	/	/	
	S6	977	17.8	87	1.6	16	/	/	
BS-240E ⁵	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	5.6	21	119	
	S3	113	2.06	10	0.18	25.6	10	130	
	S4	224	4.08	20	0.4	27.2	21	138	
	S5	414	7.53	37	0.7	5.6	/	/	
	S6	977	17.8	87	1.6	16	/	/	
BS-430 ¹¹	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	18	102	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-480 ¹²	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	24	136	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-600 ¹³	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	18	102	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-600M ¹⁴	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	61.0	1.11	5	0.10	5.6	18	102	
	S3	130	2.37	12	0.21	8.4	25	100	
	S4	233	4.24	21	0.4	16.8	25	100	
	S5	414	7.53	37	0.7	5.6	/	/	
	S6	977	17.8	87	1.6	16	/	/	
BS-800 ¹⁵ (R1:R2 :S=200:75:7)	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	7	18	102	
	S3	113	2.06	10	0.18	32	10	130	
	S4	224	4.08	20	0.4	34	21	138	
	S5	414	7.53	37	0.7	7	/	/	
	S6	977	17.8	87	1.6	20	/	/	
BS-800 ¹⁵ (R1:R2 :S=120:45:4.2)	0.9%NaCl	0.0	0.00	/	/	/	/	/	
	S2	59.0	1.07	5	0.10	4.2	18	102	
	S3	113	2.06	10	0.18	19.2	10	130	
	S4	224	4.08	20	0.4	20.4	21	138	

Prealbumin Calibrator



	S5	414	7.53	37	0.7	4.2	/	/
	S6	977	17.8	87	1.6	12	/	/
	0.9%NaCl	0.0	0.00	/	/	/	/	/
BS-2000¹⁶	S2	61.0	1.11	5	0.10	4.2	18	102
	S3	113	2.06	10	0.18	19.2	10	130
	S4	218	3.97	20	0.4	20.4	21	138
	S5	411	7.48	37	0.7	4.2	/	/
	S6	931	16.9	83	1.5	12	/	/
		0.9%NaCl	0.0	0.00	/	/	/	/
BS-2800M¹⁷	S2	62.2	1.13	6	0.10	4.2	18	102
	S3	120	2.18	11	0.20	19.2	10	130
	S4	231	4.20	21	0.4	20.4	21	138
	S5	410	7.46	37	0.7	4.2	/	/
	S6	983	17.9	88	1.6	12	/	/
		0.9%NaCl	0.0	0.00	/	/	/	/

mindray

Prealbumin Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 151023010

EXP 2025-06-12

