

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450, BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-

6.BS-300: BS-300, BS-320;

1800plus;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"),

16.BS-1000M: BS-1000M, BS-1100M;

BS-350E(Serial Number starts with "XS-")

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

17.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

330E(V35.00)(Serial Number starts with "W8-" and

software version starts with "35.00"), BS-350E(V35.00)

18.BS-2800M:BS-2600M, BS-2800M.

(Serial Number starts with "W9-" and software version

19.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

starts with "35.00");

20.For applicable models of the analyte, please refer to the reagent

9.BS-380: BS-380, BS-390;

parameter sheet and instruction.

10.BS-400: BS-400, BS-420;



LOT : 151424004



: 2025-07-31

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modelli
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model
English	Expanded Uncertainty			
Русский	Расширенная неопределенность			
Português	Incerteza Expandida			
Español	Incertidumbre expandida			
Italiano	Incertezza estesa			
Türkçe	Daha Uzun Süreli Belirsizlik			
CK-MB				
English	Creatine Kinase-MB			
Русский	MB фракцию креатинкиназы			
Português	creatina quinase-MB			
Español	creatina quinasa-MB			
Italiano	creatina chinasi-MB			
Türkçe	Kreatin Kinaz-MB			

Abbreviated name	Unit	Calibration Rule				Two-point linear	
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
CK-MB	U/L	BS-120¹	119	5	BS-400¹⁰	123	5
		BS-200²	119	5	BS-430¹¹	123	5
		BS-200E³	123	5	BS-480¹²	123	5
		BS-240⁴	119	5	BS-600¹³	123	5
		BS-240E⁵	123	5	BS-600M¹⁴	125	5
		BS-300⁶	123	5	BS-800¹⁵	123	5
		BS-330E⁷	123	5	BS-1000M¹⁶	125	5
		BS-360E⁸	123	5	BS-2000¹⁷	123	5
		BS-380⁹	123	5	BS-2800M¹⁸	125	5
	μkat/L	BS-120¹	1.99	0.08	BS-400¹⁰	2.05	0.08
		BS-200²	1.99	0.08	BS-430¹¹	2.05	0.08
		BS-200E³	2.05	0.08	BS-480¹²	2.05	0.08
		BS-240⁴	1.99	0.08	BS-600¹³	2.05	0.08
		BS-240E⁵	2.05	0.08	BS-600M¹⁴	2.09	0.08
		BS-300⁶	2.05	0.08	BS-800¹⁵	2.05	0.08
		BS-330E⁷	2.05	0.08	BS-1000M¹⁶	2.09	0.08
		BS-360E⁸	2.05	0.08	BS-2000¹⁷	2.05	0.08
		BS-380⁹	2.05	0.08	BS-2800M¹⁸	2.09	0.08

mindray

CK-MB Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 151424004

 2025-07-31

