

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

10.BS-480: BS-480, BS-490;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

11.BS-600: BS-600, BS-620;

3.BS-200E: BS-200E, BS-220E,;

12.BS-600M: BS-600M, BS-620M;

4.BS-240: BS-230, BS-240;

13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

5.BS-240E: BS240E, BS240Pro;

6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

14.BS-1000M: BS-1000M, BS-1100M;

7.BS-380: BS-380, BS-390;

15.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-400: BS-400, BS-420;

16.BS-2800M:BS-2600M, BS-2800M;

9.BS-430: BS-430, BS-450,BS-460;

17.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.

LOT : 152624001

EXP : 2025-07-02

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

MALB	
English	Microalbumin
Русский	микроальбумина
Português	Microalbuminúria
Español	Microalbúmina
Italiano	Microalbumina
Türkçe	Mikroalbumin

Abbreviated name		MALB		Calibration Rule		spline	
Model	Unit	BS-120 ¹	BS-200 ²	BS-200E ³	BS-240 ⁴	BS-200E ³	BS-240 ⁴
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/
a	mg/L	/	/	/	/	/	/
b		/	/	/	/	/	/

c	/	/	/	/	/	/	/	/	/
d	/	/	/	/	/	/	/	/	/
e	/	/	/	/	/	/	/	/	/
0.9%NaCl	/	/	/	/	/	/	/	/	/
a	/	/	/	/	/	/	/	/	/
b	μmol/L	/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
e		/	/	/	/	/	/	/	/
		/	/	/	/	/	/	/	/
Model	BS-240E⁵		BS-360E⁶		BS-380⁷		BS-400⁸		
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a		/	/	/	/	/	/	/	/
b	mg/L	/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
e		/	/	/	/	/	/	/	/
		/	/	/	/	/	/	/	/
0.9%NaCl		/	/	/	/	/	/	/	/
a		/	/	/	/	/	/	/	/
b	μmol/L	/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
e		/	/	/	/	/	/	/	/
		/	/	/	/	/	/	/	/
Model	BS-430⁹		BS-480¹⁰		BS-600¹¹		BS-600M¹²		
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	/	/	/	/	/	/
a		/	/	/	/	/	/	/	/
b	mg/L	/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
e		/	/	/	/	/	/	/	/
		/	/	/	/	/	/	/	/
0.9%NaCl		/	/	/	/	/	/	/	/
a		/	/	/	/	/	/	/	/
b	μmol/L	/	/	/	/	/	/	/	/
c		/	/	/	/	/	/	/	/
d		/	/	/	/	/	/	/	/
e		/	/	/	/	/	/	/	/
		/	/	/	/	/	/	/	/
Model	BS-800¹³		BS-1000M¹⁴		BS-2000¹⁵		BS-2800M¹⁶		
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		/	/	0.0	/	/	/	/	/
a		/	/	21.0	1.8	/	/	/	/
b	mg/L	/	/	55.9	2.6	/	/	/	/
c		/	/	103	5	/	/	/	/
d		/	/	204	13	/	/	/	/
e		/	/	381	17	/	/	/	/
		/	/	/	/	/	/	/	/
0.9%NaCl		/	/	0.000	/	/	/	/	/
a		/	/	0.319	0.028	/	/	/	/

MALB Calibrator

mindray

b	/	/	0.850	0.04	/	/	/	/
c	/	/	1.57	0.08	/	/	/	/
d	$\mu\text{mol/L}$	/	3.10	0.19	/	/	/	/
e	/	/	5.79	0.25	/	/	/	/

mindray

MALB Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 152624001

 2025-07-02

