

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-240**: BS240, BS-280;

5. **BS-400**: BS-400, BS-420;

2. **BS-240Pro**: BS-240Pro, BS-240E;

6. **BS-480**: BS-480, BS-490;

3. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)** (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

7. **BS-600**: BS-600, BS-620;

8. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

9. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

10. **BS-2800M**: BS-2600M, BS-2800M.

4. **BS-380**: BS-380, BS-390;

11. a: 0.9% NaCl is not needed, Conc. Of a=0;

Русский: a: 0,9% NaCl не необходим, конц. a=0;

Português: a: 0,9% O NaCl não é needed, Conc. de a=0;

Español: a: 0,9% El NaCl no es necesario, Conc. de a=0;

Italiano: a: 0.9% Il NaCl non è necessario, conc. di a=0;

Türkçe: a: %0,9 NaCl gerekli değildir, a Kons.=0.

LOT : 153822002

⌚ : 2023-04-25

Abbreviated name			IgE		Calibration Rule		spline	
Model	Level	Calibration Value ¹¹ (IU/mL)	Model	Level	Calibration Value ¹¹ (IU/mL)	Model	Level	Calibration Value ¹¹ (IU/mL)
BS-240 ¹	a	0	BS-480 ⁶	a	0	BS-480 ⁶	a	0
	b	83.3		b	83.3		b	83.3
	c	173		c	173		c	173
	d	510		d	510		d	510
	e	900		e	900		e	900
	f	1406		f	1406		f	1406
BS-240Pro ²	a	0	BS-600 ⁷	a	0	BS-600 ⁷	a	0
	b	83.3		b	83.3		b	83.3
	c	173		c	173		c	173
	d	510		d	510		d	510
	e	900		e	900		e	900
	f	1406		f	1406		f	1406
BS-360E ³	a	0	BS-800 ⁸	a	0	BS-800 ⁸	a	0
	b	83.3		b	83.3		b	83.3
	c	173		c	173		c	173
	d	510		d	510		d	510
	e	900		e	900		e	900
	f	1406		f	1406		f	1406
BS-380 ⁴	a	0	BS-2000 ⁹	a	0	BS-2000 ⁹	a	0
	b	83.3		b	83.4		b	83.4
	c	173		c	173		c	173
	d	510		d	510		d	510
	e	900		e	908		e	908
	f	1406		f	1411		f	1411
BS-400 ⁵	a	0	BS-2800M ¹⁰	a	0	BS-2800M ¹⁰	a	0
	b	83.3		b	85.8		b	85.8
	c	173		c	172		c	172
	d	510		d	515		d	515
	e	900		e	890		e	890
	f	1406		f	1432		f	1432

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

IgE	
English	Immunoglobulin E
Русский	Иммуноглобулин E
Português	Imunoglobulina E
Español	Inmunoglobulina E
Italiano	Immunoglobulina E
Türkçe	İmmünglobulin E