

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



There are target value changes. Please use this lot of calibrators to recalibrate, the updated control value sheet is inside box, select the corresponding value and update.

The data of each group is same.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

1. **BS-240**: BS240, BS-280;

2. **BS-240Pro**: BS-240Pro, BS-240E;

3. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-**

**330E(V35.00)** (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

4. **BS-380**: BS-380, BS-390;

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

5. **BS-400**: BS-400, BS-420;

6. **BS-480**: BS-480, BS-490;

7. **BS-600**: BS-600, BS-620;

8. **BS-600M**: BS-600M, BS-620M;

9. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

10. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

11. **BS-2800M**: BS-2600M, BS-2800M.

12. a: 0.9% NaCl is not needed, Conc. Of a=0;

Português: a: 0,9% O NaCl não é needed, Conc. de a=0;

Italiano: a: 0.9% Il NaCl non è necessario, conc. di a=0;

Русский: a: 0,9% NaCl не необходим, конц. a=0;

Español: a: 0,9% El NaCl no es necesario, Conc. de a=0;

Türkçe: a: %0,9 NaCl gerekli değildir, a Kons.=0.

**LOT** : 153822006

**EXP** : 2024-01-21

Abbreviated name			IgE		Calibration Rule		spline	
Model	Level	Calibration Value <sup>12</sup> (IU/mL)	Model	Level	Calibration Value <sup>12</sup> (IU/mL)	Model	Level	Calibration Value <sup>12</sup> (IU/mL)
BS-240 <sup>1</sup>	a	0	BS-600 <sup>7</sup>	a	0	BS-600 <sup>7</sup>	a	0
	b	94.3		b	94.3		b	94.3
	c	221		c	221		c	221
	d	661		d	661		d	661
	e	1287		e	1287		e	1287
	f	1990		f	1990		f	1990
BS-240Pro <sup>2</sup>	a	0	BS-600M <sup>8</sup>	a	0	BS-600M <sup>8</sup>	a	0
	b	94.3		b	94.8		b	94.8
	c	221		c	219		c	219
	d	661		d	663		d	663
	e	1287		e	1284		e	1284
	f	1990		f	2028		f	2028
BS-360E <sup>3</sup>	a	0	BS-800 <sup>9</sup>	a	0	BS-800 <sup>9</sup>	a	0
	b	93.5		b	94.3		b	94.3
	c	216		c	221		c	221
	d	660		d	661		d	661
	e	1287		e	1287		e	1287
	f	1980		f	1990		f	1990
BS-380 <sup>4</sup>	a	0	BS-2000 <sup>10</sup>	a	0	BS-2000 <sup>10</sup>	a	0
	b	94.3		b	92.8		b	92.8
	c	225		c	221		c	221
	d	665		d	657		d	657
	e	1287		e	1282		e	1282
	f	2000		f	2034		f	2034
BS-400 <sup>5</sup>	a	0	BS-2800M <sup>11</sup>	a	0	BS-2800M <sup>11</sup>	a	0
	b	94.3		b	94.8		b	94.8
	c	225		c	219		c	219
	d	665		d	663		d	663
	e	1287		e	1284		e	1284
	f	2000		f	2028		f	2028
BS-480 <sup>6</sup>	a	0		a	0		a	0
	b	94.3		b	94.8		b	94.8
	c	221		c	219		c	219
	d	661		d	663		d	663
	e	1287		e	1284		e	1284
	f	1990		f	2028		f	2028

# IgE Calibrator

\*Please note the target value change

**mindray**

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

## IgE

English	Immunoglobulin E
Русский	Иммуноглобулин E
Português	Imunoglobulina E
Español	Inmunoglobulina E
Italiano	Immunoglobulina E
Türkçe	İmmünglobulin E

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası

The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-230**: BS-230, BS-240, BS-280;

3. **BS-240E**: BS-240E, BS-240Pro;

4. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**

(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

5. **BS-380**: BS-380, BS-390;

6. **BS-400**: BS-400, BS-420;

7. **BS-430**: BS-430, BS-450, BS-460;

8. **BS-480**: BS-480, BS-490;

9. **BS-600**: BS-600, BS-620;

10. **BS-600M**: BS-600M, BS-620M;

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

There are target value changes. Please use this lot of calibrators to recalibrate, the updated control value sheet is inside box, select the corresponding value and update.



The following reference values are applicable to:

★ Calibrator lot No.: 153822003 or subsequent lot + Multimmun Control lot No.: 058021002, 058021003, 058021004, 058021005, 058021006, 058021007, 058021008, 058022001, 058022002; ★

Abbreviated name	Multimmun Control (L)					Multimmun Control (H)				
	Model	Unit	Assay Value	Range (AssayValue±)	1 SD	Model	Unit	Assay Value	Range (AssayValue±3SD)	1 SD
IgE  (Remark 2: The target value of IgE in BS-240 is only applicable to BS-240 and BS-280)	<b>BS-240</b> <sup>2</sup>	IU/mL	80.7	56.5 — 104.9	8.1	<b>BS-240</b> <sup>2</sup>	IU/mL	494	346 — 642	49
	<b>BS-240E</b> <sup>3</sup>	IU/mL	80.9	56.6 — 105.2	8.1	<b>BS-240E</b> <sup>3</sup>	IU/mL	479	335 — 623	48
	<b>BS-360E</b> <sup>4</sup>	IU/mL	79.3	55.5 — 103.1	7.9	<b>BS-360E</b> <sup>4</sup>	IU/mL	468	328 — 608	47
	<b>BS-380</b> <sup>5</sup>	IU/mL	78.6	55.0 — 102.2	7.9	<b>BS-380</b> <sup>5</sup>	IU/mL	490	343 — 637	49
	<b>BS-400</b> <sup>6</sup>	IU/mL	78.1	54.7 — 101.5	7.8	<b>BS-400</b> <sup>6</sup>	IU/mL	488	342 — 634	49
	<b>BS-480</b> <sup>8</sup>	IU/mL	83.2	58.2 — 108.2	8.3	<b>BS-480</b> <sup>8</sup>	IU/mL	496	347 — 645	50
	<b>BS-600</b> <sup>9</sup>	IU/mL	83.2	58.2 — 108.2	8.3	<b>BS-600</b> <sup>9</sup>	IU/mL	496	347 — 645	50
	<b>BS-600M</b> <sup>10</sup>	IU/mL	83.7	58.6 — 108.8	8.4	<b>BS-600M</b> <sup>10</sup>	IU/mL	496	347 — 645	50
	<b>BS-800</b> <sup>11</sup>	IU/mL	83.2	58.2 — 108.2	8.3	<b>BS-800</b> <sup>11</sup>	IU/mL	496	347 — 645	50
	<b>BS-2000</b> <sup>12</sup>	IU/mL	81.5	57.1 — 106.0	8.2	<b>BS-2000</b> <sup>12</sup>	IU/mL	493	345 — 641	49
<b>BS-2800M</b> <sup>13</sup>	IU/mL	83.7	58.6 — 108.8	8.4	<b>BS-2800M</b> <sup>13</sup>	IU/mL	496	347 — 645	50	