

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;

8.BS-480: BS-480, BS-490;

2.BS-240: BS-230, BS-240;

9.BS-600: BS-600, BS-620;

3.BS-240E: BS240E, BS240Pro;

10.BS-600M: BS-600M, BS-620M;

4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

5.BS-380: BS-380, BS-390;

12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

6.BS-400: BS-400, BS-420;

13.BS-2800M:BS-2600M, BS-2800M;

7.BS-430: BS-430, BS-450,BS-460;

14.Applicable models of the chemistry shall be subject to the parameter list and instructions.

LOT : 154024001



: 2025-01-18

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

English	Calibration Value	Expanded Uncertainty
Русский	Эталонное значение	Расширенная неопределенность
Português	Valor de calibração	Incerteza Expandida
Español	Valor de calibración	Incertidumbre expandida
Italiano	Valore di calibrazione	Incertezza estesa
Türkçe	Kalibrasyon Değeri	Daha Uzun Süreli Belirsizlik

English	Sample Vol for Dilution	Diluent Vol	Sample Vol For Analysis
Русский	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
Português	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
Español	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
Italiano	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
Türkçe	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

RBP

English	Retinol Binding Protein
Русский	Ретинол-связывающий белок
Português	Proteína transportadora do retinol
Español	Proteína fijadora de retinol
Italiano	Proteina legante il retinolo
Türkçe	Retinol Bağlayıcı Protein

Abbreviated name	RBP	Calibration Rule	spline
Model	Level	Calibrator Value	Expanded Uncertainty
		Sample Vol For Dilution(μL)	Dilution Vol (μL)
			Dilution Vol For Analysis(μL)

Model	Level	mg/dL	mg/dL	Sample Vol For Dilution(μL)	Dilution Vol (μL)	Dilution Vol For Analysis(μL)
BS-200E <sup>1</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.62	0.10	45	135	4
	S3	5.25	0.20	45	135	8
	S4	12.3	0.5	/	/	4
	S5	17.7	0.7	/	/	6
BS-360E <sup>4</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.56	0.10	34	102	4
	S3	5.18	0.20	34	102	8
	S4	12.3	0.5	/	/	4
	S5	17.7	0.7	/	/	6
BS-380 <sup>5</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.56	0.10	40	120	4
	S3	5.10	0.19	40	120	8
	S4	12.3	0.5	/	/	4
	S5	17.7	0.7	/	/	6
BS-400 <sup>6</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.56	0.10	40	120	4
	S3	5.20	0.20	40	120	8
	S4	12.3	0.5	/	/	4
	S5	17.7	0.7	/	/	6
Abbreviated name		RBP		Calibration Rule		spline
Model	Level	Calibrator Value	Expanded Uncertainty	Dilution Vol For Analysis(μL)	Sample Vol For Dilution(μL)	Dilution Vol (μL)
		mg/dL	mg/dL			
BS-240 <sup>2</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.62	0.10	4	35	105
	S3	5.25	0.20	8	35	105
	S4	12.3	0.5	4	/	/
	S5	17.7	0.7	6	/	/
BS-240E <sup>3</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.61	0.10	4	34	102
	S3	5.20	0.20	8	34	102
	S4	12.3	0.5	4	/	/
	S5	17.7	0.7	6	/	/
BS-430 <sup>7</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.56	0.10	4	30	90
	S3	5.10	0.19	8	30	90
	S4	12.3	0.5	4	/	/
	S5	17.7	0.7	6	/	/
BS-480 <sup>8</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.62	0.10	4	40	120
	S3	5.25	0.20	8	40	120
	S4	12.3	0.5	4	/	/
	S5	17.7	0.7	6	/	/
BS-600 <sup>9</sup>	0.9%NaCl	0.00	/	/	/	/
	S2	2.64	0.10	4	30	90
	S3	5.29	0.20	8	30	90
	S4	12.6	0.5	4	/	/
	S5	17.7	0.7	6	/	/

<b>BS-600M<sup>10</sup></b>	<b>0.9%NaCl</b>	0.00	/	/	/	/
	<b>S2</b>	2.77	0.11	4	25	75
	<b>S3</b>	5.73	0.22	8	25	75
	<b>S4</b>	12.4	0.5	4	/	/
	<b>S5</b>	17.9	0.7	6	/	/
<b>BS-800<sup>11</sup></b>	<b>0.9%NaCl</b>	0.00	/	/	/	/
	<b>S2</b>	2.56	0.10	4	30	90
	<b>S3</b>	5.10	0.19	8	30	90
	<b>S4</b>	12.3	0.5	4	/	/
	<b>S5</b>	17.7	0.7	6	/	/
<b>BS-2000<sup>12</sup></b>	<b>0.9%NaCl</b>	0.00	/	/	/	/
	<b>S2</b>	2.74	0.15	4	25	75
	<b>S3</b>	5.49	0.3	8	25	75
	<b>S4</b>	12.3	0.7	4	/	/
	<b>S5</b>	17.9	1.0	6	/	/
<b>BS-2800M<sup>13</sup></b>	<b>0.9%NaCl</b>	0.00	/	/	/	/
	<b>S2</b>	2.77	0.12	4	25	75
	<b>S3</b>	5.73	0.26	8	25	75
	<b>S4</b>	12.4	0.6	4	/	/
	<b>S5</b>	17.9	0.8	6	/	/

# mindray

## RBP Calibrator

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

**LOT** 154024001

**2025-01-18**

