

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- |   |  |
|---|--|
| 1. <b>BS-200</b> : BS-200, BS-220;  | 8. <b>BS-380</b> : BS-380, BS-390;   |
| 2. <b>BS-200E</b> : BS-200E, BS-220E;   | 9. <b>BS-400</b> : BS-400, BS-420;   |
| 3. <b>BS-230</b> : BS-230, BS-240;  | 10. <b>BS-430</b> : BS-430, BS-450, BS-460;                                    |
| 4. <b>BS-240E</b> : BS-240E, BS-240Pro;   | 11. <b>BS-480</b> : BS-480, BS-490;  |
| 5. <b>BS-300</b> : BS-300, BS-320;  | 12. <b>BS-600</b> : BS-600, BS-620;  |
| 6. <b>BS-330E</b> : <b>BS-330E</b> (Serial Number starts with "XQ-"),<br><b>BS-350E</b> (Serial Number starts with "XS-");  | 13. <b>BS-600M</b> : BS-600M, BS-620M;   |
| 7. <b>BS-360E</b> : BS-360E, BS-370E, BS-350S, BS-360S,<br><b>BS-330E(V35.00)</b> (Serial Number starts with "W8-" and<br>software version starts with "35.00"), <b>BS-350E(V35.00)</b><br>(Serial Number starts with "W9-" and software version<br>starts with "35.00"); | 14. <b>BS-800</b> : BS-800, BS-820, BS-800M, BS-820M, BS-1800,<br>BS-1800plus; |
|   | 15. <b>BS-2000</b> : BS-2000, BS-2200, BS-2000M, BS-2200M;                     |
|   | 16. <b>BS-2800M</b> : BS-2600M, BS-2800M.                                      |

17. IFCC:  $HbA1c(\%) = HbA1c(\mu mol/L) / Hb(\mu mol/L) \times 100$

NGSP/DCCT:  $HbA1c(\%) = 91.5 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 2.15$

JCCLS:  $HbA1c(\%) = 96.3 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 1.62$

**LOT**: 044622006

**EXP**: 2024-01-07

Abbreviated name	Model	Unit	Assay Value	Range(Assay Value±3SD)			1 SD	Model	Unit	Assay Value	Range(Assay Value±3SD)			1 SD
HbA1c	<b>BS-200</b> <sup>1</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-400</b> <sup>9</sup>	μmol/L	10.6	9.6	—	11.6	0.3
	<b>BS-200E</b> <sup>2</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-430</b> <sup>10</sup>	μmol/L	10.7	9.6	—	11.8	0.4
	<b>BS-230</b> <sup>3</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-480</b> <sup>11</sup>	μmol/L	10.7	9.6	—	11.8	0.4
	<b>BS-240E</b> <sup>4</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-600</b> <sup>12</sup>	μmol/L	10.7	9.6	—	11.8	0.4
	<b>BS-300</b> <sup>5</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-600M</b> <sup>13</sup>	μmol/L	10.8	9.7	—	11.9	0.4
	<b>BS-330E</b> <sup>6</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-800</b> <sup>14</sup>	μmol/L	10.7	9.6	—	11.8	0.4
	<b>BS-360E</b> <sup>7</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-2000</b> <sup>15</sup>	μmol/L	10.7	9.6	—	11.8	0.4
	<b>BS-380</b> <sup>8</sup>	μmol/L	10.7	9.6	—	11.8	0.4	<b>BS-2800M</b> <sup>16</sup>	μmol/L	10.7	9.6	—	11.8	0.4
Hb	<b>BS-200</b> <sup>1</sup>	μmol/L	114	103	—	125	4	<b>BS-400</b> <sup>9</sup>	μmol/L	116	105	—	127	4
	<b>BS-200E</b> <sup>2</sup>	μmol/L	116	105	—	127	4	<b>BS-430</b> <sup>10</sup>	μmol/L	116	105	—	127	4
	<b>BS-230</b> <sup>3</sup>	μmol/L	116	105	—	127	4	<b>BS-480</b> <sup>11</sup>	μmol/L	116	105	—	127	4
	<b>BS-240E</b> <sup>4</sup>	μmol/L	116	105	—	127	4	<b>BS-600</b> <sup>12</sup>	μmol/L	116	105	—	127	4
	<b>BS-300</b> <sup>5</sup>	μmol/L	117	105	—	129	4	<b>BS-600M</b> <sup>13</sup>	μmol/L	121	109	—	133	4
	<b>BS-330E</b> <sup>6</sup>	μmol/L	116	105	—	127	4	<b>BS-800</b> <sup>14</sup>	μmol/L	116	105	—	127	4
	<b>BS-360E</b> <sup>7</sup>	μmol/L	116	105	—	127	4	<b>BS-2000</b> <sup>15</sup>	μmol/L	116	105	—	127	4
	<b>BS-380</b> <sup>8</sup>	μmol/L	116	105	—	127	4	<b>BS-2800M</b> <sup>16</sup>	μmol/L	116	105	—	127	4

# HbA1c Control P



Abbreviated name	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	
HbA1c <sub>17</sub> (%)	IFCC	BS-200 <sup>1</sup>	%	9.39	8.46 — 10.32	0.31	BS-400 <sup>9</sup>	%	9.17	8.26 — 10.08	0.30
		BS-200E <sup>2</sup>	%	9.18	8.27 — 10.09	0.30	BS-430 <sup>10</sup>	%	9.20	8.29 — 10.11	0.30
		BS-230 <sup>3</sup>	%	9.20	8.29 — 10.11	0.30	BS-480 <sup>11</sup>	%	9.20	8.29 — 10.11	0.30
		BS-240E <sup>4</sup>	%	9.20	8.29 — 10.11	0.30	BS-600 <sup>12</sup>	%	9.20	8.29 — 10.11	0.30
		BS-300 <sup>5</sup>	%	9.13	8.23 — 10.03	0.30	BS-600M <sup>13</sup>	%	8.94	8.05 — 9.83	0.30
		BS-330E <sup>6</sup>	%	9.18	8.27 — 10.09	0.30	BS-800 <sup>14</sup>	%	9.20	8.29 — 10.11	0.30
		BS-360E <sup>7</sup>	%	9.20	8.29 — 10.11	0.30	BS-2000 <sup>15</sup>	%	9.26	8.34 — 10.18	0.31
		BS-380 <sup>8</sup>	%	9.18	8.27 — 10.09	0.30	BS-2800M <sup>16</sup>	%	9.23	8.32 — 10.14	0.30
HbA1c <sub>17</sub> (%)	NGSP / DCCT	BS-200 <sup>1</sup>	%	10.7	9.6 — 11.8	0.4	BS-400 <sup>9</sup>	%	10.5	9.5 — 11.5	0.3
		BS-200E <sup>2</sup>	%	10.6	9.6 — 11.6	0.3	BS-430 <sup>10</sup>	%	10.6	9.6 — 11.6	0.3
		BS-230 <sup>3</sup>	%	10.6	9.6 — 11.6	0.3	BS-480 <sup>11</sup>	%	10.6	9.6 — 11.6	0.3
		BS-240E <sup>4</sup>	%	10.6	9.6 — 11.6	0.3	BS-600 <sup>12</sup>	%	10.6	9.6 — 11.6	0.3
		BS-300 <sup>5</sup>	%	10.5	9.5 — 11.5	0.3	BS-600M <sup>13</sup>	%	10.3	9.3 — 11.3	0.3
		BS-330E <sup>6</sup>	%	10.6	9.6 — 11.6	0.3	BS-800 <sup>14</sup>	%	10.6	9.6 — 11.6	0.3
		BS-360E <sup>7</sup>	%	10.6	9.6 — 11.6	0.3	BS-2000 <sup>15</sup>	%	10.6	9.6 — 11.6	0.3
		BS-380 <sup>8</sup>	%	10.6	9.6 — 11.6	0.3	BS-2800M <sup>16</sup>	%	10.6	9.6 — 11.6	0.3
	JCCLS	BS-200 <sup>1</sup>	%	10.7	9.6 — 11.8	0.4	BS-400 <sup>9</sup>	%	10.4	9.4 — 11.4	0.3
		BS-200E <sup>2</sup>	%	10.5	9.5 — 11.5	0.3	BS-430 <sup>10</sup>	%	10.5	9.5 — 11.5	0.3
		BS-230 <sup>3</sup>	%	10.5	9.5 — 11.5	0.3	BS-480 <sup>11</sup>	%	10.5	9.5 — 11.5	0.3
		BS-240E <sup>4</sup>	%	10.5	9.5 — 11.5	0.3	BS-600 <sup>12</sup>	%	10.5	9.5 — 11.5	0.3
		BS-300 <sup>5</sup>	%	10.4	9.4 — 11.4	0.3	BS-600M <sup>13</sup>	%	10.2	9.2 — 11.2	0.3
		BS-330E <sup>6</sup>	%	10.5	9.5 — 11.5	0.3	BS-800 <sup>14</sup>	%	10.5	9.5 — 11.5	0.3
		BS-360E <sup>7</sup>	%	10.5	9.5 — 11.5	0.3	BS-2000 <sup>15</sup>	%	10.5	9.5 — 11.5	0.3
		BS-380 <sup>8</sup>	%	10.5	9.5 — 11.5	0.3	BS-2800M <sup>16</sup>	%	10.5	9.5 — 11.5	0.3

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione ±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

	HbA1c	Hb
English	Hemoglobin A1c	Hemoglobin
Русский	гемоглобин A1c	гемоглобин
Português	Hemoglobina A1c	Hemoglobina
Español	hemoglobina A1c	hemoglobina
Italiano	Emoglobina A1c	Emoglobina
Türkçe	Hemoglobin A1c	Hemoglobin