

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200**: BS-200, BS-220;

2. **BS-200E**: BS-200E, BS-220E;

3. **BS-230**: BS-230, BS-240;

4. **BS-240E**: BS-240E, BS-240Pro;

5. **BS-300**: BS-300, BS-320;

6. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-");

7. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00)(Serial Number starts with "W8-" and
software version starts with "35.00"), **BS-350E(V35.00)**
(Serial Number starts with "W9-" and software version
starts with "35.00");

8. **BS-380**: BS-380, BS-390;

9. **BS-400**: BS-400, BS-420;

10. **BS-430**: BS-430, BS-450, BS-460;

11. **BS-480**: BS-480, BS-490;

12. **BS-600**: BS-600, BS-620;

13. **BS-600M**: BS-600M, BS-620M;

14. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800,
BS-1800plus;

15. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

16. **BS-2800M**: BS-2600M, BS-2800M.

17. IFCC: $HbA1c(\%) = HbA1c(\mu mol/L) / Hb(\mu mol/L) \times 100$

NGSP/DCCT: $HbA1c(\%) = 91.5 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 2.15$

JCCLS: $HbA1c(\%) = 96.3 \times HbA1c(\mu mol/L) / Hb(\mu mol/L) + 1.62$

LOT: 052522002

EXP: 2023-09-17

Abbreviated name	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD
HbA1c	BS-200 ¹	μmol/L	/	/ - /	/	BS-400 ⁹	μmol/L	/	/ - /	/
	BS-200E ²	μmol/L	/	/ - /	/	BS-430 ¹⁰	μmol/L	/	/ - /	/
	BS-230 ³	μmol/L	/	/ - /	/	BS-480 ¹¹	μmol/L	/	/ - /	/
	BS-240E ⁴	μmol/L	/	/ - /	/	BS-600 ¹²	μmol/L	/	/ - /	/
	BS-300 ⁵	μmol/L	/	/ - /	/	BS-600M ¹³	μmol/L	11.0	9.9 - 12.1	0.4
	BS-330E ⁶	μmol/L	/	/ - /	/	BS-800 ¹⁴	μmol/L	/	/ - /	/
	BS-360E ⁷	μmol/L	/	/ - /	/	BS-2000 ¹⁵	μmol/L	/	/ - /	/
Hb	BS-380 ⁸	μmol/L	/	/ - /	/	BS-2800M ¹⁶	μmol/L	/	/ - /	/
	BS-200 ¹	μmol/L	/	/ - /	/	BS-400 ⁹	μmol/L	/	/ - /	/
	BS-200E ²	μmol/L	/	/ - /	/	BS-430 ¹⁰	μmol/L	/	/ - /	/
	BS-230 ³	μmol/L	/	/ - /	/	BS-480 ¹¹	μmol/L	/	/ - /	/
	BS-240E ⁴	μmol/L	/	/ - /	/	BS-600 ¹²	μmol/L	/	/ - /	/
	BS-300 ⁵	μmol/L	/	/ - /	/	BS-600M ¹³	μmol/L	126	114 - 138	4
	BS-330E ⁶	μmol/L	/	/ - /	/	BS-800 ¹⁴	μmol/L	/	/ - /	/
BS-360E ⁷	μmol/L	/	/ - /	/	BS-2000 ¹⁵	μmol/L	/	/ - /	/	
BS-380 ⁸	μmol/L	/	/ - /	/	BS-2800M ¹⁶	μmol/L	/	/ - /	/	

HbA1c Control P



Abbreviated name	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range(Assay Value±3SD)	1 SD	
HbA1c ₁₇ (%)	IFCC	BS-200 ¹	%	/	/ - /	/	BS-400 ⁹	%	/	/ - /	/
		BS-200E ²	%	/	/ - /	/	BS-430 ¹⁰	%	/	/ - /	/
		BS-230 ³	%	/	/ - /	/	BS-480 ¹¹	%	/	/ - /	/
		BS-240E ⁴	%	/	/ - /	/	BS-600 ¹²	%	/	/ - /	/
		BS-300 ⁵	%	/	/ - /	/	BS-600M ¹³	%	8.75	7.88 - 9.62	0.29
		BS-330E ⁶	%	/	/ - /	/	BS-800 ¹⁴	%	/	/ - /	/
		BS-360E ⁷	%	/	/ - /	/	BS-2000 ¹⁵	%	/	/ - /	/
		BS-380 ⁸	%	/	/ - /	/	BS-2800M ¹⁶	%	/	/ - /	/
HbA1c ₁₇ (%)	NGSP / DCCT	BS-200 ¹	%	/	/ - /	/	BS-400 ⁹	%	/	/ - /	/
		BS-200E ²	%	/	/ - /	/	BS-430 ¹⁰	%	/	/ - /	/
		BS-230 ³	%	/	/ - /	/	BS-480 ¹¹	%	/	/ - /	/
		BS-240E ⁴	%	/	/ - /	/	BS-600 ¹²	%	/	/ - /	/
		BS-300 ⁵	%	/	/ - /	/	BS-600M ¹³	%	10.2	9.2 - 11.2	0.3
		BS-330E ⁶	%	/	/ - /	/	BS-800 ¹⁴	%	/	/ - /	/
		BS-360E ⁷	%	/	/ - /	/	BS-2000 ¹⁵	%	/	/ - /	/
		BS-380 ⁸	%	/	/ - /	/	BS-2800M ¹⁶	%	/	/ - /	/
	JCCLS	BS-200 ¹	%	/	/ - /	/	BS-400 ⁹	%	/	/ - /	/
		BS-200E ²	%	/	/ - /	/	BS-430 ¹⁰	%	/	/ - /	/
		BS-230 ³	%	/	/ - /	/	BS-480 ¹¹	%	/	/ - /	/
		BS-240E ⁴	%	/	/ - /	/	BS-600 ¹²	%	/	/ - /	/
		BS-300 ⁵	%	/	/ - /	/	BS-600M ¹³	%	10.0	9.0 - 11.0	0.3
		BS-330E ⁶	%	/	/ - /	/	BS-800 ¹⁴	%	/	/ - /	/
		BS-360E ⁷	%	/	/ - /	/	BS-2000 ¹⁵	%	/	/ - /	/
		BS-380 ⁸	%	/	/ - /	/	BS-2800M ¹⁶	%	/	/ - /	/

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione ±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

	HbA1c	Hb
English	Hemoglobin A1c	Hemoglobin
Русский	гемоглобин A1c	гемоглобин
Português	Hemoglobina A1c	Hemoglobina
Español	hemoglobina A1c	hemoglobina
Italiano	Emoglobina A1c	Emoglobina
Türkçe	Hemoglobin A1c	Hemoglobin