

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

6. **BS-400**: BS-400, BS-420;

2. **BS-230**: BS-230, BS-240, BS-280;

7. **BS-430**: BS-430, BS-450, BS-460;

3. **BS-240E**: BS-240E, BS-240Pro;

8. **BS-480**: BS-480, BS-490;

4. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**

9. **BS-600**: BS-600, BS-620;

(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

10. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

5. **BS-380**: BS-380, BS-390;

11. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

12. **BS-2800M**: BS-2600M, BS-2800M.

LOT: 058022002

⌚: 2023-02-28

Abbreviated name	Multimmun Control (L)						Multimmun Control (H)							
	Model	Unit	Assay Value	Range (AssayValue±3SD)		1 SD	Model	Unit	Assay Value	Range (AssayValue±3SD)		1 SD		
BS-200E ¹	ng/mL		117	99	—	135	6	BS-200E ¹	ng/mL	433	368	—	498	22
	pmol/L		263	222	—	303	13		pmol/L	973	827	—	1119	49
BS-230 ²	ng/mL		120	102	—	138	6	BS-230 ²	ng/mL	427	363	—	491	21
	pmol/L		270	229	—	310	13		pmol/L	959	816	—	1103	47
BS-240E ³	ng/mL		117	99	—	135	6	BS-240E ³	ng/mL	419	356	—	482	21
	pmol/L		263	222	—	303	13		pmol/L	941	800	—	1083	47
BS-360E ⁴	ng/mL		117	99	—	135	6	BS-360E ⁴	ng/mL	419	356	—	482	21
	pmol/L		263	222	—	303	13		pmol/L	941	800	—	1083	47
BS-380 ⁵	ng/mL		120	102	—	138	6	BS-380 ⁵	ng/mL	431	366	—	496	22
	pmol/L		270	229	—	310	13		pmol/L	968	822	—	1115	49
BS-400 ⁶	ng/mL		119	101	—	137	6	BS-400 ⁶	ng/mL	435	370	—	500	22
	pmol/L		267	227	—	308	13		pmol/L	977	831	—	1124	49
BS-430 ⁷	ng/mL		116	99	—	133	6	BS-430 ⁷	ng/mL	432	367	—	497	22
	pmol/L		261	222	—	299	13		pmol/L	971	825	—	1117	49
BS-480 ⁸	ng/mL		117	99	—	135	6	BS-480 ⁸	ng/mL	419	356	—	482	21
	pmol/L		263	222	—	303	13		pmol/L	941	800	—	1083	47
BS-600 ⁹	ng/mL		117	99	—	135	6	BS-600 ⁹	ng/mL	419	356	—	482	21
	pmol/L		263	222	—	303	13		pmol/L	941	800	—	1083	47
BS-800 ¹⁰	ng/mL		117	99	—	135	6	BS-800 ¹⁰	ng/mL	419	356	—	482	21
	pmol/L		263	222	—	303	13		pmol/L	941	800	—	1083	47
BS-2000 ¹¹	ng/mL		117	99	—	135	6	BS-2000 ¹¹	ng/mL	420	357	—	483	21
	pmol/L		263	222	—	303	13		pmol/L	944	802	—	1085	47
BS-2800M ¹²	ng/mL		114	97	—	131	6	BS-2800M ¹²	ng/mL	418	355	—	481	21
	pmol/L		256	218	—	294	13		pmol/L	939	798	—	1081	47

FER
(Remark 2:
The target value of FER in BS-230 is only applicable to BS-230 and BS-240)

Abbreviated name	Multimmun Control (L)						Multimmun Control (H)							
	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD		
MYO (Remark 2: The target value of MYO in BS-240 is only applicable to BS-240)	BS-240 ²	ng/mL	81.1	68.9	—	93.3	4.1	BS-240 ²	ng/mL	283	241	—	325	14
	BS-240E ³	ng/mL	82.3	70.0	—	94.6	4.1	BS-240E ³	ng/mL	285	242	—	328	14
	BS-360E ⁴	ng/mL	82.3	70.0	—	94.6	4.1	BS-360E ⁴	ng/mL	285	242	—	328	14
	BS-380 ⁵	ng/mL	81.8	69.5	—	94.1	4.1	BS-380 ⁵	ng/mL	288	245	—	331	14
	BS-400 ⁶	ng/mL	82.3	70.0	—	94.6	4.1	BS-400 ⁶	ng/mL	286	243	—	329	14
	BS-430 ⁷	ng/mL	82.2	69.9	—	94.5	4.1	BS-430 ⁷	ng/mL	283	241	—	325	14
	BS-480 ⁸	ng/mL	82.3	70.0	—	94.6	4.1	BS-480 ⁸	ng/mL	285	242	—	328	14
	BS-600 ⁹	ng/mL	82.3	70.0	—	94.6	4.1	BS-600 ⁹	ng/mL	285	242	—	328	14
	BS-800 ¹⁰	ng/mL	82.3	70.0	—	94.6	4.1	BS-800 ¹⁰	ng/mL	285	242	—	328	14
	BS-2000 ¹¹	ng/mL	81.9	69.6	—	94.2	4.1	BS-2000 ¹¹	ng/mL	283	241	—	325	14
	BS-2800M ¹²	ng/mL	82.7	70.3	—	95.1	4.1	BS-2800M ¹²	ng/mL	284	241	—	327	14
	IgE (Remark 2: The target value of IgE in BS-240 is only applicable to BS-240 and BS-280)	BS-240 ²	IU/mL	78.8	55.2	—	102.4	7.9	BS-240 ²	IU/mL	373	261	—	485
BS-240E ³		IU/mL	79.4	55.6	—	103.2	7.9	BS-240E ³	IU/mL	370	259	—	481	37
BS-360E ⁴		IU/mL	78.3	54.8	—	101.8	7.8	BS-360E ⁴	IU/mL	366	256	—	476	37
BS-380 ⁵		IU/mL	76.6	53.6	—	99.6	7.7	BS-380 ⁵	IU/mL	379	265	—	493	38
BS-400 ⁶		IU/mL	76.5	53.6	—	99.5	7.7	BS-400 ⁶	IU/mL	378	265	—	491	38
BS-480 ⁸		IU/mL	78.8	55.2	—	102.4	7.9	BS-480 ⁸	IU/mL	378	265	—	491	38
BS-600 ⁹		IU/mL	78.8	55.2	—	102.4	7.9	BS-600 ⁹	IU/mL	378	265	—	491	38
BS-800 ¹⁰		IU/mL	78.8	55.2	—	102.4	7.9	BS-800 ¹⁰	IU/mL	378	265	—	491	38
BS-2000 ¹¹	IU/mL	79.5	55.7	—	103.4	8.0	BS-2000 ¹¹	IU/mL	379	265	—	493	38	
BS-2800M ¹²	IU/mL	78.2	54.7	—	101.7	7.8	BS-2800M ¹²	IU/mL	391	274	—	508	39	

English	Abbreviated name	Multimmun Control	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)
Русский	сокращенное наименование	Контроль мультииммун	модель	Прибор	Результат анализа	Диапазон (результат анализа \pm 3CO)
Português	Nome abreviado	Controle Multimmun	Modelo	Прибор	Valores da análise	Faixa (Valores da análise \pm 3SD)
Español	nombre abreviado	Control multimmun	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo \pm 3SD)
Italiano	abbreviazione	Controllo Multimmun	modelli	Unità	Valori di dosaggio	Intervallo(valore di concentrazione \pm 3 DS)
Türkçe	kısaltılmış ad	Multimmun Kontrolü	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri \pm 3SD)

	FER	MYO	IgE
English	Ferritin	Myoglobin	Immunoglobulin E
Русский	ферритин	Миоглобин	Иммуноглобулин E
Português	Ferritina	Mioglobina	Imunoglobulina E
Español	Ferritina	Mioglobina	Inmunoglobulina E
Italiano	Ferritina	Mioglobina	Immunoglobulina E
Türkçe	Ferritin	Myoglobin	İmmünglobulin E