

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- 1.BS-200E: BS-200E, BS-220E,;
- 2.BS-240: BS-230, BS-240;
- 3.BS-240E: BS240E, BS240Pro;
- 4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");
- 5.BS-380: BS-380, BS-390;
- 6.BS-400: BS-400, BS-420;
- 7.BS-430: BS-430, BS-450,BS-460;

- 8.BS-480: BS-480, BS-490;
- 9.BS-600: BS-600, BS-620;
- 10.BS-600M: BS-600M, BS-620M;
- 11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
- 12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
- 13.BS-2800M:BS-2600M, BS-2800M;
- 14.Applicable models of the chemistry shall be subject to the parameter list and instructions.

LOT : 059123008

Exp. Date: 2025-03-14

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

β2-MG

English	β2-Microglobulin
Русский	β2-микροглобулина
Português	β2-Microglobulina
Español	β2-microglobulina
Italiano	β2-Microglobulina
Türkçe	β2-Mikroglobulin

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Assay Value	Range (Assay Value±3SD)		1 SD		
β2-MG II	mg/L	BS-200E ¹	1.05	0.90	—	1.20	0.05	BS-200E ¹	2.90	2.48	—	3.32	0.14
		BS-240 ²	1.06	0.91	—	1.21	0.05	BS-240 ²	2.89	2.47	—	3.31	0.14
		BS-240E ³	1.06	0.91	—	1.21	0.05	BS-240E ³	2.87	2.45	—	3.29	0.14
		BS-360E ⁴	1.04	0.89	—	1.19	0.05	BS-360E ⁴	2.91	2.46	—	3.36	0.15
		BS-380 ⁵	1.05	0.90	—	1.20	0.05	BS-380 ⁵	2.90	2.48	—	3.32	0.14
		BS-400 ⁶	1.05	0.90	—	1.20	0.05	BS-400 ⁶	2.90	2.48	—	3.32	0.14
		BS-430 ⁷	1.06	0.91	—	1.21	0.05	BS-430 ⁷	2.94	2.49	—	3.39	0.15
		BS-480 ⁸	1.04	0.89	—	1.19	0.05	BS-480 ⁸	2.90	2.48	—	3.32	0.14
		BS-600 ⁹	1.01	0.86	—	1.16	0.05	BS-600 ⁹	2.91	2.46	—	3.36	0.15
		BS-600M ¹⁰	1.06	0.91	—	1.21	0.05	BS-600M ¹⁰	2.90	2.48	—	3.32	0.14
		BS-800 ¹¹	1.08	0.93	—	1.23	0.05	BS-800 ¹¹	2.96	2.51	—	3.41	0.15
		BS-2000 ¹²	1.07	0.92	—	1.22	0.05	BS-2000 ¹²	3.01	2.56	—	3.46	0.15
		BS-2800M ¹³	1.06	0.91	—	1.21	0.05	BS-2800M ¹³	2.90	2.48	—	3.32	0.14
	nmol/L	BS-200E ¹	88.9	76.2	—	101.6	4.2	BS-200E ¹	246	210	—	281	12
		BS-240 ²	89.8	77.1	—	102.5	4.2	BS-240 ²	245	209	—	280	12
		BS-240E ³	89.8	77.1	—	102.5	4.2	BS-240E ³	243	208	—	279	12
		BS-360E ⁴	88.1	75.4	—	100.8	4.2	BS-360E ⁴	246	208	—	285	13
		BS-380 ⁵	88.9	76.2	—	101.6	4.2	BS-380 ⁵	246	210	—	281	12

β2-Microglobulin Control



β2-MG II	nmol/L	BS-400⁶	88.9	76.2	—	101.6	4.2	BS-400⁶	246	210	—	281	12
		BS-430⁷	89.8	77.1	—	102.5	4.2	BS-430⁷	249	211	—	287	13
		BS-480⁸	88.1	75.4	—	100.8	4.2	BS-480⁸	246	210	—	281	12
		BS-600⁹	85.5	72.8	—	98.3	4.2	BS-600⁹	246	208	—	285	13
		BS-600M¹⁰	89.8	77.1	—	102.5	4.2	BS-600M¹⁰	246	210	—	281	12
		BS-800¹¹	91.5	78.8	—	104.2	4.2	BS-800¹¹	251	213	—	289	13
		BS-2000¹²	90.6	77.9	—	103.3	4.2	BS-2000¹²	255	217	—	293	13
		BS-2800M¹³	89.8	77.1	—	102.5	4.2	BS-2800M¹³	246	210	—	281	12

mindray

β 2-Microglobulin Control

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 059123008

2025-03-14

