

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;

8.BS-480: BS-480, BS-490;

2.BS-240: BS-230, BS-240;

9.BS-600: BS-600, BS-620;

3.BS-240E: BS240E, BS240Pro;

10.BS-600M: BS-600M, BS-620M;

4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-

11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-

330E(V35.00)(Serial Number starts with "W8-" and software version

1800plus;

starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-"

12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

and software version starts with "35.00");

13.BS-2800M:BS-2600M, BS-2800M;

5.BS-380: BS-380, BS-390;

14.Applicable models of the chemistry shall be subject to the

6.BS-400: BS-400, BS-420;

parameter list and instructions.

7.BS-430: BS-430, BS-450,BS-460;

LOT : 059223008

: 2025-03-24

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value $\pm$ 3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа $\pm$ 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise $\pm$ 3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo $\pm$ 3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore di concentrazione $\pm$ 3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri $\pm$ 3SD)

## ASO

English	Antistreptolysin "O"
Русский	антистрептолизина O
Português	Antiestreptolisina "O"
Español	antiestreptolisina "O"
Italiano	Anti-Streptolisina "O"
Türkçe	Antistreptolisin "O"

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	Model	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD		
ASO II	IU/mL	BS-200E <sup>1</sup>	123	81	—	165	14	BS-200E <sup>1</sup>	328	214	—	442	38
		BS-240 <sup>2</sup>	126	81	—	171	15	BS-240 <sup>2</sup>	331	214	—	448	39
		BS-240E <sup>3</sup>	123	81	—	165	14	BS-240E <sup>3</sup>	328	214	—	442	38
		BS-360E <sup>4</sup>	123	81	—	165	14	BS-360E <sup>4</sup>	328	214	—	442	38
		BS-380 <sup>5</sup>	123	81	—	165	14	BS-380 <sup>5</sup>	328	214	—	442	38
		BS-400 <sup>6</sup>	123	81	—	165	14	BS-400 <sup>6</sup>	328	214	—	442	38
		BS-430 <sup>7</sup>	123	81	—	165	14	BS-430 <sup>7</sup>	328	214	—	442	38
		BS-480 <sup>8</sup>	123	81	—	165	14	BS-480 <sup>8</sup>	328	214	—	442	38
		BS-600 <sup>9</sup>	123	81	—	165	14	BS-600 <sup>9</sup>	328	214	—	442	38
		BS-600M <sup>10</sup>	119	77	—	161	14	BS-600M <sup>10</sup>	321	207	—	435	38
		BS-800 <sup>11</sup>	123	81	—	165	14	BS-800 <sup>11</sup>	328	214	—	442	38
		BS-2000 <sup>12</sup>	119	77	—	161	14	BS-2000 <sup>12</sup>	321	207	—	435	38
		BS-2800M <sup>13</sup>	119	77	—	161	14	BS-2800M <sup>13</sup>	321	207	—	435	38

Abbreviated name	Unit	Level: L/N					Level: H/P						
		Model	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	Model	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD		
RF II	IU/mL	BS-200E <sup>1</sup>	35.8	23.2	—	48.4	4.2	BS-200E <sup>1</sup>	83.0	53.9	—	112.1	9.7
		BS-240 <sup>2</sup>	35.0	22.7	—	47.3	4.1	BS-240 <sup>2</sup>	85.0	55.3	—	114.7	9.9

RF II	IU/mL	<b>BS-240E<sup>3</sup></b>	35.8	23.2	—	48.4	4.2	<b>BS-240E<sup>3</sup></b>	84.1	54.7	—	113.5	9.8
		<b>BS-360E<sup>4</sup></b>	36.3	23.7	—	48.9	4.2	<b>BS-360E<sup>4</sup></b>	84.1	54.7	—	113.5	9.8
		<b>BS-380<sup>5</sup></b>	37.0	24.1	—	49.9	4.3	<b>BS-380<sup>5</sup></b>	85.1	55.1	—	115.1	10.0
		<b>BS-400<sup>6</sup></b>	37.0	24.1	—	49.9	4.3	<b>BS-400<sup>6</sup></b>	85.1	55.1	—	115.1	10.0
		<b>BS-430<sup>7</sup></b>	36.3	23.7	—	48.9	4.2	<b>BS-430<sup>7</sup></b>	84.1	54.7	—	113.5	9.8
		<b>BS-480<sup>8</sup></b>	36.3	23.7	—	48.9	4.2	<b>BS-480<sup>8</sup></b>	84.1	54.7	—	113.5	9.8
		<b>BS-600<sup>9</sup></b>	35.1	22.8	—	47.4	4.1	<b>BS-600<sup>9</sup></b>	79.8	51.9	—	107.7	9.3
		<b>BS-600M<sup>10</sup></b>	36.5	23.6	—	49.4	4.3	<b>BS-600M<sup>10</sup></b>	83.6	54.2	—	113.0	9.8
		<b>BS-800<sup>11</sup></b>	36.3	23.7	—	48.9	4.2	<b>BS-800<sup>11</sup></b>	84.1	54.7	—	113.5	9.8
		<b>BS-2000<sup>12</sup></b>	34.2	22.2	—	46.2	4.0	<b>BS-2000<sup>12</sup></b>	79.1	51.2	—	107.0	9.3
		Abbreviated name	Unit	Level: L/N					Level: H/P				
				Model	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD	Model	Assay Value	Range (Assay Value $\pm$ 3SD)		1 SD
RF III	IU/mL	<b>BS-800<sup>11</sup></b>	44.8	29.2	—	60.4	5.2	<b>BS-800<sup>11</sup></b>	103	67	—	139	12
		<b>BS-2000<sup>12</sup></b>	44.4	28.8	—	60.0	5.2	<b>BS-2000<sup>12</sup></b>	107	68	—	146	13
		<b>BS-2800M<sup>13</sup></b>	43.0	28.0	—	58.0	5.0	<b>BS-2800M<sup>13</sup></b>	101	65	—	137	12

# mindray

## ASO/CRP/RF Triple Control

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

LOT 059223008

2025-03-24

