

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- | | |
|---|--|
| <p>1. BS-120: BS-120, BS-130;</p> <p>2. BS-180: BS-180, BS-190;</p> <p>3. BS-200: BS-200, BS-220;</p> <p>4. BS-200E: BS-200E, BS-220E;</p> <p>5. BS-230: BS-230, BS-240, BS-280;</p> <p>6. BS-240E: BS-240E, BS-240Pro;</p> <p>7. BS-300: BS-300, BS-320;</p> <p>8. BS-330: BS-330, BS-350;</p> <p>9. BS-330E: BS-330E (Serial Number starts with "XQ-"),
BS-350E (Serial Number starts with "XS-");</p> | <p>10. BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");</p> <p>11. BS-380: BS-380, BS-390;</p> <p>12. BS-400: BS-400, BS-420;</p> <p>13. BS-430: BS-410, BS-430, BS-450, BS-460, BS-470;</p> <p>14. BS-480: BS-480, BS-490;</p> <p>15. BS-600: BS-600, BS-620;</p> <p>16. BS-600M: BS-600M, BS-620M;</p> <p>17. BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;</p> <p>18. BS-830: BS-830, BS-830S, BS-840, BS-850, BS-860, BS-870;</p> <p>19. BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M.</p> <p>20. BS-2800M: BS-2600M, BS-2800M</p> |
|---|--|

LOT: 061522005

: 2023-07-24

Abbreviated name	ADA Control(L)						ADA Control(H)							
	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD		
ADA	BS-120	¹ U/L	27.7	22.7	—	32.7	1.7	BS-120	¹ U/L	124	102	—	146	7
	BS-180	² U/L	27.7	22.7	—	32.7	1.7	BS-180	² U/L	124	102	—	146	7
	BS-200	³ U/L	27.8	22.8	—	32.8	1.7	BS-200	³ U/L	126	103	—	149	8
	BS-200E	⁴ U/L	27.4	22.5	—	32.3	1.6	BS-200E	⁴ U/L	125	103	—	148	8
	BS-230	⁵ U/L	27.2	22.3	—	32.1	1.6	BS-230	⁵ U/L	122	100	—	144	7
	BS-240E	⁶ U/L	27.2	22.3	—	32.1	1.6	BS-240E	⁶ U/L	126	103	—	149	8
	BS-300	⁷ U/L	27.3	22.4	—	32.2	1.6	BS-300	⁷ U/L	124	102	—	146	7
	BS-330	⁸ U/L	27.8	22.8	—	32.8	1.7	BS-330	⁸ U/L	126	103	—	149	8
	BS-330E	⁹ U/L	27.4	22.5	—	32.3	1.6	BS-330E	⁹ U/L	125	103	—	148	8
	BS-360E	¹⁰ U/L	27.2	22.3	—	32.1	1.6	BS-360E	¹⁰ U/L	124	102	—	146	7
	BS-380	¹¹ U/L	27.1	22.2	—	32.0	1.6	BS-380	¹¹ U/L	125	103	—	148	8
	BS-400	¹² U/L	27.2	22.3	—	32.1	1.6	BS-400	¹² U/L	125	103	—	148	8
	BS-430	¹³ U/L	27.2	22.3	—	32.1	1.6	BS-430	¹³ U/L	125	103	—	148	8
	BS-480	¹⁴ U/L	27.1	22.2	—	32.0	1.6	BS-480	¹⁴ U/L	124	102	—	146	7
	BS-600	¹⁵ U/L	27.3	22.4	—	32.2	1.6	BS-600	¹⁵ U/L	125	103	—	148	8
	BS-600M	¹⁶ U/L	/	/	—	/	/	BS-600M	¹⁶ U/L	/	/	—	/	/
	BS-800	¹⁷ U/L	27.2	22.3	—	32.1	1.6	BS-800	¹⁷ U/L	125	103	—	148	8
	BS-830	¹⁸ U/L	27.1	22.2	—	32.0	1.6	BS-830	¹⁸ U/L	125	103	—	148	8
	BS-2000	¹⁹ U/L	27.5	22.6	—	32.5	1.7	BS-2000	¹⁹ U/L	127	104	—	150	8
	BS-2800M	²⁰ U/L	27.4	22.5	—	32.3	1.6	BS-2800M	²⁰ U/L	127	104	—	150	8

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон(результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa(Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango(Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo(valore diconcentrazione±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

ADA	
English	Adenosine Deaminase
Русский	является аденозиндезаминаза
Português	Adenosina deaminase
Español	adenosina desaminasa
Italiano	adenosina deamminasi
Türkçe	Adenozin Deaminazdır