

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

- | | |
|---|--|
| <p>1. BS-120: BS-120, BS-130;</p> <p>2. BS-180: BS-180, BS-190;</p> <p>3. BS-200: BS-200, BS-220;</p> <p>4. BS-200E: BS-200E, BS-220E;</p> <p>5. BS-230: BS-230, BS-240, BS-280;</p> <p>6. BS-240E: BS-240E, BS-240Pro;</p> <p>7. BS-300: BS-300, BS-320;</p> <p>8. BS-330: BS-330, BS-350;</p> <p>9. BS-330E: BS-330E (Serial Number starts with "XQ-"),
BS-350E (Serial Number starts with "XS-");</p> | <p>10. BS-360E: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00) (Serial Number starts with "W8-"
and software version starts with "35.00"),
BS-350E(V35.00) (Serial Number starts with "W9-"
and software version starts with "35.00");</p> <p>11. BS-380: BS-380, BS-390;</p> <p>12. BS-400: BS-400, BS-420;</p> <p>13. BS-430: BS-410, BS-430, BS-450, BS-460, BS-470;</p> <p>14. BS-480: BS-480, BS-490;</p> <p>15. BS-600: BS-600, BS-620;</p> <p>16. BS-800: BS-800, BS-820, BS-800M, BS-820M,
BS-1800, BS-1800plus;</p> <p>17. BS-830: BS-830, BS-830S, BS-840, BS-850, BS-860,
BS-870;</p> <p>18. BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M.</p> <p>19. BS-2800M: BS-2600M, BS-2800M.</p> |
|---|--|

LOT: 061822001

: 2023-04-28

Abbreviate d name	CO2 Control(L)						CO2 Control(H)					
	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD		Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD	
CO2	BS-120	¹ mmol/L	/	/ - /	/		BS-120	¹ mmol/L	/	/ - /	/	
	BS-180	² mmol/L	/	/ - /	/		BS-180	² mmol/L	/	/ - /	/	
	BS-200	³ mmol/L	12.2	9.7 - 14.7	0.8		BS-200	³ mmol/L	13.2	10.5 - 15.9	0.9	
	BS-200E	⁴ mmol/L	12.5	10.0 - 15.0	0.9		BS-200E	⁴ mmol/L	12.8	10.2 - 15.4	0.9	
	BS-230	⁵ mmol/L	12.8	10.2 - 15.4	0.9		BS-230	⁵ mmol/L	12.9	10.3 - 15.5	0.9	
	BS-240E	⁶ mmol/L	12.8	10.2 - 15.4	0.9		BS-240E	⁶ mmol/L	13.4	10.7 - 16.1	0.9	
	BS-300	⁷ mmol/L	12.1	9.7 - 14.5	0.8		BS-300	⁷ mmol/L	12.2	9.7 - 14.7	0.8	
	BS-330	⁸ mmol/L	12.2	9.7 - 14.7	0.8		BS-330	⁸ mmol/L	13.2	10.5 - 15.9	0.9	
	BS-330E	⁹ mmol/L	12.5	10.0 - 15.0	0.9		BS-330E	⁹ mmol/L	12.8	10.2 - 15.4	0.9	
	BS-360E	¹⁰ mmol/L	12.8	10.2 - 15.4	0.9		BS-360E	¹⁰ mmol/L	13.5	10.8 - 16.2	0.9	
	BS-380	¹¹ mmol/L	12.5	10.0 - 15.0	0.9		BS-380	¹¹ mmol/L	13.0	10.4 - 15.6	0.9	
	BS-400	¹² mmol/L	12.7	10.1 - 15.3	0.9		BS-400	¹² mmol/L	13.1	10.5 - 15.7	0.9	
	BS-430	¹³ mmol/L	12.8	10.2 - 15.4	0.9		BS-430	¹³ mmol/L	13.8	11.0 - 16.6	0.9	
	BS-480	¹⁴ mmol/L	12.8	10.2 - 15.4	0.9		BS-480	¹⁴ mmol/L	13.7	10.9 - 16.5	0.9	
	BS-600	¹⁵ mmol/L	12.8	10.2 - 15.4	0.9		BS-600	¹⁵ mmol/L	13.2	10.5 - 15.9	0.9	
	BS-800	¹⁶ mmol/L	12.8	10.2 - 15.4	0.9		BS-800	¹⁶ mmol/L	13.9	11.1 - 16.7	0.9	
	BS-830	¹⁷ mmol/L	12.6	10.1 - 15.1	0.8		BS-830	¹⁷ mmol/L	12.6	10.1 - 15.1	0.8	
	BS-2000	¹⁸ mmol/L	12.7	10.1 - 15.3	0.9		BS-2000	¹⁸ mmol/L	14.0	11.2 - 16.8	0.9	
	BS-2800M	¹⁹ mmol/L	13.0	10.4 - 15.6	0.9		BS-2800M	¹⁹ mmol/L	14.1	11.3 - 16.9	0.9	

Abbreviate d name	TBA Control(L)					TBA Control(H)				
	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD
TBA	BS-120	¹ μmol/L	21.2	16.9 — 25.5	1.4	BS-120	¹ μmol/L	38.4	30.7 — 46.1	2.6
	BS-180	² μmol/L	21.2	16.9 — 25.5	1.4	BS-180	² μmol/L	38.4	30.7 — 46.1	2.6
	BS-200	³ μmol/L	20.6	16.5 — 24.7	1.4	BS-200	³ μmol/L	38.4	30.7 — 46.1	2.6
	BS-200E	⁴ μmol/L	21.5	17.2 — 25.8	1.4	BS-200E	⁴ μmol/L	37.9	30.3 — 45.5	2.5
	BS-230	⁵ μmol/L	21.2	16.9 — 25.5	1.4	BS-230	⁵ μmol/L	38.0	30.4 — 45.6	2.5
	BS-240E	⁶ μmol/L	21.2	16.9 — 25.5	1.4	BS-240E	⁶ μmol/L	38.6	30.8 — 46.4	2.6
	BS-300	⁷ μmol/L	21.6	17.3 — 25.9	1.4	BS-300	⁷ μmol/L	38.2	30.5 — 45.9	2.6
	BS-330	⁸ μmol/L	20.6	16.5 — 24.7	1.4	BS-330	⁸ μmol/L	38.4	30.7 — 46.1	2.6
	BS-330E	⁹ μmol/L	21.5	17.2 — 25.8	1.4	BS-330E	⁹ μmol/L	37.9	30.3 — 45.5	2.5
	BS-360E	¹⁰ μmol/L	21.2	16.9 — 25.5	1.4	BS-360E	¹⁰ μmol/L	38.6	30.8 — 46.4	2.6
	BS-380	¹¹ μmol/L	21.5	17.2 — 25.8	1.4	BS-380	¹¹ μmol/L	39.0	31.2 — 46.8	2.6
	BS-400	¹² μmol/L	21.6	17.3 — 25.9	1.4	BS-400	¹² μmol/L	38.8	31.0 — 46.6	2.6
	BS-430	¹³ μmol/L	21.2	16.9 — 25.5	1.4	BS-430	¹³ μmol/L	39.5	31.6 — 47.4	2.6
	BS-480	¹⁴ μmol/L	21.2	16.9 — 25.5	1.4	BS-480	¹⁴ μmol/L	39.0	31.2 — 46.8	2.6
	BS-600	¹⁵ μmol/L	21.2	16.9 — 25.5	1.4	BS-600	¹⁵ μmol/L	38.3	30.6 — 46.0	2.6
	BS-800	¹⁶ μmol/L	21.2	16.9 — 25.5	1.4	BS-800	¹⁶ μmol/L	39.2	31.3 — 47.1	2.6
	BS-830	¹⁷ μmol/L	21.5	17.2 — 25.8	1.4	BS-830	¹⁷ μmol/L	39.4	31.5 — 47.3	2.6
	BS-2000	¹⁸ μmol/L	21.8	17.4 — 26.2	1.5	BS-2000	¹⁸ μmol/L	40.0	32.0 — 48.0	2.7
	BS-2800M	¹⁹ μmol/L	21.1	16.9 — 25.3	1.4	BS-2800M	¹⁹ μmol/L	39.1	31.2 — 47.0	2.6

English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа ± 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise ±3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo ±3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione ±3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri±3SD)

	CO2	TBA
English	Carbon Dioxide	Total Bile Acids
Русский	Двуокись углерода	общие желчные кислоты
Português	Dióxido de Carbono	ácidos biliares totales
Español	Dióxido de carbono	ácidos biliares totais
Italiano	Anidride carbonica	acidi biliari totali
Türkçe	Karbon Dioksit	total safra asitleridir