

## Reference Value Table

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation /Analyte	Unit	Reference Value	Scope
Coagulation control (level 2)		CX series Auto Coagulation Analyzers of Shenzhen Mindray Bio-medical Electronics Co., Ltd	PT	s	34.12	27.30 ~ 40.94
	8724020118		APTT	s	60.47	48.38 ~ 72.56
			TT	s	/	/
	2025-02-27		FIB	g/L	/	/
				mg/dL	/	/

Product Name	Batch code And Use-by date	Applicable Instruments	Abbreviation /Analyte	Unit	Reference Value	Scope
Coagulation control (level 2)		C3510 and ExC810 Auto Coagulation Analyzers of Beijing Mindray Medical Instrument Co., Ltd	PT	s	31.54	25.23 ~ 37.85
	8724020118		APTT	s	63.90	51.12 ~ 76.68
			TT	s	/	/
	2025-02-27		FIB	g/L	/	/
				mg/dL	/	/