

Anti-HBc Positive Control



Target Value Sheet

Product	Lot Number and Expiration Date	Instrument Model	Unit	Target Value	Target Value and Range	1SD
Anti-HBc Positive Control	<div>LOT</div> <div>2023020111</div> <div></div> <div>2024-09-13</div>	Mindray CL-series Chemiluminescence Immunoassay Analyzer	COI	0.49	0.25 ~ 0.73	0.08