

HIV Ag/Ab Negative Control



Target Value Sheet

Product	Lot Number and Expiration Date	Instrument Model	Unit	Target Value	Target Value and Range	1SD
HIV Ag/Ab Negative Control	<div>LOT</div> <div>2022110111</div> <div></div> <div>2024-06-16</div>	Mindray CL-series Chemiluminescence Immunoassay Analyzer	COI	0.07	0.00 ~ 0.46	0.13