

HIV Ag/Ab Positive Control



Target Value Sheet

Product	Lot Number and Expiration Date	Instrument Model	Unit	Target Value	Target Value and Range	1SD
HIV Ag Positive Control	<div><div>LOT</div><div>2022110111</div></div>	Mindray CL-series Chemiluminescence Immunoassay Analyzer	COI	6.13	2.77 ~ 9.49	1.12
HIV Ab Positive Control	<div><div></div><div>2024-06-16</div></div>		COI	4.61	2.06 ~ 7.16	0.85