

ToRCH IgG/IgM Negative Control

Target Value Sheet

Product	Lot Number and Expiration Date	Analytes	Instrument Model	Unit	Target Values	Target Values and Ranges (Target Values±3SD)			1SD	International Unit	Target Values	Target Values and Ranges (Target Values±3SD)			1SD
ToRCH IgG/IgM Negative Control	LOT  2023050111	RV IgG	CL-900i Series	IU/mL	0.00	0.00	~	3.99	1.33	/	/	/	~	/	/
			CL-8000i Series		0.00	0.00	~	3.99	1.33		/	/	~	/	/
			CL-2000i/CL-1000i/CL-6000i Series		0.00	0.00	~	3.99	1.33		/	/	~	/	/
		Toxo IgG	CL-900i Series	IU/mL	0.00	0.00	~	0.81	0.27	/	/	/	~	/	/
			CL-8000i Series		0.00	0.00	~	0.81	0.27		/	/	~	/	/
			CL-2000i/CL-1000i/CL-6000i Series		0.00	0.00	~	0.81	0.27		/	/	~	/	/
		CMV IgG	CL-900i Series	IU/mL	0.00	0.00	~	0.45	0.15	/	/	/	~	/	/
			CL-8000i Series		0.00	0.00	~	0.45	0.15		/	/	~	/	/
			CL-2000i/CL-1000i/CL-6000i Series		0.00	0.00	~	0.45	0.15		/	/	~	/	/
		RV IgM	CL-900i Series	COI	0.27	0.00	~	0.69	0.14	/	/	/	~	/	/
			CL-8000i Series		0.28	0.00	~	0.70	0.14		/	/	~	/	/
			CL-2000i/CL-1000i/CL-6000i Series		0.28	0.00	~	0.70	0.14		/	/	~	/	/
		Toxo IgM	CL-900i Series	COI	0.18	0.00	~	0.60	0.14	/	/	/	~	/	/
			CL-8000i Series		0.20	0.00	~	0.62	0.14		/	/	~	/	/
			CL-2000i/CL-1000i/CL-6000i Series		0.20	0.00	~	0.62	0.14		/	/	~	/	/
	2024/12/12	CMV IgM	CL-900i Series	COI	0.07	0.00	~	0.49	0.14	/	/	/	~	/	/
			CL-8000i Series		0.07	0.00	~	0.49	0.14		/	/	~	/	/
			CL-2000i/CL-1000i/CL-6000i Series		0.07	0.00	~	0.49	0.14		/	/	~	/	/
		/	/	/	/	/	~	/	/	/	/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/
		/	/	/	/	/	~	/	/	/	/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/
		/	/	/	/	/	~	/	/	/	/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/
		/	/	/	/	/	~	/	/	/	/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/
			/		/	/	~	/	/		/	/	~	/	/

Remark: