

# HS-CRP Calibrator

## Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray  
Português : Planilha de valores do Sistema de Medição BS da Mindray  
Español : Hoja de valores del sistema de medición Mindray BS  
Italiano : Scheda dei valori del sistema di misurazione BS di Mindray  
Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский : Данные совпадают во всех группах.  
Español : la datos de cada grupo es la misma.

Português : A dados de cada grupo é a mesma.  
Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-240**;

3. **BS-240E**: BS240E, BS240Pro;

4. **BS-300**: BS-300, BS-320;

5. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),  
**BS-350E**(Serial Number starts with "XS-");

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,  
**BS-330E(V35.00)**(Serial Number starts with "W8-"  
and software version starts with "35.00"),  
**BS-350E(V35.00)** (Serial Number starts with "W9-"  
and software version starts with "35.00");

7. **BS-380**: BS-380, BS-390;

8. **BS-400**: BS-400, BS-420;

9. **BS-430**: BS-450, BS-460;

10. **BS-480**: BS-480, BS-490;

11. **BS-600**: BS-600, BS-620;

12. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,  
BS-1800, BS-1800plus;

13. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

14. **BS-2800M**: BS-2600M, BS-2800M.

15. **S0: 0.9% NaCl, Conc. Of S0=0;**

Português: S0: 0,9% NaCl, Conc. de S0=0;

Italiano: S0: 0,9% NaCl, conc. di S0=0;

Русский: S0: 0, 9% NaCl, конц. S0=0;

Español: S0: 0,9% NaCl, Conc. de S0=0;

Türkçe: S0: %0,9 NaCl, S0 Kons.=0.

**LOT** : 043422001



: 2023-07-12

Abbreviated name	HS-CRP	Calibration Rule	spline
Model	Level	Calibrator Value <sup>15</sup>	
		mg/L	nmol/L
BS-200E <sup>1</sup>	a	4.38	41.7
	b	19.2	183
	c	39.6	377
	d	152	1447
	e	320	3046
BS-240 <sup>2</sup>	a	4.38	41.7
	b	19.2	183
	c	39.6	377
	d	152	1447
	e	320	3046
BS-240E <sup>3</sup>	a	5.80	55.2
	b	19.0	181
	c	38.6	367
	d	152	1447
	e	310	2951
BS-300 <sup>4</sup>	a	4.38	41.7
	b	19.2	183
	c	39.6	377
	d	152	1447
	e	320	3046
BS-330E <sup>5</sup>	a	4.38	41.7
	b	19.2	183
	c	39.6	377
	d	152	1447
	e	320	3046
BS-360E <sup>6</sup>	a	4.38	41.7
	b	19.2	183
	c	39.6	377
	d	152	1447
	e	320	3046

# HS-CRP Calibrator

Abbreviated name	HS-CRP	Calibration Rule	spline	
Model	Level	Calibrator Value <sup>15</sup>		
		mg/L	nmol/L	
BS-380 <sup>7</sup>	a	4.38	41.7	
	b	19.2	183	
	c	39.6	377	
	d	152	1447	
	e	320	3046	
BS-400 <sup>8</sup>	a	4.38	41.7	
	b	19.2	183	
	c	39.6	377	
	d	152	1447	
	e	320	3046	
BS-430 <sup>9</sup>	a	4.38	41.7	
	b	19.2	183	
	c	39.6	377	
	d	152	1447	
	e	320	3046	
BS-480 <sup>10</sup>	a	4.38	41.7	
	b	19.2	183	
	c	39.6	377	
	d	152	1447	
	e	320	3046	
BS-600 <sup>11</sup>	a	4.38	41.7	
	b	19.2	183	
	c	39.6	377	
	d	152	1447	
	e	320	3046	
BS-800 <sup>12</sup>	a	4.38	41.7	
	b	19.2	183	
	c	39.6	377	
	d	152	1447	
	e	320	3046	
BS-2000 <sup>13</sup>	a	4.42	42.1	
	b	18.4	175	
	c	39.1	372	
	d	155	1476	
	e	313	2980	
BS-2800M <sup>14</sup>	a	4.28	40.7	
	b	18.4	175	
	c	38.8	369	
	d	153	1457	
	e	310	2951	

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

HS-CRP	
English	High Sensitivity C-reaction Protein
Русский	C-реактивный белок
Português	Proteína C Reativa de Alta Sensibilidade
Español	proteína C reactiva de alta sensibilidad
Italiano	proteina C reattiva ad alta sensibilità
Türkçe	Yüksek Duyarlıklılı C Reaksiyon Proteindir