

HS-CRP Calibrator

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
Português : Planilha de valores do Sistema de Medição BS da Mindray
Español : Hoja de valores del sistema de medición Mindray BS
Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский : Данные совпадают во всех группах.
Español : la datos de cada grupo es la misma.

Português : A dados de cada grupo é a mesma.
Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-240**;

3. **BS-240E**: BS240E, BS240Pro;

4. **BS-300**: BS-300, BS-320;

5. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-");

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00)(Serial Number starts with "W8-"
and software version starts with "35.00"),
BS-350E(V35.00) (Serial Number starts with "W9-"
and software version starts with "35.00");

7. **BS-380**: BS-380, BS-390;

8. **BS-400**: BS-400, BS-420;

9. **BS-430**: BS-450, BS-460;

10. **BS-480**: BS-480, BS-490;

11. **BS-600**: BS-600, BS-620;

12. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,
BS-1800, BS-1800plus;

13. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

14. **BS-2800M**: BS-2600M, BS-2800M.

15. **S0: 0.9% NaCl, Conc. Of S0=0;**

Português: S0: 0,9% NaCl, Conc. de S0=0;

Italiano: S0: 0,9% NaCl, conc. di S0=0;

Русский: S0: 0, 9% NaCl, конц. S0=0;

Español: S0: 0,9% NaCl, Conc. de S0=0;

Türkçe: S0: %0,9 NaCl, S0 Kons.=0.

LOT : 043422002

: 2023-07-21

| Abbreviated name | HS-CRP | Calibration Rule | spline |
|----------------------|--------|--------------------------------|--------|
| Model | Level | Calibrator Value ¹⁵ | |
| | | mg/L | nmol/L |
| BS-200E ¹ | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |
| BS-240 ² | a | 4.80 | 45.7 |
| | b | 19.4 | 185 |
| | c | 40.3 | 384 |
| | d | 168 | 1599 |
| | e | 313 | 2980 |
| BS-240E ³ | a | 5.63 | 53.6 |
| | b | 18.8 | 179 |
| | c | 39.5 | 376 |
| | d | 158 | 1504 |
| | e | 313 | 2980 |
| BS-300 ⁴ | a | 4.61 | 43.9 |
| | b | 19.2 | 183 |
| | c | 39.9 | 380 |
| | d | 158 | 1504 |
| | e | 313 | 2980 |
| BS-330E ⁵ | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |
| BS-360E ⁶ | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |

HS-CRP Calibrator

| Abbreviated name | HS-CRP | Calibration Rule | spline |
|------------------------|--------|--------------------------------|--------|
| Model | Level | Calibrator Value ¹⁵ | |
| | | mg/L | nmol/L |
| BS-380 ⁷ | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |
| BS-400 ⁸ | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |
| BS-430 ⁹ | a | 4.66 | 44.4 |
| | b | 19.5 | 186 |
| | c | 39.9 | 380 |
| | d | 160 | 1523 |
| | e | 313 | 2980 |
| BS-480 ¹⁰ | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |
| BS-600 ¹¹ | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |
| BS-800 ¹² | a | 4.52 | 43.0 |
| | b | 18.8 | 179 |
| | c | 38.7 | 368 |
| | d | 155 | 1476 |
| | e | 313 | 2980 |
| BS-2000 ¹³ | a | 4.66 | 44.4 |
| | b | 19.3 | 184 |
| | c | 39.9 | 380 |
| | d | 157 | 1495 |
| | e | 310 | 2951 |
| BS-2800M ¹⁴ | a | 4.34 | 41.3 |
| | b | 18.5 | 176 |
| | c | 39.0 | 371 |
| | d | 154 | 1466 |
| | e | 318 | 3027 |

| English | Abbreviated name | Calibration Rule | Calibration Value | Model | Level |
|-----------|--------------------------|------------------------|------------------------|---------|---------|
| Русский | сокращенное наименование | Принцип калибровки | Эталонное значение | модель | Уровень |
| Português | Nome abreviado | Regra de calibração | Valor de calibração | Modelo | Nível |
| Español | nombre abreviado | Regla de calibración | Valor de calibración | modelo | Nivel |
| Italiano | abbreviazione | Regola di calibrazione | Valore di calibrazione | modelli | Livello |
| Türkçe | kısaltılmış ad | Kalibrasyon Kuralı | Kalibrasyon Değeri | model | Düzey |

| HS-CRP | |
|-----------|--|
| English | High Sensitivity C-reaction Protein |
| Русский | C-реактивный белок |
| Português | Proteína C Reativa de Alta Sensibilidade |
| Español | proteína C reactiva de alta sensibilidad |
| Italiano | proteina C reattiva ad alta sensibilità |
| Türkçe | Yüksek Duyarlıklılı C Reaksiyon Proteindir |