

Value sheet of Mindray BS Measurement System

Русский : Таблица результатов для системы BS компании Mindray
 Português : Planilha de valores do Sistema de Medição BS da Mindray
 Español : Hoja de valores del sistema de medición Mindray BS
 Italiano : Scheda dei valori del sistema di misurazione BS di Mindray
 Türkçe : Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский : Данные совпадают во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

7. **BS-380**: BS-380, BS-390;

2. **BS-240**;

8. **BS-400**: BS-400, BS-420;

3. **BS-240E**: BS240E, BS240Pro;

9. **BS-430**: BS-450, BS-460;

4. **BS-300**: BS-300, BS-320;

10. **BS-480**: BS-480, BS-490;

5. **BS-330E**: **BS-330E**(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-");

11. **BS-600**: BS-600, BS-620;

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00)(Serial Number starts with "W8-"
 and software version starts with "35.00"),
BS-350E(V35.00) (Serial Number starts with "W9-"
 and software version starts with "35.00");

12. **BS-600M**: BS-600M, BS-620M;

13. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,
 BS-1800, BS-1800plus;

14. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15. **BS-2800M**: BS-2600M, BS-2800M.

16. **S0**: 0.9% NaCl, Conc. Of **S0=0**;

Русский : **S0**: 0, 9% NaCl, конц. **S0=0**;

Português : **S0**: 0,9% NaCl, Conc. de **S0=0**;

Español : **S0**: 0,9% NaCl, Conc. de **S0=0**;

Italiano : **S0**: 0,9% NaCl, conc. di **S0=0**;

Türkçe : **S0**: %0,9 NaCl, **S0** Kons.=0.

LOT : 043423013

: 2025-03-21

Abbreviated name	HS-CRP	Calibration Rule	spline
Model	Level	Calibrator Value ¹⁶	
		mg/L	nmol/L
BS-200E ¹	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-240 ²	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-240E ³	a	6.95	66.2
	b	18.9	180
	c	40.0	381
	d	162	1542
	e	323	3075
BS-300 ⁴	a	5.11	48.6
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-330E ⁵	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075

HS-CRP Calibrator



Abbreviated name	HS-CRP	Calibration Rule	spline
Model	Level	Calibrator Value ¹⁶	
		mg/L	nmol/L
BS-360E ⁶	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-380 ⁷	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-400 ⁸	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-430 ⁹	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-480 ¹⁰	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-600 ¹¹	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075
BS-600M ¹²	a	4.77	45.4
	b	18.6	177
	c	39.0	371
	d	158	1504
	e	323	3075
BS-800 ¹³	a	4.83	46.0
	b	18.4	175
	c	38.7	368
	d	157	1495
	e	323	3075

HS-CRP Calibrator



Abbreviated name	HS-CRP	Calibration Rule	spline
Model	Level	Calibrator Value ¹⁶	
		mg/L	nmol/L
BS-2000 ¹⁴	a	5.01	47.7
	b	18.6	177
	c	39.6	377
	d	157	1495
	e	324	3084
BS-2800M ¹⁵	a	4.77	45.4
	b	18.6	177
	c	39.0	371
	d	158	1504
	e	323	3075

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

HS-CRP	
English	High Sensitivity C-reaction Protein
Русский	C-реактивный белок
Português	Proteína C Reativa de Alta Sensibilidade
Español	proteína C reactiva de alta sensibilidad
Italiano	proteina C reattiva ad alta sensibilità
Türkçe	Yüksek Duyarlılık C Reaksiyon Proteindir