

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;	9.BS-430: BS-430, BS-450,BS-460;
2.BS-240: BS-230, BS-240;	10.BS-480: BS-480, BS-490;
3.BS-240E: BS240E, BS240Pro;	11.BS-600: BS-600, BS-620;
4.BS-300: BS-300, BS-320;	12.BS-600M: BS-600M, BS-620M;
5.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")	13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");	14.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
7.BS-380: BS-380, BS-390;	15.BS-2800M:BS-2600M, BS-2800M;
8.BS-400: BS-400, BS-420;	16.Applicable models of the chemistry shall be subject to the parameter list and instructions.

LOT : 043423015

有效期 : 2025-05-06

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

HS-CRP	
English	High Sensitivity C-reaction Protein
Русский	C-реактивный белок
Português	Proteína C Reativa de Alta Sensibilidade
Español	proteína C reactiva de alta sensibilidad
Italiano	proteina C reattiva ad alta sensibilità
Türkçe	Yüksek Duyarlıklılı C Reaksiyon Proteindir

Abbreviated name		HS-CRP		Calibration Rule		spline			
Model		BS-200E ¹		BS-240 ²		BS-240E ³		BS-300 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
生理盐水		0.00	/	0.00	/	0.00	/	0.00	/
a	mg/L	4.83	0.3	4.83	0.3	6.95	0.5	5.11	0.4
b		18.4	1.3	18.4	1.3	18.9	1.3	18.4	1.3

HS-CRP Calibrator



c		38.7	2.6	38.7	2.6	40.0	2.6	38.7	2.6
d	mg/L	157	10	157	10	162	11	157	10
e		323	22	323	22	323	22	323	22
生理盐水		0.0	/	0.0	/	0.0	/	0.0	/
a		46.0	3	46.0	3	66.2	5	48.6	3
b	nmol/L	175	12	175	12	180	13	175	12
c		368	24	368	24	381	25	368	24
d		1495	100	1495	100	1542	103	1495	100
e		3075	205	3075	205	3075	205	3075	205
Model			BS-330E⁵		BS-360E⁶		BS-380⁷		BS-400⁸
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
生理盐水		0.00	/	0.00	/	0.00	/	0.00	/
a	mg/L	4.83	0.3	4.83	0.3	4.83	0.3	4.83	0.3
b		18.4	1.3	18.4	1.3	18.4	1.3	18.4	1.3
c		38.7	2.6	38.7	2.6	38.7	2.6	38.7	2.6
d		157	10	157	10	157	10	157	10
e		323	22	323	22	323	22	323	22
生理盐水		0.0	/	0.0	/	0.0	/	0.0	/
a	nmol/L	46.0	3	46.0	3	46.0	3	46.0	3
b		175	12	175	12	175	12	175	12
c		368	24	368	24	368	24	368	24
d		1495	100	1495	100	1495	100	1495	100
e		3075	205	3075	205	3075	205	3075	205
Model		BS-430⁹		BS-480¹⁰		BS-600¹¹		BS-600M¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
生理盐水		0.00	/	0.00	/	0.00	/	0.00	/
a	mg/L	4.83	0.3	4.83	0.3	4.83	0.3	4.77	0.3
b		18.4	1.3	18.4	1.3	18.4	1.3	18.6	1.3
c		38.7	2.6	38.7	2.6	38.7	2.6	39.0	2.6
d		157	10	157	10	157	10	158	11
e		323	22	323	22	323	22	323	22
生理盐水		0.0	/	0.0	/	0.0	/	0.0	/
a	nmol/L	46.0	3	46.0	3	46.0	3	45.4	3
b		175	12	175	12	175	12	177	12
c		368	24	368	24	368	24	371	25
d		1495	100	1495	100	1495	100	1504	101
e		3075	205	3075	205	3075	205	3075	205
Model		BS-800¹³		BS-2000¹⁴		BS-2800M¹⁵			
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty		
生理盐水		0.00	/	0.00	/	0.00	/		
a	mg/L	4.83	0.3	5.01	0.3	4.77	0.3		
b		18.4	1.3	18.6	1.3	18.6	1.3		
c		38.7	2.6	39.6	2.6	39.0	2.6		
d		157	10	157	11	158	11		
e		323	22	324	22	323	22		
生理盐水	nmol/L	0.0	/	0.0	/	0.0	/		
a		46.0	3	47.7	3	45.4	3		

HS-CRP Calibrator



b		175	12	177	12	177	12
c	nmol/L	368	24	377	25	371	24
d		1495	100	1495	100	1504	100
e		3075	205	3084	206	3075	205

mindray

HS-CRP Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 043423015

2025-05-06

