

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;

9.BS-430: BS-430, BS-450,BS-460;

2.BS-240: BS-230, BS-240;

10.BS-480: BS-480, BS-490;

3.BS-240E: BS240E, BS240Pro;

11.BS-600: BS-600, BS-620;

4.BS-300: BS-300, BS-320;

12.BS-600M: BS-600M, BS-620M;

5.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and

software version starts with "35.00"), BS-

14.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

15.BS-2800M:BS-2600M, BS-2800M;

7.BS-380: BS-380, BS-390;

16.Applicable models of the chemistry shall be subject to the parameter list and instructions.

8.BS-400: BS-400, BS-420;

LOT : 043424002

Exp : 2025-07-26

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

HS-CRP

English	High Sensitivity C-reaction Protein
Русский	C-реактивный белок
Português	Proteína C Reativa de Alta Sensibilidade
Español	proteína C reactiva de alta sensibilidad
Italiano	proteina C reattiva ad alta sensibilità
Türkçe	Yüksek Duyarlıklı C Reaksiyon Proteindir

Abbreviated name	HS-CRP	Calibration Rule		spline					
Model	BS-200E ¹	BS-240 ²		BS-240E ³		BS-300 ⁴			
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a	mg/L	4.85	0.3	4.85	0.3	5.85	0.4	4.85	0.3
b		17.8	1.3	17.8	1.3	17.8	1.3	17.8	1.3

HS-CRP Calibrator



c		38.3	2.5	38.3	2.5	38.3	2.5	38.3	2.5
d		151	10	151	10	151	10	151	10
e		328	22	328	22	328	22	328	22
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	nmol/L	46.2	3	46.2	3	55.7	4	46.2	3
b		169	12	169	12	169	12	169	12
c		365	24	365	24	365	24	365	24
d		1438	96	1438	96	1438	96	1438	96
e		3123	208	3123	208	3123	208	3123	208
Model		BS-330E⁵		BS-360E⁶		BS-380⁷		BS-400⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a	mg/L	4.85	0.3	4.85	0.3	4.85	0.3	4.85	0.3
b		17.8	1.3	17.8	1.3	17.8	1.3	17.8	1.3
c		38.3	2.5	38.3	2.5	38.3	2.5	38.3	2.5
d		151	10	151	10	151	10	151	10
e		328	22	328	22	328	22	328	22
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	nmol/L	46.2	3	46.2	3	46.2	3	46.2	3
b		169	12	169	12	169	12	169	12
c		365	24	365	24	365	24	365	24
d		1438	96	1438	96	1438	96	1438	96
e		3123	208	3123	208	3123	208	3123	208
Model		BS-430⁹		BS-480¹⁰		BS-600¹¹		BS-600M¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a	mg/L	4.85	0.3	4.85	0.3	4.85	0.3	4.74	0.3
b		17.8	1.3	17.8	1.3	17.8	1.3	18.3	1.3
c		38.3	2.5	38.3	2.5	38.3	2.5	38.8	2.6
d		151	10	151	10	151	10	155	10
e		328	22	328	22	328	22	328	22
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	nmol/L	46.2	3	46.2	3	46.2	3	45.1	3
b		169	12	169	12	169	12	174	12
c		365	24	365	24	365	24	369	24
d		1438	96	1438	96	1438	96	1476	99
e		3123	208	3123	208	3123	208	3123	208
Model		BS-800¹³		BS-2000¹⁴		BS-2800M¹⁵			
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty		
0.9%NaCl		0.00	/	0.00	/	0.00	/		
a	mg/L	4.85	0.3	4.78	0.3	4.74	0.3		
b		17.8	1.3	18.0	1.2	18.3	1.2		
c		38.3	2.5	39.3	2.6	38.8	2.6		
d		151	10	157	11	155	10		
e		328	22	325	22	328	22		
0.9%NaCl		0.0	/	0.0	/	0.0	/		
a		46.2	3	45.5	3	45.1	3		

HS-CRP Calibrator

mindray

b		169	12	171	12	174	12
c	nmol/L	365	24	374	25	369	24
d		1438	96	1495	100	1476	99
e		3123	208	3094	207	3123	208

mindray

HS-CRP Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 043424002

2025-07-26

