

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;	10.BS-480: BS-480, BS-490;
2.BS-240: BS-230, BS-240;	11.BS-600: BS-600, BS-620;
3.BS-240E: BS240E, BS240Pro;	12.BS-600M: BS-600M, BS-620M;
4.BS-300: BS-300, BS-320;	13.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
5.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")	14.BS-1000M: BS-1000M, BS-1100M;
6.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");	15.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
7.BS-380: BS-380, BS-390;	16.BS-2800M:BS-2600M, BS-2800M;
8.BS-400: BS-400, BS-420;	17.For applicable models of the analyte, please refer to the reagent parameter sheet and instruction.
9.BS-430: BS-430, BS-450,BS-460;	

LOT : 043424004

Exp. Date : 2025-09-22

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzye	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

HS-CRP	
English	High Sensitivity C-reaction Protein
Русский	C-реактивный белок
Português	Proteína C Reativa de Alta Sensibilidade
Español	proteína C reactiva de alta sensibilidad
Italiano	proteina C reattiva ad alta sensibilità
Türkçe	Yüksek Duyarlıklılı C Reaksiyon Proteindir

Abbreviated name	HS-CRP		Calibration Rule		spline				
Model	BS-200E ¹		BS-240 ²		BS-240E ³		BS-300 ⁴		
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.89	0.3	4.65	0.3	6.17	0.4	4.89	0.3

HS-CRP Calibrator



b		18.9	1.3	18.9	1.3	18.9	1.3	18.9	1.3
c	mg/L	39.6	2.6	39.6	2.6	38.8	2.6	39.6	2.6
d		157	10	157	10	155	10	157	10
e		321	21	321	21	321	21	321	21
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a		46.6	3	44.3	3	58.7	4	46.6	3
b	nmol/L	180	13	180	13	180	13	180	13
c		377	25	377	25	369	24	377	25
d		1495	100	1495	100	1476	99	1495	100
e		3056	204	3056	204	3056	204	3056	204
0.9%NaCl			0.0	/	0.0	/	0.0	/	0.0
Model		BS-330E⁵		BS-360E⁶		BS-380⁷		BS-400⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.89	0.3	4.89	0.3	4.72	0.3	4.72	0.3
b	mg/L	18.9	1.3	18.9	1.3	18.3	1.3	18.3	1.3
c		39.6	2.6	39.6	2.6	39.1	2.6	39.1	2.6
d		157	10	157	10	156	10	156	10
e		321	21	321	21	324	22	324	22
0.9%NaCl			0.0	/	0.0	/	0.0	/	0.0
a		46.6	3	46.6	3	44.9	3	44.9	3
b	nmol/L	180	13	180	13	174	12	174	12
c		377	25	377	25	372	25	372	25
d		1495	100	1495	100	1485	99	1485	99
e		3056	204	3056	204	3084	206	3084	206
0.9%NaCl			0.0	/	0.0	/	0.0	/	0.0
Model		BS-430⁹		BS-480¹⁰		BS-600¹¹		BS-600M¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.89	0.3	4.72	0.3	4.72	0.3	4.72	0.3
b	mg/L	18.9	1.3	18.3	1.3	18.3	1.3	18.3	1.3
c		39.6	2.6	39.1	2.6	39.1	2.6	39.1	2.6
d		157	10	156	10	156	10	156	10
e		321	21	324	22	324	22	324	22
0.9%NaCl			0.0	/	0.0	/	0.0	/	0.0
a		46.6	3	44.9	3	44.9	3	44.9	3
b	nmol/L	180	13	174	12	174	12	174	12
c		377	25	372	25	372	25	372	25
d		1495	100	1485	99	1485	99	1485	99
e		3056	204	3084	206	3084	206	3084	206
0.9%NaCl			0.0	/	0.0	/	0.0	/	0.0
Model		BS-800¹³		BS-1000M¹⁴		BS-2000¹⁵		BS-2800M¹⁶	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.00	/	0.00	/	0.00	/	0.00	/
a		4.89	0.3	4.72	0.3	4.95	0.3	4.72	0.3
b	mg/L	18.9	1.3	18.3	1.2	18.3	1.3	18.3	1.2
c		39.6	2.6	39.1	2.6	39.6	2.6	39.1	2.6
d		157	10	156	10	157	11	156	10
e		321	21	324	22	324	22	324	22
0.9%NaCl			0.0	/	0.0	/	0.0	/	0.0

HS-CRP Calibrator

mindray

a		46.6	3	44.9	3	47.1	3	44.9	3
b		180	13	174	12	174	12	174	12
c	nmol/L	377	25	372	24	377	25	372	24
d		1495	100	1485	99	1495	100	1485	99
e		3056	204	3084	206	3084	206	3084	206

mindray

HS-CRP Calibrator

For use on: BS-1000M, BS-2000,
BS-2800M, BS-600M,
BS-800

LOT 043424004

🕒 2025-09-22

