

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1. **BS-240**;
2. **BS-240E**: BS-240E, BS-240Pro;
3. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, **BS-330E(V35.00)**(Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)**(Serial Number starts with "W9-" and software version starts with "35.00");
4. **BS-380**: BS-380, BS-390;
5. **BS-400**: BS-400, BS-420;
6. **BS-430**: BS-450, BS-460;
7. **BS-480**: BS-480, BS-490;
8. **BS-600**: BS-600, BS-620;
9. **BS-600M**: BS-600M, BS-620M;
10. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
11. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;
12. **BS-2800M**: BS-2600M, BS-2800M.

13. a:0.9% NaCl is not needed, Conc. Of a=0;

Русский: a: 0,9% NaCl не необходим, конц. a=0;

Português: a: 0,9% O NaCl não é needed, Conc. de a=0;

Español: a: 0,9% El NaCl no es necesario, Conc. de a=0;

Italiano: a: 0.9% Il NaCl non è necessario, conc. di a=0;

Türkçe: a: %0,9 NaCl gerekli değildir, a Kons.=0.

**LOT**: 045723009

: 2024-04-30

Abbreviated name		D-Dimer		Calibration Rule			spline	
Model	Level	Calibration Value <sup>13</sup> (µg/mL)	Model	Level	Calibration Value <sup>13</sup> (µg/mL)	Model	Level	Calibration Value <sup>13</sup> (µg/mL)
BS-240 <sup>1</sup>	a	0	BS-240E <sup>2</sup>	a	0	BS-360E <sup>3</sup>	a	0
	b	2.36		b	2.36		b	2.36
	c	8.70		c	8.60		c	8.60
	d	17.4		d	17.0		d	17.0
	e	34.6		e	34.0		e	34.0
	f	50.4		f	50.4		f	50.4
BS-380 <sup>4</sup>	a	0	BS-400 <sup>5</sup>	a	0	BS-430 <sup>6</sup>	a	0
	b	2.36		b	2.36		b	2.36
	c	8.70		c	8.70		c	8.70
	d	17.2		d	17.2		d	17.3
	e	34.5		e	34.5		e	34.6
	f	50.4		f	50.4		f	50.8
BS-480 <sup>7</sup>	a	0	BS-600 <sup>8</sup>	a	0	BS-600M <sup>9</sup>	a	0
	b	2.36		b	2.36		b	2.37
	c	8.60		c	8.60		c	8.66
	d	17.0		d	17.0		d	17.0
	e	34.0		e	34.0		e	33.9
	f	50.4		f	50.4		f	50.1

# D-Dimer Calibrator

Abbreviated name			D-Dimer			Calibration Rule			spline		
Model	Level	Calibration Value <sup>13</sup> (µg/mL)	Model	Level	Calibration Value <sup>13</sup> (µg/mL)	Model	Level	Calibration Value <sup>13</sup> (µg/mL)			
BS-800 <sup>10</sup>	a	0	BS-2000 <sup>11</sup>	a	0	BS-2800M <sup>12</sup>	a	0			
	b	2.36		b	2.38		b	2.37			
	c	8.60		c	8.68		c	8.66			
	d	17.0		d	17.0		d	17.0			
	e	34.0		e	34.3		e	33.9			
	f	50.4		f	50.7		f	50.1			

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

D-Dimer	
English	D-Dimer
Русский	D-димера
Português	D-Dimer
Español	Dímero D
Italiano	D-Dimero
Türkçe	D-Dimer