

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-240: BS-230, BS-240;

8.BS-600: BS-600, BS-620;

2.BS-240E: BS240E, BS240Pro;

9.BS-600M: BS-600M, BS-620M;

3.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

10.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

4.BS-380: BS-380, BS-390;

11.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

5.BS-400: BS-400, BS-420;

12.BS-2800M:BS-2600M, BS-2800M;

6.BS-430: BS-430, BS-450,BS-460;

13.Applicable models of the chemistry shall be subject to the parameter list and instructions.

7.BS-480: BS-480, BS-490;

LOT : 045723015

: 2025-01-07

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

	D-Dimer
English	D-Dimer
Русский	D-димера
Português	D-Dimer
Español	Dímero D
Italiano	D-Dimero
Türkçe	D-Dimer

Abbreviated name	D-Dimer	Calibration Rule	spline						
Model	BS-240 ¹	BS-240E ²	BS-360E ³						
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty

a		0.00	/	0.00	/	0.00	/	0.00	/
b		1.93	0.07	1.93	0.07	1.93	0.07	1.93	0.07
c	µg/mL	8.40	0.17	8.40	0.17	8.40	0.17	8.40	0.17
d		16.3	0.4	16.3	0.4	16.3	0.4	16.3	0.4
e		32.4	1.1	32.4	1.1	32.4	1.1	32.4	1.1

D-Dimer Calibrator



f	µg/mL	46.3	1.6	46.3	1.6	46.3	1.6	46.3	1.6
Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
a		0.00	/	0.00	/	0.00	/	0.00	/
b		1.93	0.07	1.93	0.07	1.93	0.07	1.93	0.07
c	µg/mL	8.40	0.17	8.40	0.17	8.40	0.17	8.40	0.17
d		16.3	0.4	16.3	0.4	16.3	0.4	16.3	0.4
e		32.4	1.1	32.4	1.1	32.4	1.1	32.4	1.1
f		46.3	1.6	46.3	1.6	46.3	1.6	46.3	1.6
Model		BS-600M ⁹		BS-800 ¹⁰		BS-2000 ¹¹		BS-2800M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
a		0.00	/	0.00	/	0.00	/	0.00	/
b		1.98	0.08	1.93	0.07	2.02	0.08	1.98	0.08
c	µg/mL	8.35	0.17	8.40	0.17	8.40	0.17	8.35	0.17
d		16.1	0.4	16.3	0.4	15.9	0.4	16.1	0.4
e		31.0	1.0	32.4	1.1	31.6	1.0	31.0	1.0
f		46.0	1.6	46.3	1.6	46.0	1.6	46.0	1.6

mindray

D-Dimer Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 045723015

2025-01-07

