

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



We have updated the control value for new version MYO reagent. Please select the corresponding value and update.

The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-240: BS-230, BS-240;

8.BS-600: BS-600, BS-620;

2.BS-240E: BS240E, BS240Pro;

9.BS-600M: BS-600M, BS-620M;

3.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");

10.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

4.BS-380: BS-380, BS-390;

11.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

5.BS-400: BS-400, BS-420;

12.BS-2800M:BS-2600M, BS-2800M;

6.BS-430: BS-430, BS-450,BS-460;

13.Applicable models of the chemistry shall be subject to the parameter list and instructions.

7.BS-480: BS-480, BS-490;

LOT : 045824001

: 2024-12-18

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzyey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

	MYO
English	Myoglobin
Русский	Миоглобин
Português	Mioglobina
Español	Mioglobina
Italiano	Mioglobina
Türkçe	Myoglobin

★ The following reference values are applicable to:Reagent lot No.: 045823006 or Previous lot ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	ng/mL	88.0	1.2	88.0	1.2	88.0	1.2	88.0	1.2
b		181	4	181	4	181	4	181	4

MYO Calibrator *Please note the target value change



c	373	8	373	8	373	8	373	8
d	714	12	714	12	714	12	714	12

Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	ng/mL	88.0	1.2	88.0	1.2	88.0	1.2	88.0	1.2
b		181	4	181	4	181	4	181	4
c		373	8	373	8	373	8	373	8
d		714	12	714	12	714	12	714	12

Model		BS-600M ⁹		BS-800 ¹⁰		BS-2000 ¹¹		BS-2800M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0.0	/	0.0	/	0.0	/	0.0	/
a	ng/mL	92.0	1.3	88.0	1.2	93.0	1.3	92.0	1.3
b		183	4	181	4	183	4	183	4
c		373	8	373	8	375	8	373	8
d		718	13	714	12	716	12	718	12

★ The following reference values are applicable to: Reagent lot No.: 045823007 or subsequent lot ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	102	1.4	102	1.4	102	1.4	102	1.4
b		196	4	196	4	196	4	196	4
c		406	9	406	9	406	9	406	9
d		812	14	812	14	812	14	812	14

Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	102	1.4	102	1.4	102	1.4	102	1.4
b		196	4	196	4	196	4	196	4
c		406	9	406	9	406	9	406	9
d		812	14	812	14	812	14	812	14

Model		BS-600M ⁹		BS-800 ¹⁰		BS-2000 ¹¹		BS-2800M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	104	1.5	102	1.4	104	1.4	104	1.5
b		197	4	196	4	198	4	197	4
c		410	9	406	9	414	9	410	9
d		808	14	812	14	811	14	808	14

★ The following reference values are applicable to: Reagent lot No.: 045823007 or subsequent lot +Calibrator lot No.: 045822010、153722003 ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty

MYO Calibrator *Please note the target value change



0.9%NaCl		0	/	0	/	0	/	0	/
a		101	1.4	101	1.4	101	1.4	101	1.4
b	ng/mL	194	4	194	4	194	4	194	4
c		410	9	410	9	410	9	410	9
d		806	14	806	14	806	14	806	14

Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a		101	1.4	101	1.4	101	1.4	101	1.4
b	ng/mL	194	4	194	4	194	4	194	4
c		410	9	410	9	410	9	410	9
d		806	14	806	14	806	14	806	14

Model		BS-600M ⁹		BS-800 ¹⁰		BS-2000 ¹¹		BS-2800M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a		103	1.5	101	1.4	104	1.4	103	1.5
b	ng/mL	192	4	194	4	192	4	192	4
c		413	9	410	9	412	9	413	9
d		790	14	806	14	803	14	790	14

★ The following reference values are applicable to: Reagent lot No.: 045823007 or subsequent lot +Calibrator lot No.: 045822015, 045822017, 045823001 ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a		100	1.4	100	1.4	100	1.4	100	1.4
b	ng/mL	195	4	195	4	195	4	195	4
c		425	10	425	10	425	10	425	10
d		825	14	825	14	825	14	825	14

Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a		100	1.4	100	1.4	100	1.4	100	1.4
b	ng/mL	195	4	195	4	195	4	195	4
c		425	10	425	10	425	10	425	10
d		825	14	825	14	825	14	825	14

Model		BS-600M ⁹		BS-800 ¹⁰		BS-2000 ¹¹		BS-2800M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a		102	1.4	100	1.4	100	1.4	102	1.5
b	ng/mL	195	4	195	4	194	4	195	4
c		428	10	425	10	423	10	428	10
d		805	14	825	14	810	14	805	14

★ The following reference values are applicable to: Reagent lot No.: 045823007 or subsequent lot +Calibrator lot No.: 153723001 ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	101	1.4	101	1.4	101	1.4	101	1.4
b		199	4	199	4	199	4	199	4
c		423	10	423	10	423	10	423	10
d		820	14	820	14	820	14	820	14

Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	101	1.4	101	1.4	101	1.4	101	1.4
b		199	4	199	4	199	4	199	4
c		423	10	423	10	423	10	423	10
d		820	14	820	14	820	14	820	14

Model		BS-600M ⁹		BS-800 ¹⁰		BS-2000 ¹¹		BS-2800M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	102	1.4	101	1.4	103	1.4	102	1.5
b		199	4	199	4	202	4	199	4
c		423	10	423	10	418	9	423	10
d		800	14	820	14	815	14	800	14

★ The following reference values are applicable to: Reagent lot No.: 045823007 or subsequent lot +Calibrator lot No.: 045823005 ★

Abbreviated name		MYO		Calibration Rule		spline			
Model		BS-240 ¹		BS-240E ²		BS-360E ³		BS-380 ⁴	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	105	1.5	105	1.5	105	1.5	105	1.5
b		196	4	196	4	196	4	196	4
c		406	9	406	9	406	9	406	9
d		797	14	797	14	797	14	797	14

Model		BS-400 ⁵		BS-430 ⁶		BS-480 ⁷		BS-600 ⁸	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	105	1.5	105	1.5	105	1.5	105	1.5
b		196	4	196	4	196	4	196	4
c		406	9	406	9	406	9	406	9
d		797	14	797	14	797	14	797	14

Model		BS-600M ⁹		BS-800 ¹⁰		BS-2000 ¹¹		BS-2800M ¹²	
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a	ng/mL	107	1.5	105	1.5	105	1.4	107	1.5
b		197	4	196	4	197	4	197	4

MYO Calibrator *Please note the target value change



c	407	9	406	9	408	9	407	9
d	785	14	797	14	794	14	785	14

mindray

MYO Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 045824001

EXP 2024-12-18

