

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

2.**BS-200**: BS-200, BS-220;

3.**BS-200E**: BS-200E, BS-220E;

4.**BS-240**: BS-230, BS-240;

5.**BS-240E**: BS-240E, BS-240Pro;

6.**BS-360E**: BS-360E,BS-370E,BS-350S,BS-360S,

**BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");**

7.**BS-380**: BS-380, BS-390;

8.**BS-400**: BS-400, BS-420;

9.**BS-430**: BS-430, BS-450, BS-460;

10.**BS-480**: BS-480, BS-490;

11.**BS-600**: BS-600, BS-620;

12.**BS-600M**: BS-600M, BS-620M;

13.**BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-880, BS-890, BS-1800, BS-1800plus;

14.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15.**BS-2800M**: BS-2600M, BS-2800M.

16. Cal0: 0.9% NaCl, Conc. Of Cal0=0;

Português: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Italiano: Cal0: 0,9% NaCl, conc. di Cal0=0;

Русский: Cal0: 0,9% NaCl, конц. Cal0=0;

Español: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Türkçe: Cal0: %0,9 NaCl, Cal0 Kons.=0.

**LOT** : 046022007

**EXP** : 2024-02-17

Abbreviated name		MALB				Calibration Rule				spline			
Model	Level	Calibration Value <sup>16</sup>				Model	Level	Calibration Value <sup>16</sup>					
		Urine		CSF				Urine		CSF			
		mg/L	µmol/L	mg/L	µmol/L			mg/L	µmol/L	mg/L	µmol/L		
BS-120 <sup>1</sup>	a	19.2	0.292	/	/	BS-430 <sup>9</sup>	a	19.2	0.292	/	/		
	b	53.2	0.809	/	/		b	53.2	0.809	/	/		
	c	101	1.54	/	/		c	101	1.54	/	/		
	d	197	2.99	/	/		d	197	2.99	/	/		
	e	374	5.68	/	/		e	374	5.68	/	/		
BS-200 <sup>2</sup>	a	19.2	0.292	/	/	BS-480 <sup>10</sup>	a	19.2	0.292	/	/		
	b	53.2	0.809	/	/		b	53.2	0.809	/	/		
	c	101	1.54	/	/		c	101	1.54	/	/		
	d	197	2.99	/	/		d	197	2.99	/	/		
	e	374	5.68	/	/		e	374	5.68	/	/		
BS-200E <sup>3</sup>	a	19.2	0.292	/	/	BS-600 <sup>11</sup>	a	19.2	0.292	/	/		
	b	53.2	0.809	/	/		b	53.2	0.809	/	/		
	c	101	1.54	/	/		c	101	1.54	/	/		
	d	197	2.99	/	/		d	197	2.99	/	/		
	e	374	5.68	/	/		e	374	5.68	/	/		

Abbreviated name		MALB				Calibration Rule		spline			
Model	Level	Calibration Value <sup>16</sup>				Model	Level	Calibration Value <sup>16</sup>			
		Urine		CSF				Urine		CSF	
		mg/L	µmol/L	mg/L	µmol/L			mg/L	µmol/L	mg/L	µmol/L
BS-240 <sup>4</sup>	a	19.2	0.292	/	/	BS-600M <sup>12</sup>	a	19.2	0.292	/	/
	b	53.2	0.809	/	/		b	53.2	0.809	/	/
	c	101	1.54	/	/		c	101	1.54	/	/
	d	197	2.99	/	/		d	197	2.99	/	/
	e	374	5.68	/	/		e	374	5.68	/	/
BS-240E <sup>5</sup>	a	19.2	0.292	/	/	BS-800 <sup>13</sup>	a	19.2	0.292	/	/
	b	53.2	0.809	/	/		b	53.2	0.809	/	/
	c	101	1.54	/	/		c	101	1.54	/	/
	d	197	2.99	/	/		d	197	2.99	/	/
	e	374	5.68	/	/		e	374	5.68	/	/
BS-360E <sup>6</sup>	a	19.2	0.292	/	/	BS-2000 <sup>14</sup>	a	20.4	0.310	/	/
	b	53.2	0.809	/	/		b	55.3	0.841	/	/
	c	101	1.54	/	/		c	103	1.57	/	/
	d	197	2.99	/	/		d	199	3.02	/	/
	e	374	5.68	/	/		e	369	5.61	/	/
BS-380 <sup>7</sup>	a	19.2	0.292	/	/	BS-2800M <sup>15</sup>	a	20.3	0.309	/	/
	b	53.2	0.809	/	/		b	54.7	0.831	/	/
	c	101	1.54	/	/		c	102	1.55	/	/
	d	197	2.99	/	/		d	198	3.01	/	/
	e	374	5.68	/	/		e	363	5.52	/	/
BS-400 <sup>8</sup>	a	19.2	0.292	/	/						
	b	53.2	0.809	/	/						
	c	101	1.54	/	/						
	d	197	2.99	/	/						
	e	374	5.68	/	/						

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

MALB	
English	Microalbumin
Русский	микроальбумина
Português	Microalbuminúria
Español	Microalbúmina
Italiano	Microalbumina
Türkçe	Mikroalbumin