

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

1. **BS-120**: BS-120, BS-130;

2. **BS-200**: BS-200, BS-220;

3. **BS-200E**: BS-200E, BS-220E;

4. **BS-230**: BS-230, BS-240;

5. **BS-240E**: BS-240E, BS-240Pro;

6. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

**BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");**

16. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Português: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Italiano: Cal0: 0,9% NaCl, conc. di Cal0=0;

7. **BS-380**: BS-380, BS-390;

8. **BS-400**: BS-400, BS-420;

9. **BS-430**: BS-430, BS-450, BS-460;

10. **BS-480**: BS-480, BS-490;

11. **BS-600**: BS-600, BS-620;

12. **BS-600M**: BS-600M, BS-620M;

13. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

14. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

15. **BS-2800M**: BS-2600M, BS-2800M.

Русский: Cal0: 0,9% NaCl, конц. Cal0=0;

Español: Cal0: 0,9% NaCl, Conc. de Cal0=0;

Türkçe: Cal0: %0,9 NaCl, Cal0 Kons.=0.

**LOT** : 046023009

**EXP** : 2025-05-12



The following reference values are only applicable to:  
Reagent lot No.: 046023007, 046023008 and 046023009



Abbreviated name		MALB		Calibration Rule		spline	
Model	Level	Calibration Value <sup>16</sup>		Model	Level	Calibration Value <sup>16</sup>	
		mg/L	µmol/L			mg/L	µmol/L
BS-120 <sup>1</sup>	a	18.6	0.283	BS-430 <sup>9</sup>	a	18.6	0.283
	b	52.1	0.792		b	52.1	0.792
	c	98.3	1.49		c	98.3	1.49
	d	198	3.01		d	198	3.01
	e	380	5.78		e	380	5.78
BS-200 <sup>2</sup>	a	18.6	0.283	BS-480 <sup>10</sup>	a	18.6	0.283
	b	52.1	0.792		b	52.1	0.792
	c	98.3	1.49		c	98.3	1.49
	d	198	3.01		d	198	3.01
	e	380	5.78		e	380	5.78
BS-200E <sup>3</sup>	a	18.6	0.283	BS-600 <sup>11</sup>	a	18.6	0.283
	b	52.1	0.792		b	52.1	0.792
	c	98.3	1.49		c	98.3	1.49
	d	198	3.01		d	198	3.01
	e	380	5.78		e	380	5.78

# MALB Calibrator <sup>\*Please note the calibration value change</sup>



Abbreviated name		MALB		Calibration Rule		spline	
Model	Level	Calibration Value <sup>16</sup>		Model	Level	Calibration Value <sup>16</sup>	
		mg/L	μmol/L			mg/L	μmol/L
BS-230 <sup>4</sup>	a	18.6	0.283	BS-600M <sup>12</sup>	a	18.6	0.283
	b	52.1	0.792		b	52.1	0.792
	c	98.3	1.49		c	98.3	1.49
	d	198	3.01		d	198	3.01
	e	380	5.78		e	380	5.78
BS-240E <sup>5</sup>	a	18.6	0.283	BS-800 <sup>13</sup>	a	18.6	0.283
	b	52.1	0.792		b	52.1	0.792
	c	98.3	1.49		c	98.3	1.49
	d	198	3.01		d	198	3.01
	e	380	5.78		e	380	5.78
BS-360E <sup>6</sup>	a	18.6	0.283	BS-2000 <sup>14</sup>	a	19.1	0.290
	b	52.1	0.792		b	52.7	0.801
	c	98.3	1.49		c	99.2	1.51
	d	198	3.01		d	199	3.02
	e	380	5.78		e	373	5.67
BS-380 <sup>7</sup>	a	18.6	0.283	BS-2800M <sup>15</sup>	a	18.4	0.280
	b	52.1	0.792		b	52.6	0.800
	c	98.3	1.49		c	98.8	1.50
	d	198	3.01		d	199	3.02
	e	380	5.78		e	384	5.84
BS-400 <sup>8</sup>	a	18.6	0.283				
	b	52.1	0.792				
	c	98.3	1.49				
	d	198	3.01				
	e	380	5.78				

# MALB Calibrator

\*Please note the calibration value change



The following reference values are applicable to:

Reagent lot No.: Lot reagent except 046023007, 046023008 and 046023009



Abbreviated name		MALB		Calibration Rule		spline	
Model	Level	Calibration Value <sup>16</sup>		Model	Level	Calibration Value <sup>16</sup>	
		mg/L	μmol/L			mg/L	μmol/L
BS-120 <sup>1</sup>	a	21.4	0.325	BS-430 <sup>9</sup>	a	21.4	0.325
	b	56.0	0.851		b	56.0	0.851
	c	104	1.58		c	104	1.58
	d	202	3.07		d	202	3.07
	e	381	5.79		e	381	5.79
BS-200 <sup>2</sup>	a	21.4	0.325	BS-480 <sup>10</sup>	a	21.4	0.325
	b	56.0	0.851		b	56.0	0.851
	c	104	1.58		c	104	1.58
	d	202	3.07		d	202	3.07
	e	381	5.79		e	381	5.79
BS-200E <sup>3</sup>	a	21.4	0.325	BS-600 <sup>11</sup>	a	21.4	0.325
	b	56.0	0.851		b	56.0	0.851
	c	104	1.58		c	104	1.58
	d	202	3.07		d	202	3.07
	e	381	5.79		e	381	5.79
BS-230 <sup>4</sup>	a	21.4	0.325	BS-600M <sup>12</sup>	a	21.4	0.325
	b	56.0	0.851		b	56.0	0.851
	c	104	1.58		c	104	1.58
	d	202	3.07		d	202	3.07
	e	381	5.79		e	381	5.79
BS-240E <sup>5</sup>	a	21.4	0.325	BS-800 <sup>13</sup>	a	21.4	0.325
	b	56.0	0.851		b	56.0	0.851
	c	104	1.58		c	104	1.58
	d	202	3.07		d	202	3.07
	e	381	5.79		e	381	5.79
BS-360E <sup>6</sup>	a	21.4	0.325	BS-2000 <sup>14</sup>	a	22.1	0.336
	b	56.0	0.851		b	56.9	0.865
	c	104	1.58		c	105	1.60
	d	202	3.07		d	205	3.12
	e	381	5.79		e	373	5.67
BS-380 <sup>7</sup>	a	21.4	0.325	BS-2800M <sup>15</sup>	a	21.2	0.322
	b	56.0	0.851		b	56.4	0.857
	c	104	1.58		c	104	1.58
	d	202	3.07		d	204	3.10
	e	381	5.79		e	383	5.82
BS-400 <sup>8</sup>	a	21.4	0.325				
	b	56.0	0.851				
	c	104	1.58				
	d	202	3.07				
	e	381	5.79				

# MALB Calibrator

\*Please note the calibration value change

**mindray**

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
<b>Русский</b>	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
<b>Português</b>	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
<b>Español</b>	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
<b>Italiano</b>	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
<b>Türkçe</b>	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

<b>MALB</b>	
<b>English</b>	Microalbumin
<b>Русский</b>	микроальбумина
<b>Português</b>	Microalbuminúria
<b>Español</b>	Microalbúmina
<b>Italiano</b>	Microalbumina
<b>Türkçe</b>	Mikroalbumin