

## Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português : A dados de cada grupo é a mesma.

Español : la datos de cada grupo es la misma.

Italiano : la dati di ogni gruppo è la stessa.

Türkçe : her grubun veri aynıdır.

1.BS-200E: BS-200E, BS-220E,;	8.BS-480: BS-480, BS-490;
2.BS-240: BS-230, BS-240;	9.BS-600: BS-600, BS-620;
3.BS-240E: BS240E, BS240Pro;	10.BS-600M: BS-600M, BS-620M;
4.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with"W9-" and software version starts with "35.00");	11.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
5.BS-380: BS-380, BS-390;	12.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
6.BS-400: BS-400, BS-420;	13.BS-2800M:BS-2600M, BS-2800M;
7.BS-430: BS-430, BS-450,BS-460;	14.Applicable models of the chemistry shall be subject to the parameter list and instructions.

LOT : 046124001

: 2025-04-09

English	Abbreviated name	Calibration Value	Level	Model
Русский	сокращенное наименование	Эталонное значение	Уровень	модель
Português	Nome abreviado	Valor de calibração	Nível	Modelo
Español	nombre abreviado	Valor de calibración	Nivel	modelo
Italiano	abbreviazione	Valore di calibrazione	Livello	modelli
Türkçe	kısaltılmış ad	Kalibrasyon Değeri	Düzey	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

English	FER
English	Ferritin
Русский	ферритин
Português	Ferritina
Español	Ferritina
Italiano	Ferritina
Türkçe	Ferritin

Abbreviated name	FER	Calibration Rule		spline					
Model	BS-200E <sup>1</sup>	BS-240 <sup>2</sup>		BS-240E <sup>3</sup>		BS-360E <sup>4</sup>			
Level	Unit	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty	Calibrator Value	Expanded Uncertainty
0.9%NaCl		0	/	0	/	0	/	0	/
a		129	9	129	9	119	9	117	8
b	ng/mL	260	6	260	6	238	5	230	5
c		560	24	560	24	569	25	563	24
d		1130	40	1130	40	1070	38	1070	38

# FER Calibrator



<b>0.9%NaCl</b>		0	/	0	/	0	/	0	/
<b>a</b>		290	21	290	21	267	19	263	19
<b>b</b>	pmol/L	584	13	584	13	535	12	517	11
<b>c</b>		1258	54	1258	54	1279	55	1265	55
<b>d</b>		2539	89	2539	89	2404	84	2404	84
<b>Model</b>		<b>BS-380<sup>5</sup></b>		<b>BS-400<sup>6</sup></b>		<b>BS-430<sup>7</sup></b>		<b>BS-480<sup>8</sup></b>	
<b>Level</b>	<b>Unit</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>
<b>0.9%NaCl</b>		0	/	0	/	0	/	0	/
<b>a</b>		121	9	121	9	117	8	117	8
<b>b</b>	ng/mL	232	5	232	5	242	5	230	5
<b>c</b>		565	24	565	24	568	25	563	24
<b>d</b>		1070	38	1070	38	1070	38	1070	38
<b>0.9%NaCl</b>		0	/	0	/	0	/	0	/
<b>a</b>		272	19	272	19	263	19	263	19
<b>b</b>	pmol/L	521	12	521	12	544	12	517	11
<b>c</b>		1270	55	1270	55	1276	55	1265	55
<b>d</b>		2404	84	2404	84	2404	84	2404	84
<b>Model</b>		<b>BS-600<sup>9</sup></b>		<b>BS-600M<sup>10</sup></b>		<b>BS-800<sup>11</sup></b>		<b>BS-2000<sup>12</sup></b>	
<b>Level</b>	<b>Unit</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>
<b>0.9%NaCl</b>		0	/	0	/	0	/	0	/
<b>a</b>		117	8	115	8	117	8	116	8
<b>b</b>	ng/mL	230	5	227	5	230	5	229	5
<b>c</b>		563	24	557	24	563	24	545	24
<b>d</b>		1070	38	1093	38	1070	38	1066	37
<b>0.9%NaCl</b>		0	/	0	/	0	/	0	/
<b>a</b>		263	19	258	18	263	19	261	19
<b>b</b>	pmol/L	517	11	510	11	517	11	515	11
<b>c</b>		1265	55	1252	54	1265	55	1225	53
<b>d</b>		2404	84	2456	86	2404	84	2395	83
<b>Model</b>		<b>BS-2800M<sup>13</sup></b>							
<b>Level</b>	<b>Unit</b>	<b>Calibrator Value</b>	<b>Expanded Uncertainty</b>						
<b>0.9%NaCl</b>		0	/						
<b>a</b>		115	8						
<b>b</b>	ng/mL	227	5						
<b>c</b>		557	24						
<b>d</b>		1093	38						
<b>0.9%NaCl</b>		0	/						
<b>a</b>		258	18						
<b>b</b>	pmol/L	510	11						
<b>c</b>		1252	54						
<b>d</b>		2456	84						

# mindray

## FER Calibrator

For use on: BS-2000, BS-2800M,  
BS-600M, BS-800

LOT 046124001

2025-04-09

