

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

9.**BS-430**: BS-430, BS-450, BS-460;

2.**BS-200**: BS-200, BS-220;

10.**BS-480**: BS-480, BS-490;

3.**BS-200E**: BS-200E, BS-220E;

11.**BS-600**: BS-600, BS-620;

4.**BS-230**: BS-230, BS-240;

12.**BS-600M**: BS-600M, BS-620M;

5.**BS-240E**: BS-240E, BS-240Pro;

13.**BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.**BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,

14.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)**

15.**BS-2800M**: BS-2600M, BS-2800M.

(Serial Number starts with "W9-" and software version starts with "35.00");

16. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Português: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Italiano: **Cal0: 0,9% NaCl, conc. di Cal0=0;**

Русский: **Cal0: 0, 9% NaCl, конц. Cal0=0;**

Español: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Türkçe: **Cal0: %0,9 NaCl, Cal0 Kons.=0.**

7.**BS-380**: BS-380, BS-390;

8.**BS-400**: BS-400, BS-420;

LOT : 046223005

: 2025-06-05

Abbreviated name		TRF		Calibration Rule		spline	
Model	Level	Calibration Value ¹⁶		Model	Level	Calibration Value ¹⁶	
		g/L	µmol/L			g/L	µmol/L
BS-120 ¹	a	0.760	9.58	BS-430 ⁹	a	0.760	9.58
	b	1.76	22.2		b	1.76	22.2
	c	2.46	31.0		c	2.46	31.0
	d	3.66	46.1		d	3.66	46.1
	e	4.63	58.3		e	4.63	58.3
BS-200 ²	a	0.760	9.58	BS-480 ¹⁰	a	0.760	9.58
	b	1.76	22.2		b	1.76	22.2
	c	2.46	31.0		c	2.46	31.0
	d	3.66	46.1		d	3.66	46.1
	e	4.63	58.3		e	4.63	58.3
BS-200E ³	a	0.760	9.58	BS-600 ¹¹	a	0.760	9.58
	b	1.76	22.2		b	1.76	22.2
	c	2.46	31.0		c	2.46	31.0
	d	3.66	46.1		d	3.66	46.1
	e	4.63	58.3		e	4.63	58.3

Abbreviated name		TRF		Calibration Rule		spline	
Model	Level	Calibration Value ¹⁶		Model	Level	Calibration Value ¹⁶	
		g/L	μmol/L			g/L	μmol/L
BS-230 ⁴	a	0.760	9.58	BS-600M ¹²	a	0.750	9.45
	b	1.76	22.2		b	1.75	22.1
	c	2.46	31.0		c	2.46	31.0
	d	3.66	46.1		d	3.70	46.6
	e	4.63	58.3		e	4.72	59.5
BS-240E ⁵	a	0.760	9.58	BS-800 ¹³	a	0.760	9.58
	b	1.76	22.2		b	1.76	22.2
	c	2.46	31.0		c	2.46	31.0
	d	3.66	46.1		d	3.66	46.1
	e	4.63	58.3		e	4.63	58.3
BS-360E ⁶	a	0.760	9.58	BS-2000 ¹⁴	a	0.790	10.0
	b	1.76	22.2		b	1.77	22.3
	c	2.46	31.0		c	2.48	31.2
	d	3.66	46.1		d	3.72	46.9
	e	4.63	58.3		e	4.67	58.8
BS-380 ⁷	a	0.760	9.58	BS-2800M ¹⁵	a	0.750	9.45
	b	1.76	22.2		b	1.75	22.1
	c	2.46	31.0		c	2.46	31.0
	d	3.66	46.1		d	3.70	46.6
	e	4.63	58.3		e	4.72	59.5
BS-400 ⁸	a	0.760	9.58				
	b	1.76	22.2				
	c	2.46	31.0				
	d	3.66	46.1				
	e	4.63	58.3				

English	Abbreviated name	Calibration Rule	Calibration Value	Model	Level
Русский	сокращенное наименование	Принцип калибровки	Эталонное значение	модель	Уровень
Português	Nome abreviado	Regra de calibração	Valor de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	Valor de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	Valore di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	Kalibrasyon Değeri	model	Düzey

TRF	
English	Transferrin
Русский	трансферрина
Português	Transferrina
Español	Transferrina
Italiano	Transferrina
Türkçe	Transferin