

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов измерительной системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадает во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

1. **BS-200E**: BS-200E, BS-220E;

2. **BS-230**: BS-230,BS-240;

3. **BS-240E**: BS-240E, BS-240Pro;

4. **BS-360E**: BS-360E, BS-370E,BS-350S,BS-360S,

BS-330E(V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), **BS-350E(V35.00)** (Serial Number starts with "W9-" and software version starts with "35.00");

5. **BS-380**: BS-380, BS-390;

14. **Cal0: 0.9% NaCl, Conc. Of Cal0=0;**

Português: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Italiano: **Cal0: 0,9% NaCl, conc. di Cal0=0;**

6. **BS-400**: BS-400,BS-420;

7. **BS-430**: BS-430,BS-450,BS-460;

8. **BS-480**: BS-480,BS-490;

9. **BS-600**: BS-600, BS-620;

10. **BS-600M**: BS-600M, BS-620M;

11. **BS-800**: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

12. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

13. **BS-2800M**: BS-2600M, BS-2800M.

Русский: **Cal0: 0,9% NaCl, конц. Cal0=0;**

Español: **Cal0: 0,9% NaCl, Conc. de Cal0=0;**

Türkçe: **Cal0: %0,9 NaCl, Cal0 Kons.=0**

LOT : 046323008

EXP : 2024-07-05

Abbreviated name		RBP	Calibration Rule		Spline
Model	Level	Calibration Value ¹⁴ (mg/dL)	Sample Vol for Dilution / μ L	Diluent Vol/ μ L	Sample Vol For Analysis / μ L
BS-200E ¹ R1: R2: S= 180: 60: 4	Cal1	2.74	45	135	4
	Cal2	5.58	45	135	8
	Cal3	12.5	/	/	4
	Cal4	17.2	/	/	6
BS-380 ⁵ R1: R2: S= 180: 60: 4	Cal1	2.68	40	120	4
	Cal2	5.23	40	120	8
	Cal3	12.2	/	/	4
	Cal4	16.8	/	/	6
BS-400 ⁶ R1: R2: S= 180: 60: 4	Cal1	2.68	40	120	4
	Cal2	5.38	40	120	8
	Cal3	12.4	/	/	4
	Cal4	17.1	/	/	6
Abbreviated name		RBP	Calibration Rule		Spline
Model	Level	Calibration Value ¹⁴ (mg/dL)	Sample Vol For Analysis / μ L	Sample Vol for Dilution / μ L	Diluent Vol/ μ L
BS-230 ² R1: R2: S= 180: 60: 4	Cal1	2.74	4	35	105
	Cal2	5.58	8	35	105
	Cal3	12.5	4	/	/
	Cal4	17.2	6	/	/

Abbreviated name		RBP	Calibration Rule		Spline
Model	Level	Calibration Value ¹⁴ (mg/dL)	Sample Vol For Analysis / μ L	Sample Vol for Dilution / μ L	Diluent Vol/ μ L
BS-240E ³ R1: R2: S= 180: 60: 4	Cal1	2.68	4	34	102
	Cal2	5.23	8	34	102
	Cal3	12.2	4	/	/
	Cal4	16.8	6	/	/
BS-360E ⁴ R1: R2: S= 180: 60: 4	Cal1	2.68	4	34	102
	Cal2	5.23	8	34	102
	Cal3	12.2	4	/	/
	Cal4	16.8	6	/	/
BS-430 ⁷ R1: R2: S= 180: 60: 4	Cal1	2.68	4	30	90
	Cal2	5.23	8	30	90
	Cal3	12.2	4	/	/
	Cal4	16.8	6	/	/
BS-480 ⁸ R1: R2: S= 180: 60: 4	Cal1	2.78	4	40	120
	Cal2	5.44	8	40	120
	Cal3	12.6	4	/	/
	Cal4	16.8	6	/	/
BS-600 ⁹ R1: R2: S= 180: 60: 4	Cal1	2.68	4	30	90
	Cal2	5.23	8	30	90
	Cal3	12.2	4	/	/
	Cal4	16.8	6	/	/
BS-600M ¹⁰ R1: R2: S= 180: 60: 4	Cal1	2.81	4	25	75
	Cal2	5.75	8	25	75
	Cal3	12.4	4	/	/
	Cal4	17.3	6	/	/
BS-800 ¹¹ R1: R2: S= 180: 60: 4	Cal1	2.68	4	30	90
	Cal2	5.23	8	30	90
	Cal3	12.2	4	/	/
	Cal4	16.8	6	/	/
BS-2000 ¹² R1: R2: S= 180: 60: 4	Cal1	2.81	4	25	75
	Cal2	5.60	8	25	75
	Cal3	12.3	4	/	/
	Cal4	17.1	6	/	/
BS-2800M ¹³ R1: R2: S= 180: 60: 4	Cal1	2.81	4	25	75
	Cal2	5.75	8	25	75
	Cal3	12.4	4	/	/
	Cal4	17.3	6	/	/

English	Abbreviated name	Calibration Rule	Model	Level
Русский	сокращенное наименование	Принцип калибровки	модель	Уровень
Português	Nome abreviado	Regra de calibração	Modelo	Nível
Español	nombre abreviado	Regla de calibración	modelo	Nivel
Italiano	abbreviazione	Regola di calibrazione	modelli	Livello
Türkçe	kısaltılmış ad	Kalibrasyon Kuralı	model	Düzey

RBP Calibrator



English	Calibration Value	Sample Vol. for Dilution	Diluent Vol.	Sample Vol. For Analysis
Русский	Эталонное значение	Объем пробы для разбавления	Объем разбавителя	Объем пробы для анализа
Português	Valor de calibração	Volume da amostra para diluição	Volume de diluente	Volume da amostra para análise
Español	Valor de calibración	Vol. muestra para dilución	Vol. diluyente	Vol. muestra para análisis
Italiano	Valore di calibrazione	Vol. campione per la diluizione	Vol. diluente	Vol. campione per analisi
Türkçe	Kalibrasyon Değeri	Dilüsyon için Numune Hacmi	Seyreltici Hacmi	Analiz için Numune Hacmi

RBP	
English	Retinol Binding Protein
Русский	Ретинол-связывающий белок
Português	Proteína transportadora do retinol
Español	Proteína fijadora de retinol
Italiano	Proteina legante il retinolo
Türkçe	Retinol Bağlayıcı Protein