

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Español: la datos de cada grupo es la misma.

Türkçe: her grubun veri aynıdır.

Português: A dados de cada grupo é a mesma.

Italiano: la dati di ogni gruppo è la stessa.

1. **BS-120**: BS-120, BS-130;

2. **BS-180**: BS-180, BS-190;

3. **BS-200**: BS-200, BS-220;

4. **BS-200E**: BS-200E, BS-220E;

5. **BS-230**: BS-230, BS-240;

6. **BS-240E**: BS-240E, BS-240Pro;

7. **BS-300**: BS-300, BS-320;

8. **BS-330**: BS-330, BS-350;

9. **BS-330E**: BS-330E (Serial Number starts with "XQ-"),

BS-350E (Serial Number starts with "XS-");

20. **S0: 0.9% NaCl, Conc. Of S0=0;**

Português: S0: 0,9% NaCl, Conc. de S0=0;

Italiano: S0: 0,9% NaCl, conc. di S0=0;

10. **BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E (V35.00) (Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E (V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

11. **BS-380**: BS-380, BS-390;

12. **BS-400**: BS-400, BS-420;

13. **BS-430**: BS-430, BS-450, BS-460;

14. **BS-480**: BS-480, BS-490;

15. **BS-600**: BS-600, BS-620;

16. **BS-600M**: BS-600M, BS-620M;

17. **BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

18. **BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M;

19. **BS-2800M**: BS-2600M, BS-2800M.

Русский: S0: 0, 9% NaCl, конц. S0=0;

Español: S0: 0,9% NaCl, Conc. de S0=0;

Türkçe: S0: %0,9 NaCl, S0 Kons.=0.

LOT : 143222007

⌚ : 2023-07-28

Abbreviated name	Unit	Model	Calibration Value ²⁰	Model	Calibration Value ²⁰
TBA	µmol/L	BS-120 ¹	32.4	BS-380 ¹¹	31.8
		BS-180 ²	32.4	BS-400 ¹²	31.8
		BS-200 ³	31.8	BS-430 ¹³	32.1
		BS-200E ⁴	33.4	BS-480 ¹⁴	32.4
		BS-230 ⁵	32.6	BS-600 ¹⁵	31.8
		BS-240E ⁶	31.8	BS-600M ¹⁶	32.1
		BS-300 ⁷	31.8	BS-800 ¹⁷	31.8
		BS-330 ⁸	31.8	BS-2000 ¹⁸	33.0
		BS-330E ⁹	33.4	BS-2800M ¹⁹	33.5
		BS-360E ¹⁰	31.8		

English	Abbreviated name	Model	Unit	Calibration Value
Русский	сокращенное наименование	модель	Прибор	Принцип калибровки
Português	Nome abreviado	Modelo	Unidade	Regra de calibração
Español	nombre abreviado	modelo	Unidad	Regla de calibración
Italiano	abbreviazione	modelli	Unità	Regola di calibrazione
Türkçe	kısaltılmış ad	model	Ünite	Kalibrasyon Kuralı

TBA	
English	Total Bile Acids
Русский	желчные кислоты
Português	Ácido Biliar Total
Español	ácidos biliares totales
Italiano	acidi biliari totali
Türkçe	Total Safra Asitleridir

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.**BS-120**: BS-120, BS-130;

2.**BS-180**: BS-180, BS-190;

3.**BS-200**: BS-200, BS-220;

4.**BS-200E**: BS-200E, BS-220E;

5.**BS-230**: BS-230, BS-240, BS-280;

6.**BS-240E**: BS-240E, BS-240Pro;

7.**BS-300**: BS-300, BS-320;

8.**BS-330**: BS-330, BS-350;

9.**BS-330E**:**BS-330E**(Serial Number starts with "XQ-"),

BS-350E(Serial Number starts with "XS-");

10.**BS-360E**: BS-360E, BS-370E, BS-350S, BS-360S,
BS-330E(V35.00)(Serial Number starts with "W8-"
and software version starts with "35.00"),
BS-350E(V35.00) (Serial Number starts with "W9-"
and software version starts with "35.00");

11.**BS-380**: BS-380, BS-390;

12.**BS-400**: BS-400, BS-420;

13.**BS-430**: BS-410, BS-430, BS-450, BS-460, BS-470;

14.**BS-480**: BS-480, BS-490;

15.**BS-600**: BS-600, BS-620;

16.**BS-600M**: BS-600M, BS-620M;

17.**BS-800**: BS-800, BS-820, BS-800M, BS-820M,

BS-1800, BS-1800plus;

18.**BS-830**: BS-830, BS-830S,BS-840, BS-850, BS-860,

BS-870;

19.**BS-2000**: BS-2000, BS-2200, BS-2000M, BS-2200M.

20.**BS-2800M**: BS-2600M, BS-2800M.

Abbreviated name	TBA Control(L)					TBA Control(H)				
	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)	1 SD
TBA (apply to lot 061821001)	BS-120	¹ μmol/L	/	/ - /	/	BS-120	¹ μmol/L	/	/ - /	/
	BS-180	² μmol/L	/	/ - /	/	BS-180	² μmol/L	/	/ - /	/
	BS-200	³ μmol/L	/	/ - /	/	BS-200	³ μmol/L	/	/ - /	/
	BS-200E	⁴ μmol/L	/	/ - /	/	BS-200E	⁴ μmol/L	/	/ - /	/
	BS-230	⁵ μmol/L	/	/ - /	/	BS-230	⁵ μmol/L	/	/ - /	/
	BS-240E	⁶ μmol/L	/	/ - /	/	BS-240E	⁶ μmol/L	/	/ - /	/
	BS-300	⁷ μmol/L	/	/ - /	/	BS-300	⁷ μmol/L	/	/ - /	/
	BS-330	⁸ μmol/L	/	/ - /	/	BS-330	⁸ μmol/L	/	/ - /	/
	BS-330E	⁹ μmol/L	/	/ - /	/	BS-330E	⁹ μmol/L	/	/ - /	/
	BS-360E	¹⁰ μmol/L	/	/ - /	/	BS-360E	¹⁰ μmol/L	/	/ - /	/
	BS-380	¹¹ μmol/L	/	/ - /	/	BS-380	¹¹ μmol/L	/	/ - /	/
	BS-400	¹² μmol/L	/	/ - /	/	BS-400	¹² μmol/L	/	/ - /	/
	BS-430	¹³ μmol/L	/	/ - /	/	BS-430	¹³ μmol/L	/	/ - /	/
	BS-480	¹⁴ μmol/L	/	/ - /	/	BS-480	¹⁴ μmol/L	/	/ - /	/
	BS-600	¹⁵ μmol/L	/	/ - /	/	BS-600	¹⁵ μmol/L	/	/ - /	/
	BS-600M	¹⁶ μmol/L	/	/ - /	/	BS-600M	¹⁶ μmol/L	/	/ - /	/
	BS-800	¹⁷ μmol/L	/	/ - /	/	BS-800	¹⁷ μmol/L	/	/ - /	/
	BS-830	¹⁸ μmol/L	/	/ - /	/	BS-830	¹⁸ μmol/L	/	/ - /	/
	BS-2000	¹⁹ μmol/L	/	/ - /	/	BS-2000	¹⁹ μmol/L	/	/ - /	/
	BS-2800M	²⁰ μmol/L	/	/ - /	/	BS-2800M	²⁰ μmol/L	/	/ - /	/

CO2 and TBA Multi Control



TBA Control(L)							TBA Control(H)					
Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD
TBA (apply to lot 061821002)	BS-120	¹ μmol/L	/	/	-	/	BS-120	¹ μmol/L	/	/	-	/
	BS-180	² μmol/L	/	/	-	/	BS-180	² μmol/L	/	/	-	/
	BS-200	³ μmol/L	/	/	-	/	BS-200	³ μmol/L	/	/	-	/
	BS-200E	⁴ μmol/L	/	/	-	/	BS-200E	⁴ μmol/L	/	/	-	/
	BS-230	⁵ μmol/L	/	/	-	/	BS-230	⁵ μmol/L	/	/	-	/
	BS-240E	⁶ μmol/L	/	/	-	/	BS-240E	⁶ μmol/L	/	/	-	/
	BS-300	⁷ μmol/L	/	/	-	/	BS-300	⁷ μmol/L	/	/	-	/
	BS-330	⁸ μmol/L	/	/	-	/	BS-330	⁸ μmol/L	/	/	-	/
	BS-330E	⁹ μmol/L	/	/	-	/	BS-330E	⁹ μmol/L	/	/	-	/
	BS-360E	¹⁰ μmol/L	/	/	-	/	BS-360E	¹⁰ μmol/L	/	/	-	/
	BS-380	¹¹ μmol/L	/	/	-	/	BS-380	¹¹ μmol/L	/	/	-	/
	BS-400	¹² μmol/L	/	/	-	/	BS-400	¹² μmol/L	/	/	-	/
	BS-430	¹³ μmol/L	/	/	-	/	BS-430	¹³ μmol/L	/	/	-	/
	BS-480	¹⁴ μmol/L	/	/	-	/	BS-480	¹⁴ μmol/L	/	/	-	/
	BS-600	¹⁵ μmol/L	/	/	-	/	BS-600	¹⁵ μmol/L	/	/	-	/
	BS-600M	¹⁶ μmol/L	/	/	-	/	BS-600M	¹⁶ μmol/L	/	/	-	/
	BS-800	¹⁷ μmol/L	/	/	-	/	BS-800	¹⁷ μmol/L	/	/	-	/
	BS-830	¹⁸ μmol/L	/	/	-	/	BS-830	¹⁸ μmol/L	/	/	-	/
	BS-2000	¹⁹ μmol/L	/	/	-	/	BS-2000	¹⁹ μmol/L	/	/	-	/
	BS-2800M	²⁰ μmol/L	/	/	-	/	BS-2800M	²⁰ μmol/L	/	/	-	/

TBA Control(L)							TBA Control(H)							
Abbreviated name	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value±3SD)		1 SD		
TBA (apply to lot 061821003, 061821004)	BS-120	¹ μmol/L	20.7	16.5	-	24.9	1.4	BS-120	¹ μmol/L	39.5	31.6	-	47.4	2.6
	BS-180	² μmol/L	20.7	16.5	-	24.9	1.4	BS-180	² μmol/L	39.5	31.6	-	47.4	2.6
	BS-200	³ μmol/L	20.1	16.1	-	24.1	1.3	BS-200	³ μmol/L	39.3	31.4	-	47.2	2.6
	BS-200E	⁴ μmol/L	20.1	16.1	-	24.1	1.3	BS-200E	⁴ μmol/L	40.4	32.3	-	48.5	2.7
	BS-230	⁵ μmol/L	20.1	16.1	-	24.1	1.3	BS-230	⁵ μmol/L	38.8	31.0	-	46.6	2.6
	BS-240E	⁶ μmol/L	19.3	15.4	-	23.2	1.3	BS-240E	⁶ μmol/L	37.3	29.8	-	44.8	2.5
	BS-300	⁷ μmol/L	20.5	16.4	-	24.6	1.4	BS-300	⁷ μmol/L	39.3	31.4	-	47.2	2.6
	BS-330	⁸ μmol/L	20.1	16.1	-	24.1	1.3	BS-330	⁸ μmol/L	39.3	31.4	-	47.2	2.6
	BS-330E	⁹ μmol/L	20.1	16.1	-	24.1	1.3	BS-330E	⁹ μmol/L	40.4	32.3	-	48.5	2.7
	BS-360E	¹⁰ μmol/L	20.2	16.1	-	24.3	1.4	BS-360E	¹⁰ μmol/L	38.2	30.5	-	45.9	2.6
	BS-380	¹¹ μmol/L	20.1	16.1	-	24.1	1.3	BS-380	¹¹ μmol/L	39.1	31.2	-	47.0	2.6
	BS-400	¹² μmol/L	20.5	16.4	-	24.6	1.4	BS-400	¹² μmol/L	38.7	30.9	-	46.5	2.6
	BS-430	¹³ μmol/L	20.6	16.5	-	24.7	1.4	BS-430	¹³ μmol/L	39.2	31.3	-	47.1	2.6
	BS-480	¹⁴ μmol/L	20.4	16.3	-	24.5	1.4	BS-480	¹⁴ μmol/L	39.3	31.4	-	47.2	2.6
	BS-600	¹⁵ μmol/L	20.9	16.7	-	25.1	1.4	BS-600	¹⁵ μmol/L	39.8	31.8	-	47.8	2.7
	BS-600M	¹⁶ μmol/L	20.7	16.5	-	24.9	1.4	BS-600M	¹⁶ μmol/L	39.2	31.3	-	47.1	2.6
	BS-800	¹⁷ μmol/L	20.5	16.4	-	24.6	1.4	BS-800	¹⁷ μmol/L	39.3	31.4	-	47.2	2.6
	BS-830	¹⁸ μmol/L	21.2	16.9	-	25.5	1.4	BS-830	¹⁸ μmol/L	40.2	32.1	-	48.3	2.7
	BS-2000	¹⁹ μmol/L	21.3	17.0	-	25.6	1.4	BS-2000	¹⁹ μmol/L	40.6	32.4	-	48.8	2.7
	BS-2800M	²⁰ μmol/L	21.4	17.1	-	25.7	1.4	BS-2800M	²⁰ μmol/L	40.7	32.5	-	48.9	2.7

CO2 and TBA Multi Control



		TBA Control(L)					TBA Control(H)				
Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)	1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)	1 SD	
TBA (apply to lot 061822001)	BS-120	¹ $\mu\text{mol/L}$	20.7	16.5 — 24.9	1.4	BS-120	¹ $\mu\text{mol/L}$	37.9	30.3 — 45.5	2.5	
	BS-180	² $\mu\text{mol/L}$	20.7	16.5 — 24.9	1.4	BS-180	² $\mu\text{mol/L}$	37.9	30.3 — 45.5	2.5	
	BS-200	³ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-200	³ $\mu\text{mol/L}$	37.2	29.7 — 44.7	2.5	
	BS-200E	⁴ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-200E	⁴ $\mu\text{mol/L}$	37.4	29.9 — 44.9	2.5	
	BS-230	⁵ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-230	⁵ $\mu\text{mol/L}$	37.0	29.6 — 44.4	2.5	
	BS-240E	⁶ $\mu\text{mol/L}$	19.3	15.4 — 23.2	1.3	BS-240E	⁶ $\mu\text{mol/L}$	35.9	28.7 — 43.1	2.4	
	BS-300	⁷ $\mu\text{mol/L}$	20.5	16.4 — 24.6	1.4	BS-300	⁷ $\mu\text{mol/L}$	37.2	29.7 — 44.7	2.5	
	BS-330	⁸ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-330	⁸ $\mu\text{mol/L}$	37.2	29.7 — 44.7	2.5	
	BS-330E	⁹ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-330E	⁹ $\mu\text{mol/L}$	37.4	29.9 — 44.9	2.5	
	BS-360E	¹⁰ $\mu\text{mol/L}$	20.2	16.1 — 24.3	1.4	BS-360E	¹⁰ $\mu\text{mol/L}$	36.2	28.9 — 43.5	2.4	
	BS-380	¹¹ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-380	¹¹ $\mu\text{mol/L}$	37.1	29.6 — 44.6	2.5	
	BS-400	¹² $\mu\text{mol/L}$	20.5	16.4 — 24.6	1.4	BS-400	¹² $\mu\text{mol/L}$	36.8	29.4 — 44.2	2.5	
	BS-430	¹³ $\mu\text{mol/L}$	20.6	16.5 — 24.7	1.4	BS-430	¹³ $\mu\text{mol/L}$	37.5	30.0 — 45.0	2.5	
	BS-480	¹⁴ $\mu\text{mol/L}$	20.4	16.3 — 24.5	1.4	BS-480	¹⁴ $\mu\text{mol/L}$	37.1	29.6 — 44.6	2.5	
	BS-600	¹⁵ $\mu\text{mol/L}$	20.9	16.7 — 25.1	1.4	BS-600	¹⁵ $\mu\text{mol/L}$	37.8	30.2 — 45.4	2.5	
	BS-600M	¹⁶ $\mu\text{mol/L}$	20.7	16.5 — 24.9	1.4	BS-600M	¹⁶ $\mu\text{mol/L}$	37.5	30.0 — 45.0	2.5	
	BS-800	¹⁷ $\mu\text{mol/L}$	20.5	16.4 — 24.6	1.4	BS-800	¹⁷ $\mu\text{mol/L}$	37.3	29.8 — 44.8	2.5	
	BS-830	¹⁸ $\mu\text{mol/L}$	21.2	16.9 — 25.5	1.4	BS-830	¹⁸ $\mu\text{mol/L}$	38.4	30.7 — 46.1	2.6	
	BS-2000	¹⁹ $\mu\text{mol/L}$	21.3	17.0 — 25.6	1.4	BS-2000	¹⁹ $\mu\text{mol/L}$	38.7	30.9 — 46.5	2.6	
	BS-2800M	²⁰ $\mu\text{mol/L}$	21.4	17.1 — 25.7	1.4	BS-2800M	²⁰ $\mu\text{mol/L}$	38.7	30.9 — 46.5	2.6	

		TBA Control(L)					TBA Control(H)				
Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)	1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)	1 SD	
TBA (apply to lot 061822002)	BS-120	¹ $\mu\text{mol/L}$	20.7	16.5 — 24.9	1.4	BS-120	¹ $\mu\text{mol/L}$	36.2	28.9 — 43.5	2.4	
	BS-180	² $\mu\text{mol/L}$	20.7	16.5 — 24.9	1.4	BS-180	² $\mu\text{mol/L}$	36.2	28.9 — 43.5	2.4	
	BS-200	³ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-200	³ $\mu\text{mol/L}$	36.3	29.0 — 43.6	2.4	
	BS-200E	⁴ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-200E	⁴ $\mu\text{mol/L}$	37.7	30.1 — 45.3	2.5	
	BS-230	⁵ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-230	⁵ $\mu\text{mol/L}$	35.5	28.4 — 42.6	2.4	
	BS-240E	⁶ $\mu\text{mol/L}$	19.3	15.4 — 23.2	1.3	BS-240E	⁶ $\mu\text{mol/L}$	34.3	27.4 — 41.2	2.3	
	BS-300	⁷ $\mu\text{mol/L}$	20.5	16.4 — 24.6	1.4	BS-300	⁷ $\mu\text{mol/L}$	36.9	29.5 — 44.3	2.5	
	BS-330	⁸ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-330	⁸ $\mu\text{mol/L}$	36.3	29.0 — 43.6	2.4	
	BS-330E	⁹ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-330E	⁹ $\mu\text{mol/L}$	37.7	30.1 — 45.3	2.5	
	BS-360E	¹⁰ $\mu\text{mol/L}$	20.2	16.1 — 24.3	1.4	BS-360E	¹⁰ $\mu\text{mol/L}$	35.0	28.0 — 42.0	2.3	
	BS-380	¹¹ $\mu\text{mol/L}$	20.1	16.1 — 24.1	1.3	BS-380	¹¹ $\mu\text{mol/L}$	35.7	28.5 — 42.9	2.4	
	BS-400	¹² $\mu\text{mol/L}$	20.5	16.4 — 24.6	1.4	BS-400	¹² $\mu\text{mol/L}$	35.5	28.4 — 42.6	2.4	
	BS-430	¹³ $\mu\text{mol/L}$	20.6	16.5 — 24.7	1.4	BS-430	¹³ $\mu\text{mol/L}$	35.9	28.7 — 43.1	2.4	
	BS-480	¹⁴ $\mu\text{mol/L}$	20.4	16.3 — 24.5	1.4	BS-480	¹⁴ $\mu\text{mol/L}$	35.4	28.3 — 42.5	2.4	
	BS-600	¹⁵ $\mu\text{mol/L}$	20.9	16.7 — 25.1	1.4	BS-600	¹⁵ $\mu\text{mol/L}$	36.4	29.1 — 43.7	2.4	
	BS-600M	¹⁶ $\mu\text{mol/L}$	20.7	16.5 — 24.9	1.4	BS-600M	¹⁶ $\mu\text{mol/L}$	36.0	28.8 — 43.2	2.4	
	BS-800	¹⁷ $\mu\text{mol/L}$	20.5	16.4 — 24.6	1.4	BS-800	¹⁷ $\mu\text{mol/L}$	35.9	28.7 — 43.1	2.4	
	BS-830	¹⁸ $\mu\text{mol/L}$	21.2	16.9 — 25.5	1.4	BS-830	¹⁸ $\mu\text{mol/L}$	36.8	29.4 — 44.2	2.5	
	BS-2000	¹⁹ $\mu\text{mol/L}$	21.3	17.0 — 25.6	1.4	BS-2000	¹⁹ $\mu\text{mol/L}$	37.3	29.8 — 44.8	2.5	
	BS-2800M	²⁰ $\mu\text{mol/L}$	21.4	17.1 — 25.7	1.4	BS-2800M	²⁰ $\mu\text{mol/L}$	37.3	29.8 — 44.8	2.5	

CO2 and TBA Multi Control



TBA Control(L)							TBA Control(H)							
Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD		
TBA (apply to lot 061822003, 061822004, 061822005)	BS-120	¹ $\mu\text{mol/L}$	21.5	17.2	—	25.8	1.4	BS-120	¹ $\mu\text{mol/L}$	36.2	28.9	—	43.5	2.4
	BS-180	² $\mu\text{mol/L}$	21.5	17.2	—	25.8	1.4	BS-180	² $\mu\text{mol/L}$	36.2	28.9	—	43.5	2.4
	BS-200	³ $\mu\text{mol/L}$	20.8	16.6	—	25.0	1.4	BS-200	³ $\mu\text{mol/L}$	36.3	29.0	—	43.6	2.4
	BS-200E	⁴ $\mu\text{mol/L}$	21.0	16.8	—	25.2	1.4	BS-200E	⁴ $\mu\text{mol/L}$	37.7	30.1	—	45.3	2.5
	BS-230	⁵ $\mu\text{mol/L}$	21.0	16.8	—	25.2	1.4	BS-230	⁵ $\mu\text{mol/L}$	35.5	28.4	—	42.6	2.4
	BS-240E	⁶ $\mu\text{mol/L}$	20.3	16.2	—	24.4	1.4	BS-240E	⁶ $\mu\text{mol/L}$	34.3	27.4	—	41.2	2.3
	BS-300	⁷ $\mu\text{mol/L}$	21.3	17.0	—	25.6	1.4	BS-300	⁷ $\mu\text{mol/L}$	36.9	29.5	—	44.3	2.5
	BS-330	⁸ $\mu\text{mol/L}$	20.8	16.6	—	25.0	1.4	BS-330	⁸ $\mu\text{mol/L}$	36.3	29.0	—	43.6	2.4
	BS-330E	⁹ $\mu\text{mol/L}$	21.0	16.8	—	25.2	1.4	BS-330E	⁹ $\mu\text{mol/L}$	37.7	30.1	—	45.3	2.5
	BS-360E	¹⁰ $\mu\text{mol/L}$	20.9	16.7	—	25.1	1.4	BS-360E	¹⁰ $\mu\text{mol/L}$	35.0	28.0	—	42.0	2.3
	BS-380	¹¹ $\mu\text{mol/L}$	21.3	17.0	—	25.6	1.4	BS-380	¹¹ $\mu\text{mol/L}$	35.7	28.5	—	42.9	2.4
	BS-400	¹² $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-400	¹² $\mu\text{mol/L}$	35.5	28.4	—	42.6	2.4
	BS-430	¹³ $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-430	¹³ $\mu\text{mol/L}$	35.9	28.7	—	43.1	2.4
	BS-480	¹⁴ $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-480	¹⁴ $\mu\text{mol/L}$	35.4	28.3	—	42.5	2.4
	BS-600	¹⁵ $\mu\text{mol/L}$	21.9	17.5	—	26.3	1.5	BS-600	¹⁵ $\mu\text{mol/L}$	36.4	29.1	—	43.7	2.4
	BS-600M	¹⁶ $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-600M	¹⁶ $\mu\text{mol/L}$	36.0	28.8	—	43.2	2.4
	BS-800	¹⁷ $\mu\text{mol/L}$	21.4	17.1	—	25.7	1.4	BS-800	¹⁷ $\mu\text{mol/L}$	35.9	28.7	—	43.1	2.4
	BS-830	¹⁸ $\mu\text{mol/L}$	22.1	17.7	—	26.5	1.5	BS-830	¹⁸ $\mu\text{mol/L}$	36.8	29.4	—	44.2	2.5
	BS-2000	¹⁹ $\mu\text{mol/L}$	22.3	17.8	—	26.8	1.5	BS-2000	¹⁹ $\mu\text{mol/L}$	37.3	29.8	—	44.8	2.5
	BS-2800M	²⁰ $\mu\text{mol/L}$	22.3	17.8	—	26.8	1.5	BS-2800M	²⁰ $\mu\text{mol/L}$	37.3	29.8	—	44.8	2.5

TBA Control(L)							TBA Control(H)							
Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)		1 SD		
TBA (apply to lot 061822006)	BS-120	¹ $\mu\text{mol/L}$	21.5	17.2	—	25.8	1.4	BS-120	¹ $\mu\text{mol/L}$	36.1	28.8	—	43.4	2.4
	BS-180	² $\mu\text{mol/L}$	21.5	17.2	—	25.8	1.4	BS-180	² $\mu\text{mol/L}$	36.1	28.8	—	43.4	2.4
	BS-200	³ $\mu\text{mol/L}$	20.8	16.6	—	25.0	1.4	BS-200	³ $\mu\text{mol/L}$	35.7	28.5	—	42.9	2.4
	BS-200E	⁴ $\mu\text{mol/L}$	21.0	16.8	—	25.2	1.4	BS-200E	⁴ $\mu\text{mol/L}$	38.2	30.5	—	45.9	2.6
	BS-230	⁵ $\mu\text{mol/L}$	21.0	16.8	—	25.2	1.4	BS-230	⁵ $\mu\text{mol/L}$	35.6	28.4	—	42.8	2.4
	BS-240E	⁶ $\mu\text{mol/L}$	20.3	16.2	—	24.4	1.4	BS-240E	⁶ $\mu\text{mol/L}$	34.3	27.4	—	41.2	2.3
	BS-300	⁷ $\mu\text{mol/L}$	21.3	17.0	—	25.6	1.4	BS-300	⁷ $\mu\text{mol/L}$	37.4	29.9	—	44.9	2.5
	BS-330	⁸ $\mu\text{mol/L}$	20.8	16.6	—	25.0	1.4	BS-330	⁸ $\mu\text{mol/L}$	35.7	28.5	—	42.9	2.4
	BS-330E	⁹ $\mu\text{mol/L}$	21.0	16.8	—	25.2	1.4	BS-330E	⁹ $\mu\text{mol/L}$	38.2	30.5	—	45.9	2.6
	BS-360E	¹⁰ $\mu\text{mol/L}$	20.9	16.7	—	25.1	1.4	BS-360E	¹⁰ $\mu\text{mol/L}$	35.3	28.2	—	42.4	2.4
	BS-380	¹¹ $\mu\text{mol/L}$	21.3	17.0	—	25.6	1.4	BS-380	¹¹ $\mu\text{mol/L}$	35.6	28.4	—	42.8	2.4
	BS-400	¹² $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-400	¹² $\mu\text{mol/L}$	35.6	28.4	—	42.8	2.4
	BS-430	¹³ $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-430	¹³ $\mu\text{mol/L}$	35.9	28.7	—	43.1	2.4
	BS-480	¹⁴ $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-480	¹⁴ $\mu\text{mol/L}$	35.7	28.5	—	42.9	2.4
	BS-600	¹⁵ $\mu\text{mol/L}$	21.9	17.5	—	26.3	1.5	BS-600	¹⁵ $\mu\text{mol/L}$	36.1	28.8	—	43.4	2.4
	BS-600M	¹⁶ $\mu\text{mol/L}$	21.6	17.3	—	25.9	1.4	BS-600M	¹⁶ $\mu\text{mol/L}$	35.8	28.6	—	43.0	2.4
	BS-800	¹⁷ $\mu\text{mol/L}$	21.4	17.1	—	25.7	1.4	BS-800	¹⁷ $\mu\text{mol/L}$	35.9	28.7	—	43.1	2.4
	BS-830	¹⁸ $\mu\text{mol/L}$	22.1	17.7	—	26.5	1.5	BS-830	¹⁸ $\mu\text{mol/L}$	36.6	29.2	—	44.0	2.5
	BS-2000	¹⁹ $\mu\text{mol/L}$	22.3	17.8	—	26.8	1.5	BS-2000	¹⁹ $\mu\text{mol/L}$	37.3	29.8	—	44.8	2.5
	BS-2800M	²⁰ $\mu\text{mol/L}$	22.3	17.8	—	26.8	1.5	BS-2800M	²⁰ $\mu\text{mol/L}$	37.3	29.8	—	44.8	2.5

CO2 and TBA Multi Control



English	Abbreviated name	Model	Unit	Assay Value	Range (Assay Value \pm 3SD)
Русский	сокращенное наименование	модель	Прибор	Результат анализа	Диапазон (результат анализа \pm 3CO)
Português	Nome abreviado	Modelo	Unidade	Valores da análise	Faixa (Valores da análise \pm 3SD)
Español	nombre abreviado	modelo	Unidad	Valor de ensayo	Rango (Valor de ensayo \pm 3SD)
Italiano	abbreviazione	modelli	Unità	Valori di dosaggio	Intervallo (valore diconcentrazione \pm 3 DS)
Türkçe	kısaltılmış ad	model	Ünite	Tayin Değeri	Aralık (Tayin Değeri \pm 3SD)
English		CO2		TBA	
English		Carbon Dioxide		Total Bile Acids	
Русский		Двуокись углерода		общие желчные кислоты	
Português		Dióxido de Carbono		ácidos biliares totales	
Español		Dióxido de carbono		ácidos biliares totais	
Italiano		Anidride carbonica		acidi biliari totali	
Türkçe		Karbon Dioksit		total safra asitleridir	