

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-120: BS-120, BS-130, BS-180, BS-190;

11.BS-430: BS-430, BS-450, BS-460;

2.BS-200: BS-200, BS-220, BS-330, BS-350;

12.BS-480: BS-480, BS-490;

3.BS-200E: BS-200E, BS-220E,;

13.BS-600: BS-600, BS-620;

4.BS-240: BS-230, BS-240;

14.BS-600M: BS-600M, BS-620M;

5.BS-240E: BS240E, BS240Pro;

15.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

6.BS-300: BS-300, BS-320;

7.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")

16.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00)

17.BS-2800M:BS-2600M, BS-2800M.

(Serial Number starts with "W9-" and software version starts with "35.00");

18.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

9.BS-380: BS-380, BS-390;

19.For applicable analyzers of the analyte, please subject to the

10.BS-400: BS-400, BS-420;

parameter sheet and instrument.



LOT : 143224005



: 2025-03-07

| English | Abbreviated name | Unit | Calibration Rule | Model |
|-----------|------------------------------|---------|------------------------|---------|
| Русский | сокращенное наименование | Прибор | Принцип калибровки | модель |
| Português | Nome abreviado | Unidade | Regra de calibração | Modelo |
| Español | nombre abreviado | Unidad | Regla de calibración | modelo |
| Italiano | abbreviazione | Unità | Regola di calibrazione | modelli |
| Türkçe | kısaltılmış ad | Ünite | Kalibrasyon Kuralı | model |
| English | Expanded Uncertainty | | | |
| Русский | Расширенная неопределенность | | | |
| Português | Incerteza Expandida | | | |
| Español | Incertidumbre expandida | | | |
| Italiano | Incertezza estesa | | | |
| Türkçe | Daha Uzun Süreli Belirsizlik | | | |
| TBA | | | | |
| English | Total Bile Acids | | | |
| Русский | желчные кислоты | | | |
| Português | Ácido Biliar Total | | | |
| Español | ácidos biliares totales | | | |
| Italiano | acidi biliari totali | | | |
| Türkçe | Total Safra Asitleridir | | | |

| Abbreviated name | Unit | Calibration Rule | | | Two-point linear | | |
|------------------|----------------|----------------------|------------------|----------------------|------------------------|------------------|----------------------|
| | | Model | Calibrator Value | Expanded Uncertainty | Model | Calibrator Value | Expanded Uncertainty |
| TBA | μ mol/L | BS-120 ¹ | 35.8 | 1.0 | BS-400 ¹⁰ | 34.4 | 1.0 |
| | | BS-200 ² | 34.4 | 1.0 | BS-430 ¹¹ | 34.5 | 1.0 |
| | | BS-200E ³ | 35.0 | 1.0 | BS-480 ¹² | 34.8 | 1.0 |
| | | BS-240 ⁴ | 34.4 | 1.0 | BS-600 ¹³ | 34.4 | 1.0 |
| | | BS-240E ⁵ | 34.4 | 1.0 | BS-600M ¹⁴ | 34.4 | 1.0 |
| | | BS-300 ⁶ | 34.4 | 1.0 | BS-800 ¹⁵ | 34.4 | 1.0 |
| | | BS-330E ⁷ | 35.0 | 1.0 | BS-2000 ¹⁶ | 34.8 | 1.0 |
| | | BS-360E ⁸ | 34.4 | 1.0 | BS-2800M ¹⁷ | 35.5 | 1.0 |
| | | BS-380 ⁹ | 34.4 | 1.0 | | | |

mindray

TBA Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 143224005

🕒 2025-03-07

