

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-200: BS-200, BS-220, BS-330, BS-350;	10.BS-430: BS-430, BS-450, BS-460;
2.BS-200E: BS-200E, BS-220E,;	11.BS-480: BS-480, BS-490;
3.BS-240: BS-230, BS-240;	12.BS-600: BS-600, BS-620;
4.BS-240E: BS240E, BS240Pro;	13.BS-600M: BS-600M, BS-620M;
5.BS-300: BS-300, BS-320;	14.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;
6.BS-330E:BS-330E(Serial Number starts with "XQ-"), BS-350E(Serial Number starts with "XS-")	15.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;
7.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");	16.BS-2800M:BS-2600M, BS-2800M.
8.BS-380: BS-380, BS-390;	17.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;
9.BS-400: BS-400, BS-420;	18.Applicable models of the chemistry shall be subject to the parameter list and instructions.



LOT : 145123013



EXP : 2024-09-14

English	Abbreviated name	Unit	Calibration Rule	Model
Русский	сокращенное наименование	Прибор	Принцип калибровки	модель
Português	Nome abreviado	Unidade	Regra de calibração	Modelo
Español	nombre abreviado	Unidad	Regla de calibración	modelo
Italiano	abbreviazione	Unità	Regola di calibrazione	modell
Türkçe	kısaltılmış ad	Ünite	Kalibrasyon Kuralı	model

English	Expanded Uncertainty
Русский	Расширенная неопределенность
Português	Incerteza Expandida
Español	Incertidumbre expandida
Italiano	Incertezza estesa
Türkçe	Daha Uzun Süreli Belirsizlik

CO2	
English	Carbon Dioxide
Русский	Двуокись углерода
Português	Dióxido de Carbono
Español	Dióxido de carbono
Italiano	Anidride carbonica
Türkçe	Karbon Dioksit

CO2 Calibrator



Abbreviated name	Unit	Calibration Rule				Two-point linear	
		Model	Calibrator Value	Expanded Uncertainty	Model	Calibrator Value	Expanded Uncertainty
CO2	mmol/ L	BS-200¹	29.2	0.4	BS-400⁹	29.2	0.4
		BS-200E²	29.2	0.4	BS-430¹⁰	29.2	0.4
		BS-240³	29.2	0.4	BS-480¹¹	29.2	0.4
		BS-240E⁴	29.2	0.4	BS-600¹²	29.2	0.4
		BS-300⁵	29.2	0.4	BS-600M¹³	28.8	0.4
		BS-330E⁶	29.2	0.4	BS-800¹⁴	29.2	0.4
		BS-360E⁷	29.2	0.4	BS-2000¹⁵	29.5	0.4
		BS-380⁸	29.2	0.4	BS-2800M¹⁶	28.8	0.4
	mol/L	BS-200¹	0.0292	0.0004	BS-400⁹	0.0292	0.0004
		BS-200E²	0.0292	0.0004	BS-430¹⁰	0.0292	0.0004
		BS-240³	0.0292	0.0004	BS-480¹¹	0.0292	0.0004
		BS-240E⁴	0.0292	0.0004	BS-600¹²	0.0292	0.0004
		BS-300⁵	0.0292	0.0004	BS-600M¹³	0.0288	0.0004
		BS-330E⁶	0.0292	0.0004	BS-800¹⁴	0.0292	0.0004
		BS-360E⁷	0.0292	0.0004	BS-2000¹⁵	0.0295	0.0004
		BS-380⁸	0.0292	0.0004	BS-2800M¹⁶	0.0288	0.0004

mindray

CO2 Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 145123013

2024-09-14

