

Value sheet of Mindray BS Measurement System

Русский: Таблица результатов для системы BS компании Mindray

Português: Planilha de valores do Sistema de Medição BS da Mindray

Español: Hoja de valores del sistema de medición Mindray BS

Italiano: Scheda dei valori del sistema di misurazione BS di Mindray

Türkçe: Mindray BS Ölçüm Sistemi'nin değer sayfası



The data of each group is same.

Русский: Данные совпадают во всех группах.

Português: A dados de cada grupo é a mesma.

Español: la datos de cada grupo es la misma.

Italiano: la dati di ogni gruppo è la stessa.

Türkçe: her grubun veri aynıdır.

1.BS-200: BS-200, BS-220, BS-330, BS-350;

10.BS-430: BS-430, BS-450, BS-460;

2.BS-200E: BS-200E, BS-220E,;

11.BS-480: BS-480, BS-490;

3.BS-240: BS-230, BS-240;

12.BS-600: BS-600, BS-620;

4.BS-240E: BS240E, BS240Pro;

13.BS-600M: BS-600M, BS-620M;

5.BS-300: BS-300, BS-320;

6.BS-330E:BS-330E(Serial Number starts with "XQ-"),
BS-350E(Serial Number starts with "XS-")

14.BS-800: BS-800, BS-820, BS-800M, BS-820M, BS-1800, BS-1800plus;

7.BS-360E: BS-360E, BS-370E, BS-350S, BS-360S, BS-330E(V35.00)(Serial Number starts with "W8-" and software version starts with "35.00"), BS-350E(V35.00) (Serial Number starts with "W9-" and software version starts with "35.00");

15.BS-2000: BS-2000, BS-2200, BS-2000M, BS-2200M;

8.BS-380: BS-380, BS-390;

16.BS-2800M:BS-2600M, BS-2800M.

9.BS-400: BS-400, BS-420;

17.0.9% NaCl is needed, Conc. Of 0.9% NaCl =0;

18.Applicable models of the chemistry shall be subject to the parameter list and instructions.

LOT

: 145124004



: 2025-01-20

| English | Abbreviated name | Unit | Calibration Rule | Model |
|-----------|------------------------------|---------|------------------------|---------|
| Русский | сокращенное наименование | Прибор | Принцип калибровки | модель |
| Português | Nome abreviado | Unidade | Regra de calibração | Modelo |
| Español | nombre abreviado | Unidad | Regla de calibración | modelo |
| Italiano | abbreviazione | Unità | Regola di calibrazione | modelli |
| Türkçe | kısaltılmış ad | Ünite | Kalibrasyon Kuralı | model |
| English | Expanded Uncertainty | | | |
| Русский | Расширенная неопределенность | | | |
| Português | Incerteza Expandida | | | |
| Español | Incertidumbre expandida | | | |
| Italiano | Incertezza estesa | | | |
| Türkçe | Daha Uzun Süreli Belirsizlik | | | |
| CO2 | | | | |
| English | Carbon Dioxide | | | |
| Русский | Двуокись углерода | | | |
| Português | Dióxido de Carbono | | | |
| Español | Dióxido de carbono | | | |
| Italiano | Anidride carbonica | | | |
| Türkçe | Karbon Dioksit | | | |


| Abbreviated name | Unit | Calibration Rule | | | | Two-point linear | |
|------------------|--------|----------------------|------------------|----------------------|------------------------|------------------|----------------------|
| | | Model | Calibrator Value | Expanded Uncertainty | Model | Calibrator Value | Expanded Uncertainty |
| CO2 | mmol/L | BS-200 ¹ | 28.9 | 0.4 | BS-400 ⁹ | 28.9 | 0.4 |
| | | BS-200E ² | 28.9 | 0.4 | BS-430 ¹⁰ | 28.9 | 0.4 |
| | | BS-240 ³ | 28.9 | 0.4 | BS-480 ¹¹ | 28.9 | 0.4 |
| | | BS-240E ⁴ | 28.9 | 0.4 | BS-600 ¹² | 28.9 | 0.4 |
| | | BS-300 ⁵ | 28.9 | 0.4 | BS-600M ¹³ | 29.8 | 0.4 |
| | | BS-330E ⁶ | 28.9 | 0.4 | BS-800 ¹⁴ | 28.9 | 0.4 |
| | | BS-360E ⁷ | 28.9 | 0.4 | BS-2000 ¹⁵ | 29.5 | 0.4 |
| | | BS-380 ⁸ | 28.9 | 0.4 | BS-2800M ¹⁶ | 29.8 | 0.4 |
| | mol/L | BS-200 ¹ | 0.0289 | 0.0004 | BS-400 ⁹ | 0.0289 | 0.0004 |
| | | BS-200E ² | 0.0289 | 0.0004 | BS-430 ¹⁰ | 0.0289 | 0.0004 |
| | | BS-240 ³ | 0.0289 | 0.0004 | BS-480 ¹¹ | 0.0289 | 0.0004 |
| | | BS-240E ⁴ | 0.0289 | 0.0004 | BS-600 ¹² | 0.0289 | 0.0004 |
| | | BS-300 ⁵ | 0.0289 | 0.0004 | BS-600M ¹³ | 0.0298 | 0.0004 |
| | | BS-330E ⁶ | 0.0289 | 0.0004 | BS-800 ¹⁴ | 0.0289 | 0.0004 |
| | | BS-360E ⁷ | 0.0289 | 0.0004 | BS-2000 ¹⁵ | 0.0295 | 0.0004 |
| | | BS-380 ⁸ | 0.0289 | 0.0004 | BS-2800M ¹⁶ | 0.0298 | 0.0004 |

mindray

CO2 Calibrator

For use on: BS-2000, BS-2800M,
BS-600M, BS-800

LOT 145124004

 2025-01-20

